

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

29-06760-08


Docket or Reference number

030-12688

Amendment No. 04

Radiology-Ultrasound-Nuclear Consultants, PA
Radiation-Oncology Institute
303 West Main Street
Freehold, New Jersey 07728

In accordance with the letter dated February 22, 1993, License Number 29-06760-08 is hereby terminated.



Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions 2
FOIA- 2006-0146

F/2

9312020512 931025
PDR ADDCA 03012688
PDR

For the U.S. Nuclear Regulatory Commission

Original Signed By:

Francis M. Costello

By

Nuclear Materials Safety Branch
Region I

King of Prussia, Pennsylvania 19406

Date

OCT 25 1993

OFFICIAL RECORD COPY ML 10

NOTE TO DMB:

THE ATTACHED DOCUMENTS ARE TO BE PROCESSED AS ONE MATERIALS LICENSING PACKAGE.

LICENSE NUMBER: 17-06762-08

DOCKET NUMBER: 12677

CONTROL NUMBER: 1717

THIS SHEET MAY BE DISCARDED AFTER PROCESSING.

THANK YOU!

29-151

ML101

OCT 25 1993

License No. 29-06760-08
Docket No. 030-12688
Control No. 117810

Radiology-Ultrasound-Nuclear Consultants, PA
ATTN: G. A. Doener, M.D.
Radiation-Oncology Institute
303 West Main Street
Freehold, New Jersey 07728

Dear Dr. Doener:

As requested in your letter dated February 22, 1993 and based on the results of a closeout safety inspection conducted by an NRC inspector at the above address on April 30, 1993, please find enclosed Amendment No. 04 terminating License No. 29-06760-08.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:
Francis M. Costello

Francis M. Costello, Chief
Medical Licensing Section
Division of Radiation Safety
and Safeguards

Enclosure:
Amendment No. 04

DRSS:RI
Cahill/gc/MC

FMC
DRSS:RI
Costello

10/25/93

10/26/93

OFFICIAL RECORD COPY - G:\WPS\MLTR\2906760.08 - 10/21/93

ML 10

10 pages

~~03/12/93~~

TELEPHONE CONVERSATION RECORD	Date: 8/3/93	Time: 2:00p.m.
Mail Control No.:	License No.: 29-06760-08	Docket No.: 030-12688
Person Called: Dr. Doener	Organization: Radiology Ultrasound Nuclear Consultants	telephone Number: 908- 222-3456
Person Calling: Mary Cahill		
Subject: Deficiency letter dated 6/11/93 in response to termination request for license no. 29-06760-08		
<p>Summary: Dr. Doener stated that he believed his license had already been terminated. He did not understand why he needed to do a survey when the NRC already did one. He stated that he had been turning back mail received from the NRC. He finally agreed to do a survey but did not understand why he should not use an ion chamber. After an attempt to explain this was made, Dr. Doener said he would try to borrow a G M to do the survey. He said he would send the results to the NRC within 30 days.</p> <p>Dr. Doener said that he was now referring all patients to Hillary Clinton.</p>		
Action Required/Taken: <i>To Section Chief for approval of extension</i>		
Signature: <i>Mary Cahill</i>	Date: <i>8/3/93</i>	

OFFICE OF THE ATTORNEY GENERAL, MI, 10

117810

JUN 11 1993

License No. 29-06760-08
Docket No. 030-12688
Control No. 117810

Radiology Ultrasound Nuclear Consultants, PA
Radiation Oncology Institute
ATTN: G. A. Doener, M.D.
303 West Main Street
Freehold, New Jersey 07728

Dear Dr. Doener:

This is in reference to your request in a letter dated February 22, 1993 to terminate License No. 29-06760-08. In order to terminate your license, you must complete and submit to this office the enclosed RC Form 314 - Certificate of Disposition of Materials. In addition, prior to release of your facility for unrestricted use, it will be necessary for you to perform a closeout survey (guide enclosed) of your facility and submit these results to us for review. In submitting the survey results, please indicate the manufacturer, model number, and sensitivity of the survey instruments used to measure residual activity.

We will continue our review of your request to terminate your license upon receipt of this information. Please reply to my attention at the Region I office and refer to Mail Control No. 117810. If you have any technical questions regarding this letter please call Ms. Mary Cahill at (215) 337-5109.

In order to continue our prompt review, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

Sincerely,

Original Signed By:
Thomas K. Thompson

Thomas K. Thompson
Nuclear Materials Safety Branch
Division of Radiation Safety
and Safeguards

OFFICIAL RECORD COPY - S:\PENDING\DOENER.TER - 06/03/93

Enclosures:

1. Guidelines for Decontamination of Facilities and Equipment Prior to Release for Unrestricted Use or Termination of Licenses for Byproduct, Source, or Special Nuclear Material
2. NRC Form 314-Certificate of Disposition of Materials

RET
NMSS:RI

for Cahill/M:amw

6/8/93

RET
NMSS:RI

Thompson

6/8/93

CONVERSATION RECORD

TIME

2:50 PM

DATE

4/14/93

TYPE

 VISIT CONFERENCE TELEPHONE INCOMING OUTGOING

ROUTINE

NAME/SYMBOL / INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

ORGANIZATION (Offic. empl. Bureau, etc.)

TELEPHONE NO.

Dr. Doener

Sutro

(408)
780-2036

SUBJECT

Letter dated 12/10/92 + request to terminate

SUMMARY

Dr. Doener stated that he will vacate the facility by April 30, 1993. He also stated that his physicist performed wipe test of CS-132 source and concluded that it was leaking. I asked the activity removed he did not know of Rand but indicated it was slightly leaking. I asked if he had notified the NRC when the source was determined to be leaking he believes the physicist did. I told him that the 12/10 letter referring to disposal is all I have. I informed him that I will terminate the Co-60 license but Mr. Cahill had the nuclear med. license action and that when RAM is removed he must perform a close out survey for our review. I told Tim Dwyer of 12/10/92 letter + he is very source incing we decide to perform our own survey before we terminate the N.M. license

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

J. A. Jovston

4/14/93

ACTION TAKEN

SIGNATURE

TITLE

DATE

50271-101

OFFICIAL RECORD COPY

OPTIONAL FORM 271 (12-78)
DEPARTMENT OF DEFENSE

117810

RADIOLOGY  ULTRASOUND NUCLEAR CONSULTANTS, P.A.

RADIATION • ONCOLOGY INSTITUTE
DR. G.A. DOENER
FREEHOLD PLAZA
303 WEST MAIN STREET • LOWER LEVEL RM. 8
FREEHOLD, NEW JERSEY 07728
TELEPHONE 908-780-2030

February 22, 1993

United States Nuclear Regulatory Commission
Region I
475 Allendale Road
King of Prussia, Pa. 19406

Re: License Number 29-06760-07
29-06760-08

Dear Ms. Joustra

With reference to my letter dated December 10, 1992 (copy attached), I wish to inform you that the date for closing the radiation facility at 303 W. Main St., Freehold, NJ 07728 (Picker 1000) has been set for March 3, 1993.

Will you please terminate the license as of March 3, 1993.

If there is any refund of the licensing fee due, will you please take care of that.

Will you also please terminate the license for nuclear materials by March 3, 1993.

I wish to reiterate that the Cobalt 60 wires, as well as a leaking Cesium 137 source, were discarded by Teledyne Isotopes in Westwood, NJ.

Sincerely yours,

G.A. Doener
G. A. Doener, M.D.
GAD/tam

FEE EXEMPT

Termination

RECEIVED BY LFDCB	
Date	<i>3/27/93</i>
Log	<i>mpk</i>
By	<i>ES</i>
Date Completed	<i>3/27/93</i>

OFFICIAL RECORD COPY

ML 10

Diplomate of The American Board of Radiology
American Board of Nuclear Medicine, Canadian Board of Radiology

117810

MAR 9 1993

RADIOLOGY ULTRASOUND NUCLEAR CONSULTANTS, P.A.

RADIATION • ONCOLOGY INSTITUTE
DR. G.A. DOENER
FREEHOLD PLAZA
303 WEST MAIN STREET • LOWER LEVEL RM. 8
FREEHOLD, NEW JERSEY 07728
TELEPHONE 908-780-2030

December 10, 1992

U. S. Nuclear Regulatory Commission
Region I
475 Allendale Rd.
King of Prussia, Pa. 19406-1415

117809

Re: License # 29-06760-07

Attn: Mohamed M. Shanbaky, Chief

Dear Dr. Shanbaky:

Exemption 2

Since our rental contract expires on March 15, 1993, we will remove the Cobalt unit Picker 1000 from our office. We have contacted [redacted] to remove the Cobalt sources and the Picker 1000 unit. Ms. Selby Stein informed us that they do not have a license for a Picker 1000 unit. Would you please issue a license for removal of the Picker 1000 unit located at 303 W. Mai Street, Freehold, NJ to AIC.

We wish to inform you that all of the Cobalt 60 wires which were stored in our Cobalt room, and the leaking Cesium 137 source were removed by Teledyne Isotopes.

Sincerely yours,

G.A. Doener
G. A. Doener, M.D.
GAD/tam

P.S. Please try to locate and return the tape on breast cancer which was loaned to your office.