

(10-2003)
10 CFR 2.201**Docket File Information
SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE PSI Energy, Inc. REPORT NUMBER(S) 2006-001		2. NRC/REGIONAL OFFICE Region III	
3. DOCKET NUMBER(S) 03009317	4. LICENSE NUMBER(S) 13-15544-01	5. DATE(S) OF INSPECTION May 17, 2006	
6. INSPECTION PROCEDURES USED 87124		7. INSPECTION FOCUS AREAS 03.01 - 03.07	
SUPPLEMENTAL INSPECTION INFORMATION			
1. PROGRAM CODE(S) 03120	2. PRIORITY 5	3. LICENSEE CONTACT Becky Sparks, RSO	4. TELEPHONE NUMBER 812-386-4202
<input type="checkbox"/> Main Office Inspection	Next Inspection Date: May 2011		
<input checked="" type="checkbox"/> Field Office	Gibson Station, RR 1, Hwy 64, Owensville, IN 47665		
<input type="checkbox"/> Temporary Job Site			

PROGRAM SCOPE

The licensee is authorized for possession and use of cesium-137 in sealed sources registered either with NRC under 10 CFR 32.210 or with an Agreement State and compatibly incorporated in Kay Ray Models 7062BP and 7063P, TN Technologies Models 5190 and 5211, and Ohmart Model SH-F1A gauging device source holders for density measurements with no single source to exceed the maximum activity specified in the certificate of registration issued by NRC or an Agreement State at a coal-fired electrical generating plant located as specified on the license. The licensee currently is in possession of 15 devices as authorized and all are currently in use with no gauges in storage and total activity is well below license limits. Staff consists of the local plant RSO, ARSO, and authorized instrument/maintenance staff at this plant. The licensee is not authorized to perform service, repair, or exchange/removal of gauges. The gauges are leak-tested and shutter tested on a six-month cycle.

Performance Observations

During the inspection, plant staff demonstrated/discussed: (1) signage; (2) security of licensed material; (3) gauge maintenance; (4) staff training; (5) leak test procedures and the latest leak tests performed on 4/14/06; and (7) shutter tests as specified in license condition No. 16.

Handwritten signature