

FAX COVER SHEET

Radiological Physics Service, Inc.
 3839 Napier Road
 Plymouth, MI 48170
 (734) 455-4730
 Fax (734) 453-8851

Send to: LUREN HUBZER	From: Ray Carlson
Fax Number: 630-829-9782	Date: 6/15/06

- Urgent
- Reply ASAP
- Please comment
- Please review
- For your information

Total pages, including cover: **2**

Comments:

Loren,
 I redid the leak tests here
 tonight.
 Ray

Radiological Physics Service, Inc.

Ray A. Carlson, M.S.
 President

Medical Nuclear & Diagnostic Physicist
 Certified American Board of Radiology
 American Board of Medical Physics

(734) 455-4730
 3839 Napier Road
 Plymouth, MI 48170

E-Mail: rcarlson@voyager.net
 Cell (734) 395-7361
 FAX (734) 453-8851

CWO-4 U.S.M.C. (Retired)

Radiological Physics Service, Inc.

Dan L. Marx, M.S., M.B.A.
 Medical Health Physicist

(734) 455-4730
 (734) 453-8851 (Fax)
 3839 Napier Road
 Plymouth, MI 48170

Cell: (810) 730-6004
 Home: (810) 631-8914

Major U.S. Army (Retired)

Location: South Allen Radiology - Flat Rock (Stored at RPS)

Date of Inspection: June 15, 2006

SEALED SOURCE LEAK TEST

Nuclide	Type	Location	Calibration Activity	Calibration Date	Mfgr.	M/N	S/N
Cs-137	vial	Hot Lab	265 uCi	09/02/85	Syncor	2I-2004	3194/28
Ba-133	vial	Hot Lab	311.4 uCi	10/01/82	Amersham	BDC.VI	2129 MA
Co-57	flood	Hot Lab	10 mCi	05/14/00	DuPont	NES-392	S3920813-02

COMMENT: The sources listed above were leak tested using a wipe technique and were found to have less than 0.005 uCi removable activity.

TESTED BY: K. A. C. [Signature] MS DATE: 6/15/06

RSO: _____