



11438 Cronridge Drive
Suite G
Owings Mills, Maryland 21117
Tel: (410) 363-8344
Fax: (410) 363-9978
E-mail: ctibaltimore@cti-consultants.com

June 13, 2006

United States Nuclear Regulatory Commission
Security and Industrial Branch
Division of Materials Safety
Region 1
475 Allendale Road
King of Prussia
PA 19406-1415

Q-2

Attn: Mr. Sattar Lodhr, PhD

03020195

Re: License Amendment (45-23031-01)

Dear Mr. Lodhr:

Attached are the resent leak tests results provided to me by Chesapeake branch.

Please let me know if I can provide any additional information.

Sincerely,

A. Kuriatnikov
RSO

RECEIVED
REGION 1
2006 JUN 19 AM 10:31

139013

NMOS/RONI MATERIALS-002

CHANTILLY, VA ■ RICHMOND, VA ■ ROCKVILLE, MD ■ CHARLOTTESVILLE, VA ■ BALTIMORE, MD
BLACKSBURG, VA ■ FREDERICKSBURG, VA ■ CHESAPEAKE, VA ■ WOODBRIDGE, VA



North East Technical Services, Inc.

75 Allerton Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

1407 Stephanie Way

Chesapeake, VA 23320

ATTN: Darrell Darrow

Shipping Address: 1407 Stephanie Way

Chesapeake, VA 23320

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model 3411 Gauge S/N 5155 Leak Test Date 5/30/2006

Source	Reading in microCuries
CC-2212	0.000088
CAA-1261	0.00000

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

6-5-06

Monday, June 05, 2006



#52

North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

1407 Stephanie Way
Chesapeake, VA 23320
ATTN: Darell Darrow

*Shipping Address: 1407 Stephanie Way
Chesapeake, VA 23320*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001	Gauge S/N	114402	Leak Test Date	5/30/2006
Source	Reading in microCuries				
1394CM	0.0000556				
NJ04640	0.00000				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

6-5-06

Monday, June 05, 2006



#53

*North East Technical Services, Inc.**75 Allerton Ct., Suite 4**Westminster, MD 21157**Ph: 410.751.5090 Fax: 410.751.5091**CTI Consultants - Chesapeake**1407 Stephanie Way**Chesapeake, VA 23320**ATTN: Darrell Darrow**Shipping Address: 1407 Stephanie Way**Chesapeake, VA 23320*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001 EZ	Gauge S/N	114403	Leak Test Date	5/30/2006
Source					
1396 CM			0.0000648		
NJ04641			0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Date:

5-5-06

Monday, June 05, 2006



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

*1407 Stephanie Way
Chesapeake, VA 23320
ATTN: Darell Darrow*

*Shipping Address: 1407 Stephanie Way
Chesapeake, VA 23320*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3430	Gauge S/N	22632	Leak Test Date	5/23/2006
Source		Reading in microCuries			
	75-4388	0.0000538			
	47-18476	0.00000			

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

5-25-06

Thursday, May 25, 2006

This is to acknowledge the receipt of your letter/application dated

6/13/2006, and to inform you that the initial processing which includes an administrative review has been performed.

☒ Adm. 45-23031-01
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 139013.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.

NRC FORM 532 (RI)
(6-98)

Sincerely,
Licensing Assistance Team Leader