

## JAMES A. SEWELL & ASSOCIATES

\*Civil Engineering \*Electrical Engineering \*Land Surveying

Phone WA (509) 447-3626 Fax (509) 447-2112

Phone ID (208) 437-2641

Sandpoint (208) 263 - 4160  
Fax 263 - 5229

# fax

To: Roberto J. Torres - U.S. Nuclear Regulatory  
Commission, Region IV Division of Nuclear  
Materials Safety Nuclear Materials Licensing

Fax #: 817-860-8188

817-810-8263

From: RAY HENRIKSEN - J. A. SEWELL & ASSOC.

Date: May 30, 2006

Subject: Certificate of Leak Test for J. A. Sewell & Assoc.

Pages: 4 including cover page

NOTES: Roberto,

We not longer have the CPN - MC-3. We have acquired another Troxler 3440. This happened the end of 2005 / the beginning of 2006.


If there are any problems please contact me.

Thanks, Ray

470938



Advanced Instrumentation for Density  
& Moisture Testing of Soils & Pavements



CPN International, Inc.  
2830 Howe Road  
Martinez, CA 94553 USA  
Phone: (925) 228-9770  
Fax: (925) 228-3183  
e-mail: cpn@cpn-intl.com

## CERTIFICATE OF LEAK TEST

JAMES A. SEWELL & ASSOC.

RSO: JOE OLMSTEAD

600 4TH ST WEST

NEWPORT, WA 99156

Customer No: 19280600

Date Sample Collected: 1/12/2005

Test Number: 42177

This is to certify that the leak test on the indicated source (s) was counted on 1/21/2005 and the results shown accurately represent the level of removable contamination.

Model: MC-3

Source(s)/Gauge Serial Number: M37087847

<u>Source</u>	<u>Activity</u>	* <u>Removable Contamination</u> <u>( in microCuries)</u>
AmBe 241	50 mCi	.0000033
Cs 137	10 mCi	.0000238

\* 0.005 microcuries (185 Bq) or greater constitutes a leaking source.


This Source : May Remain in use.

*Maria Lee*

Reviewed by: Leak Testing Quality Control  
CA License No. 1100-07



Advanced Instrumentation for Density  
& Moisture Testing of Soils & Pavements



CPN International, Inc.  
2830 Howe Road  
Martinez, CA 94553 USA  
Phone: (925) 228-9770  
Fax: (925) 228-3183  
e-mail: cpn@cpn-intl.com

## CERTIFICATE OF LEAK TEST

JAMES A. SEWELL & ASSOC.

RSO: JOE OLMSTEAD

600 4TH ST WEST

NEWPORT, WA 99156

Customer No: 19280600

Date Sample Collected: 6/9/2005

Test Number: 43184

This is to certify that the leak test on the indicated source (s) was counted on 6/17/2005 and the results shown accurately represent the level of removable contamination.

Model: 3440

Source(s)/Gauge Serial Number: 29270

<u>Source</u>	<u>Activity</u>	* <u>Removable Contamination</u> <u>( in microCuries)</u>
AmBe 241	40 mCi	.0000032
Cs 137	8 mCi	.0000187

\* 0.005 microcuries (185 Bq) or greater constitutes a leaking source.


This Source : May Remain in use.

*Maria Lee*

Reviewed by: Leak Testing Quality Control  
CA License No. 1100-07



Advanced Instrumentation for Density  
& Moisture Testing of Soils & Pavements



CPN International, Inc.  
2830 Howe Road  
Martinez, CA 94553 USA  
Phone: (925) 228-9770  
Fax: (925) 228-3183  
e-mail: cpn@cpn-intl.com

## CERTIFICATE OF LEAK TEST

JAMES A. SEWELL & ASSOC.

RSO: JOE OLMSTEAD

600 4TH ST WEST

NEWPORT, WA 99156

Customer No: 19280600

Date Sample Collected: 6/9/2005

Test Number: 43183

This is to certify that the leak test on the indicated source (s) was counted on 6/17/2005 and the results shown accurately represent the level of removable contamination.

Model: 3440

Source(s)/Gauge Serial Number: 28535

<u>Source</u>	<u>Activity</u>	* <u>Removable Contamination</u> <u>( in microCuries)</u>
AmBe 241	40 mCi	.0000032
Cs 137	8 mCi	.0000165

\* 0.005 microcuries (185 Bq) or greater constitutes a leaking source.

This Source : May Remain in use.

*Theresa Lau*

Reviewed by: Leak Testing Quality Control  
CA License No. 1100-07

470938