

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
34-26655-02

Docket or Reference Number
030-35067

Amendment No. 02

- B. through D. To be used for instrument calibration and testing.
- E. To be used for instrument calibration and testing for shielding evaluations.

CONDITIONS

10. A. Tests for leakage and/or contamination may be performed at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material. Analysis of leak test samples may be performed at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
- B. Licensed material listed in Subitems 6.B. through 6.D. may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. Licensed material shall be used by, or under the supervision of, Sharon L. Long, David Close, Samuel A. Pontillo, Michael W. Lairmore and Kelly Stoneberg.
12. The Radiation Safety Officer for this license is Sharon L. Long.
13. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed 6 months or at such other intervals as specified on the certificate of registration referred to in 10 CFR 32.210.
- B. In the absence of a certificate of a transfer indicating that a leak test has been made within 6 months prior to the transfer, sealed sources transferred from another person shall not be put into use until tested.
- C. Sealed sources need not be leak tested if they contain not more than 100 microcuries of beta and/or gamma emitting material and are in storage, and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
34-26655-02

Docket or Reference Number
030-35067

Amendment No. 02

- D. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
- E. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically licensed by the Commission or an Agreement State to Perform such services.
14. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license.
15. Sealed sources containing licensed material shall not be opened or sources removed from source holders by the licensee.
16. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
17. The licensee is authorized to hold radioactive material with a physical half-life of less than 120 days for decay-in-storage before disposal in accordance with the following conditions:
- A. Before disposal as ordinary waste, byproduct material shall be surveyed at the container surface with the appropriate survey meter when its probe isens the scale and with no interposed shielding to determine that its radioactivity is not distinguishable from background. All radiation labels shall be removed or obliterated.
- B. A record of each disposal permitted under this License Condition shall be retained for three years. The record must include the date of disposal, the date on which the byproduct material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.
18. Licensed material shall not be used in or on human beings.
19. The licensee shall develop, maintain and implement procedures for obtaining an agreement with customers outlining the responsibilities of both the customer and service provider when performing service operations at customer's facility.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
34-26655-02Docket or Reference Number
030-35067

Amendment No. 02


20. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated July 1, 2005; and
 - B. Facsimile dated September 1, 2005, with attached update dated August 30, 2005.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date JUN 26 2006

By


James R. Mullauer, M.H.S.
Materials Licensing Branch
Region III