

CONVERSATION RECORD

TIME | DATE  
6-15-2006

VISIT       CONFERENCE       TELEPHONE  
  
 INCOMING  
 OUTGOING

030-18220

NAME OF PERSON(S) CONTACTED OR IN CONTACT  
James H. Granger  
Radiation Safety Officer

ORGANIZATION (OFFICE, DEPT. ETC.)  
Granger and Associates, Inc.  
Cheboygan, Michigan

TELEPHONE NO.  
231-627-2763

SUBJECT

Control No. 315359 -- Re - Amendment request by letter dated March 29, 2006, regarding request to terminate their NRC license, License No. 21-20302-01.

SUMMARY

In a telephone conversation with Mr. Granger on this date, I related that we will need the following additional information before proceeding with their request:

1. Please provide a copy of leak test records (current as of the time of the transfer) for the sources in both gauges transferred to InstroTek, Inc., in March of 2006, which confirms the non-leaking status of all sources; and
2. Please confirm that no leaking sealed sources were ever possessed by the licensee, based on routine leak test data.

Licensee was requested to provide response, in writing, to me (as additional information to Control Number 315359) by facsimile at 630-829-9782. No follow-up hard copy is needed. Any questions may be directed to me at 630-829-9828.

ACTION REQUIRED

Licensee to respond as requested.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Loren J. Hueter



6-15-2006

ACTION TAKEN

SIGNATURE

TITLE

DATE