

CONVERSATION RECORD

TIME | DATE  
6-15-2006

VISIT       CONFERENCE       TELEPHONE  
  
 INCOMING  
 OUTGOING

| NAME OF PERSON(S) CONTACTED OR IN CONTACT    | ORGANIZATION (OFFICE, DEPT. ETC.)                   | TELEPHONE NO. |
|--|---|---------------|
| James H. Granger<br>Radiation Safety Officer | Granger and Associates, Inc.<br>Cheboygan, Michigan | 231-627-2763  |

**SUBJECT**  
Control No. 315359 - Re - Amendment request by letter dated March 29, 2006, regarding request to terminate their NRC license, License No. 21-20302-01.

**SUMMARY**  
In a telephone conversation with Mr. Granger on this date, I related that we will need the following additional information before proceeding with their request:

1. Please provide a copy of leak test records (current as of the time of the transfer) for the sources in both gauges transferred to InstroTek, Inc., in March of 2006, which confirms the non-leaking status of all sources; and *COPIES OF LEAK TESTS ATTACHED JWH*
2. Please confirm that no leaking sealed sources were ever possessed by the licensee, based on routine leak test data. *CONFIRMING NO LEAKING SEALED SOURCES JWH*

Licensee was requested to provide response, in writing, to me (as additional information to Control Number 315359) by facsimile at 630-829-9782. No follow-up hard copy is needed. Any questions may be directed to me at 630-829-9828.

**ACTION REQUIRED**  
Licensee to respond as requested.

| NAME OF PERSON DOCUMENTING CONVERSATION | SIGNATURE              | DATE      |
|---|------------------------|-----------|
| Loren J. Hueter                         | <i>Loren J. Hueter</i> | 6-15-2006 |

**ACTION TAKEN**

| SIGNATURE               | TITLE          | DATE          |
|-------------------------|----------------|---------------|
| <i>James H. Granger</i> | VICE PRESIDENT | JUNE 15, 2006 |

Jun 15 08 11:46a InstroTek, Inc.

9198758328

p.2

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 6/15/2006  
Test Number: 1

JOHNNY  
REFURBISHED GAUGES 2006  
5908 TRIANGLE DR  
RALEIGH, NC 27617

Phone: (919) 875 -8371

### LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

|              |                  |                               |
|--------------|------------------|-------------------------------|
| Gauge model: | Gauge S/N: 18527 | Leak Test Date: 5/ 2/2006     |
|              | <u>Source</u>    | <u>Reading in microCuries</u> |
|              | 47-13983         | 0.00000                       |
|              | 50-8081          | 0.00038                       |

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Johnny Wilkoff Date: 6-15-06

InstroTek, Inc.  
5908 Triangle Dr  
Raleigh, NC 27617  
(919) 875-8371 Fax (919) 875-8328

Date: 6/15/2006  
Test Number: 1

JOHNNY  
REFURBISHED GAUGES 2006  
5908 TRIANGLE DR  
RALEIGH, NC 27617

Phone: (919) 875 -8371

### LEAK TEST CERTIFICATE

NC Materials License # 092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

|               |                               |                           |
|---------------|-------------------------------|---------------------------|
| Gauge model:  | Gauge S/N: 16941              | Leak Test Date: 5/ 2/2006 |
| <u>Source</u> | <u>Reading in microCuries</u> |                           |
| 47-12364      | 0.00000                       |                           |
| 50-6110       | 0.00033                       |                           |

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Johnny Williams Date: 6-15-2006