

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303069397	Mfr.	Fisher-Scientific	Technician	blt
Asset No.	002471	Model	14-983-10B		
Serial No.	183305	Type.	THERMOMETER	Cal Date.	02-Jun-06
Remarks: No accuracy is provided by manufacturer. +/- 1 Degree C resolution is used as the accuracy specification.					
TOTAL IMMERSION					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	°C	°C	°C	°C	°C	Result
	-19.89	-20.5	-0.6	1.0	1.2	Pass
	0.04	0.0	0.0	1.0	1.2	Pass
	49.97	50.0	0.0	1.0	1.2	Pass
	99.90	100.9	1.0	1.0	1.2	Pass
END OF REPORT						