



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

June 27, 2006

Docket No. 03034820
Control No. 138861

License No. 29-30472-01

Gina Diaz
President
Bergenline Imaging Center
d/b/a Bergenline X-Ray Diagnostic Center
402 43rd Street
Union City, NJ 07087

SUBJECT: BERGENLINE IMAGING CENTER, REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO. 138861

Dear Mrs. Diaz:

This is in reference to your letter dated April 14, 2006 requesting to amend Nuclear Regulatory Commission License No. 29-30472-01. In order to continue our review, we need the following additional information:

1. In your letter, you have appointed Arnold Brenner, D.O. as your Radiation Safety Officer (RSO). Pursuant to 10 CFR 35.50 (c)(2), (d) and (e), please provide a written attestation statement by a preceptor radiation safety officer that the individual has satisfactorily completed the applicable requirements in 10 CFR 35.50 and has achieved a level of radiation safety knowledge sufficient to function independently as an RSO. Also, please provide a description of training in radiation safety, regulatory issues and emergency procedures for the types of use listed in the license. You may find Section 8.10 of NUREG-1556, Volume 9, Revision 1 useful in preparing your response.

Please note that pursuant to 10 CFR 35.24 a licensee may permit an authorized user to function as a temporary RSO for up to 60 days each year provided that the licensee takes the necessary actions required in 10 CFR 35.24 and notifies the commission in accordance with 10 CFR 35.14(b)

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Toolkit Index Page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 138861. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5143.

G. Diaz
Bergenline Imaging Center

2

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Farrah C. Gaskins

Farrah C. Gaskins
Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

cc:
Arnold Brenner, D.O., Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML061800216.wpd

SUNSI Review Complete: FGaskins

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	DNMS/RI	<input type="checkbox"/>		
NAME	FGaskins /FCG/							
DATE	6/27/06							

OFFICIAL RECORD COPY