

U.S. NUCLEAR REGULATORY COMMISSION		Date: 7/25/2006
TELEPHONE CONVERSATION RECORD		Time: 10:30 AM
Mail Control or Report No(s).	139131	License No(s). 45-31174-01 Docket No(s). 030-37309
Name of Licensee:	Volkert and Associates, Inc.	
Name of Participant(s):	William D. McDowall II, P.E. Kathy Modes, NRC	
Telephone No.	804-452-2862	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	new portable gauge license	
Summary:	<p>Please provide your fax number and email address for our records.</p> <p>Please re-submit following the exact wording in Appendix B of NUREG-1556, Volume 1, Revision 1. Your submission of 7/11/2006 had the incorrect leak test frequency of 6 months. The correct one is 12 months. We need you to confirm that when you are requesting the Model 3400 Series - you are requesting to use Models 3430, 3430-M, 3440, 3440-M, 3450 and 3451. You indicated that you would use equivalent training, but did not specifically state equivalent to Appendix D of NUREG-1556, Vol 1 Rev 1 as required by this guidance. Inventory records are to be retained for 5 years, not 3 as written in this application.</p> <p>Since Volkert is located in various states, please provide a copy of Volkert's agreement state license(s).</p>	
Action Required:	Wait for additional information	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SUNSI Review Completed By:	K Modes / RA / Date: 7/25/2006	