

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 7/25/2006

**TELEPHONE CONVERSATION RECORD**

Time: 2:30 PM

Mail Control 139066 License No(s) 37-30089-01 Docket No(s) 03033342  
or Report No(s).

Name of Licensee: Scrubgrass Generating Company, L.P.

Name of Participant(s): Bradley Knight, proposed RSO  
Kathy Modes, NRC

Telephone No. 814-385-6661 x15

Subject: Non-routine maintenance on fixed gauges

(NOTE: This will be used as the Documents Title in ADAMS)

Summary: Left a message:

Please provide Mr. Knight's phone, fax and email address for our records.

Please provide a copy of Mr. Knight's fixed gauge training courses.

Review License Condition 18 and inform us as to which of these non-routine maintenance activities you would like to be authorized to perform.

Please provide a sample calculation of a prospective evaluation using the highest activity source.

Monthly surveys are not acceptable. Surveys should be performed prior to, during and after any non-routine maintenance activity. Please revise this in your procedure.

Your letter dated June 27, 2006 did not include specific procedures. Please provide specific detailed procedures.

Action Required: Wait for response from licensee

Document Availability:  Publicly Available  Non-Publicly Available

Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright

Immediate Release  Normal Release  Delay Release Date

Prepared &amp; SUNSI Review Completed By: K Modes / RA / Date: 7/25/2006