



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



## Certificate of Calibration

Certificate #  
0972-01

**Submitted By:** DIV20  
**Address:** B57  
**Contact:** DON BANNON  
**Manufacturer Model:** OMEGA DP465-KC-MDSSD  
**Description:** TEMPERATURE METER  
**Serial No:** 3130900  
**Asset No:** 002524  
**Procedure:** DIGITAL THERMOMETERS - 9 MAY 2006

**Work Order:** 303070139  
**Date Issued:** Jul 6, 2006  
**Calibration Date:** Jul 6, 2006  
**\*Calibration Due:** Jan 6, 2007  
**Calibration Location:** Bldg. 64  
**Environment:** Temp. 73.0°F Hum. 40 %RH  
**\*\*Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.


Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.


**Remarks:** None

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
010329	7949003	FLUKE	525A	TEMPERATURE/PRESSURE CALIBRATOR	Oct 06, 06

Instrument calibration record for Omega Digital ThermoCouple Meter, Model # DP465KC-MDSSD, Serial # 3130900 (07/06/2006)

Reviewed by:   
Metrology Technician  
m:\a2la1.rpt Rev date August 15, 2005

  
Measurements by: Paul Depmore  
Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303070139	Mfr:	Omega	Technician:	PRD
Asset No:	002524	Model:	DP465	Cal Date:	06-Jul-06
Serial No:	3130900	Type:	Temperature Meter		
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Type K	°C	°C	°C	°C	°C	Result
Ch1	0	1	1	2	1.2	Pass
	-200	-198	2	2	1.2	Pass
	320	320	0	2	1.2	Pass
	840	840	0	2	1.2	Pass
	1372	1373	1	2	1.2	Pass
Ch2	1372	1373	1	2	1.2	Pass
Ch3	1372	1373	1	2	1.2	Pass
Ch4	1372	1373	1	2	1.2	Pass
Ch5	1372	1373	1	2	1.2	Pass
Ch6	1372	1372	0	2	1.2	Pass
Ch7	1372	1373	1	2	1.2	Pass
Ch8	1372	1372	0	2	1.2	Pass
Ch9	1372	1373	1	2	1.2	Pass
Ch10	1372	1372	0	2	1.2	Pass

END OF REPORT