

U.S. NUCLEAR REGULATORY COMMISSION

Date: 7/26/2006

TELEPHONE CONVERSATION RECORD

Time: 8:45 AM

Mail Control 139066 License No(s). 37-30089-01 Docket No(s). 03033342
or Report No(s).

Name of Licensee: Scrubgrass Generating Company, L.P.

Name of Participant(s): Bradley Knight, proposed RSO
Kathy Modes, NRC

Telephone No. 814-385-6661 x12

Subject: Non-routine maintenance on fixed gauges
(NOTE: This will be used as the Documents Title in ADAMS)

Summary: LEFT MESSAGE:

Mr. Knight returned my call and indicated that in Amendment No 4 of their license, he was authorized to perform installation, initial radiation surveys, relocation, removal from service, and alignment. After a review of the supporting document for this action I noted that I will need the following information to process this amendment request:

Mr. Knight's fax and email address for our records.

Confirmation that they would like to be authorized to perform: installation, initial radiation surveys, relocation, removal from service, and alignment.

Please provide a sample calculation of a prospective evaluation using the highest activity source.

Monthly surveys are not acceptable. Surveys should be performed prior to, during and after any non-routine maintenance activity. Please revise this in your procedure.

Specific procedures for each task identified above - can use instructions from manufacturer's training or instruction manual.

Action Required: Wait for response from licensee

Document Availability: ☒ Publicly Available ☐ Non-Publicly Available

☒ Non-Sensitive ☐ Non-Sensitive Copyright ☐ Sensitive ☐ Sensitive Copyright

☐ Immediate Release ☒ Normal Release ☐ Delay Release Date

Prepared & SUNSI Review Completed By: K Modes / RA / Date: 7/26/2006