



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20
Address: B57
Contact: DON BANNON
Manufacturer Model: OMEGA Type B
Description: THERMOCOUPLE
Serial No: 11116
Asset No: 011116
Procedure: TEMPERATURE PROBES - 5 JUNE, 2006

Work Order: 303070141
Date Issued: Jul 7, 2006
Calibration Date: Jul 7, 2006
***Calibration Due:** Jan 7, 2007
Calibration Location: Bldg. 64
Environment: Temp. 73.0°F Hum. 40 %RH
****Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.


*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.


Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: Cal'd at 900,1000,and1100 °C

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
010814	A44625	HART SCIENTIFIC	1529	THERMOCOUPLE THERMOMETER	Oct 26, 06
010813	9287	HART SCIENTIFIC	5650	THERMOCOUPLE	Sep 28, 06

Reviewed by: 
Metrology Technician

Measurements by: 
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order: 303070141	Mfr: OMEGA	Technician: BLT
Asset No: 011116	Model: TYPE B	
Serial No: 11116	Type: THERMOCOUPLE	Cal Date: 07-Jul-06
Remarks:		
Limits are based on ASTM E230-02		

Function/Range	Test Point	TI Reading	Difference	+/-Limits	+/-Uncertainty	Found/Left
TEMPERATURE	°C	°C	°C	°C	°C	Result
	899.332	898.84	0.49	4.5	2	Pass
	999.930	998.97	0.96	5.0	2	Pass
	1099.914	1099.94	-0.03	5.5	2	Pass

END OF REPORT