

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Plastatech Engineering, Ltd. 725 Morley Drive Saginaw, MI 48601		2. NRC/REGIONAL OFFICE REGION III US NUCLEAR REGULATORY COMMISSION 2443 WARRENVILLE ROAD, SUITE 210 LISLE, ILLINOIS 60532	
REPORT	2006-001		
3. DOCKET NUMBER(S) 030-34117	4. LICENSEE NUMBER(S) 21-26714-01	5. DATE(S) OF INSPECTION July 1, 2006	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Ed Kulzer	<i>E. D. Kulzer</i>	7/19/06

Docket File Information
**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Plastatech Engineering, Ltd.	2. NRC/REGIONAL OFFICE Region III
--	---

3. DOCKET NUMBER(S) 030-34117	4. LICENSE NUMBER(S) 21-26714-01	5. DATE(S) OF INSPECTION July 25, 2006
----------------------------------	-------------------------------------	---

6. INSPECTION PROCEDURES USED 87124	7. INSPECTION FOCUS AREAS 03.01 - 03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 03120	2. PRIORITY 5	3. LICENSEE CONTACT Crint LaBrosse	4. TELEPHONE NUMBER 989.754.6500
---------------------	------------------	---------------------------------------	-------------------------------------

Main Office Inspection Next Inspection Date: July 2011

Field _____

Temporary Job Site _____

PROGRAM SCOPE

The licensee is a manufacturing firm that employs 124 individuals. The licensee possesses one specifically licensed fixed gauge used in the production of roofing products. Currently, five individuals are involved with use and oversight of the devices including the RSO. Inventories and shutter checks are performed at required frequencies.

The RSO stated that the gauge is relocated, removed, etc. by the manufacturer. The licensee does not perform any service or maintenance activities on its gauges; these services are performed by the manufacturer.

Performance Observations

The plant facility was toured and the gauge was observed in conjunction with the licensee's current inventory. The RSO demonstrated survey techniques used if/when a gauge is relocated. Proper lock-out and leak test procedures were also adequately described. Interviews conducted with plant workers revealed that under no circumstances would a vessel be entered without RSO supervision of lock-out/tag-out procedures. The inspector performed independent and confirmatory radiation measurements which indicated similar results as noted in the licensee's survey records