Cardinal Health Nuclear Pharmacy Services 7000 Cardinal Place Dublin, OH 43017 614.757.5000 tel



www.cardinalhealth.com

P.7

July 26, 2006

RE:

Tom Thompson United States Nuclear Regulatory Commission, Region I 475 Allendale Road King of Prussia, PA 19406-1415

Terminated Radioactive Materials License Number 29-30867-01.

Dear Mr. Thompson:

As discussed over the phone, enclosed is the sealed source transfer record and sealed source leak test record for the cesium-137 source authorized under the above referenced license (now terminated. This was the only source authorized under the license, and all NRC regulated material has been removed from the facility. The facility will continue to utilize only accelerator produced material and will operate under New Jersey radioactive materials license number NJ-29-30867-01.

If there are any questions concerning this request, please contact me at 614.757.3116.

Sincerely,

Dave Breuning Health Physicist

Quality & Regulatory, Nuclear Pharmacy Services

CC: MRSO, Location 5862

Location File 5862 (3)

测则31 附2:

REGION 1

139002 NMSS/RGNI MATERIALS-002

PRS-20T

To: Cardinal Health NPS Location

30 Murray Hill Parkway, Suite

From: Cardinal Health PET

Manufacturing Services

30 Murray Hill Parkway, Suite

450 East Rutherford, I (201) 438-4044 NRC RAML#: 29-			PA 190' ford, <b>NJ</b> #:				
NO. PIECES ISOTOP	E CH	IEMICAL FORM/PHY	SICAL STA	ATE	Total Volume	Activity	
( Cs-137	Ser.# 1030-15-1	NA/Solid	Rod S	DUCE	NA	0.6951-	
1 (cs-137	Ser # 1014-60-6	NA/Solid NA/Solid	Q E-V	al	NA	204.2 m	
		7					
				-1			
Wipe test and packaging performed By Cardinal Health PET Manufacturing Services Date: 7 (0)06	<u></u>			Received By Cardinal He Nuclear Pha	aith	Services	

RADIOACTIVE MATERIAL TRANSFER RECORD

Shipper

Number: <u>071006</u> 3

Cardinal Health Nuclear Pharmacy Services

## Sealed Source Leak Test (RS-40L)

Facility:

**EAST RUTHERFORD** 30 MURRAY HILL PKWY

SUITE #400 PIC - ABDUL KAMARA EAST RUTHERFORD, NJ 07073

Test Date/Time: 7/11/2006 1:41PM

Next True (1772/007 4/44/79)

Test Instrument:

Make:Spectrum Techniques

Model:ICS-10 Serial # 100501

Instrument Location: Restricted Area

Isotope

Efficiency **Test Date**  MDA in

MDA In

Counting Efficiency

Cs-137

06/07/2006

1.266E-004

281

18.25 %

Leak Test Data:

Source Information: Cs-137 Make: Isotope Products Lab Model: RV-137-200-U Serial # 1014-60-6 Source Type: Vial Reference Srce										
Test [	<u>Data:</u> Calib. Activity	Calib. Date	Current Activity	Count Time(Min)	Gross Counts	Background Counts	Net Sample CPM	Test Results	Status	
	200.00 uCi	01/01/2004	188.69 <b>u</b> Ci	1.00	99	87.00	12	< 0.005 uCi	PASS	

Performed By: CDG Candice Goodyear Comments and Recommendations:

> Initials: **RSO Signature:**