

# RI - DNMS Licensee Event Report Disposition

Licensee: Bayhealth Medical Center

Event Description: Medical Event

License No: 07-14850-01      Docket No: 03007565      MLER-RI: 2006-024

Event Date: 06/12/06      Report Date: 6-12-06      HQ Ops Event #: 42634

**1. REPORTING REQUIREMENT**

<input type="checkbox"/>	10 CFR 20.1906 Package Contamination	<input type="checkbox"/>	10 CFR 30.50 Report
<input type="checkbox"/>	10 CFR 20.2201 Theft or Loss	<input checked="" type="checkbox"/>	10 CFR 35.3045 Medical Event
<input type="checkbox"/>	10 CFR 20.2203 30 Day Report	<input type="checkbox"/>	License Condition
<input type="checkbox"/>	Other _____		

**2. REGION I RESPONSE**

<input checked="" type="checkbox"/>	Immediate Site Inspection	Inspector/Date	
<input checked="" type="checkbox"/>	Special Inspection	Inspector/Date	<u>Gabriel/Simmons 6-16-06</u>
<input checked="" type="checkbox"/>	Telephone Inquiry	Inspector/Date	<u>Gabriel 6-13-06</u>
<input checked="" type="checkbox"/>	Preliminary Notification/Report	<input type="checkbox"/>	Daily Report
<input checked="" type="checkbox"/>	Information Entered in RI Log	<input type="checkbox"/>	Review at Next Inspection
<input type="checkbox"/>	Report Referred To	_____	

**3. REPORT EVALUATION**

<input checked="" type="checkbox"/>	Description of Event	<input checked="" type="checkbox"/>	Corrective Actions
<input checked="" type="checkbox"/>	Levels of RAM Involved	<input checked="" type="checkbox"/>	Calculations Adequate
<input checked="" type="checkbox"/>	Cause of Event	<input checked="" type="checkbox"/>	Additional Information Requested from Licensee

**4. MANAGEMENT DIRECTIVE 8.3 EVALUATION**

<input type="checkbox"/>	Release w/Exposure > Limits	<input type="checkbox"/>	Deliberate Misuse w/Exposure > Limits
<input type="checkbox"/>	Repeated Inadequate Control	<input type="checkbox"/>	Pkgng Failure > 10 rads/hr or Contamination > 1000x Limits
<input type="checkbox"/>	Exposure 5x Limits	<input type="checkbox"/>	Large# Indivs w/Exp > Limits or Medical Deterministic Effects
<input type="checkbox"/>	Potential Fatality	<input type="checkbox"/>	Unique Circumstances or Safeguards Concerns
<input type="checkbox"/>	If any of the above are involved:		
<input type="checkbox"/>	Considered Need for IIT	<input type="checkbox"/>	Considered Need for AIT
<input type="checkbox"/>	Decision/Made By/Date _____		

**5. MANAGEMENT DIRECTIVE 8.10 EVALUATION (additional evaluation for medical events only)**

<input checked="" type="checkbox"/>	Timeliness - Inspection Meets Requirements (5 days for overdose / 10 days for underdose)
<input checked="" type="checkbox"/>	Medical Consultant Used-Name of Consultant/Date of Report: <u>Dr. Nag, 7-3-06</u>
<input type="checkbox"/>	Medical Consultant Determined Event Directly Contributed to Fatality
<input type="checkbox"/>	Device Failure with Possible Adverse Generic Implications
<input type="checkbox"/>	HQ or Contractor Support Required to Evaluate Consequences

**6. SPECIAL INSTRUCTIONS OR COMMENTS**

\_\_\_\_\_

Non-Public

Inspector Signature: Sandra Gabriel

Date: 7-12-06

Public-SISP REVIEW COMPLETE

Branch Chief Initials: [Signature]

Date: 7/29/06

Oncology Services

Department of Radiation Oncology

793 South Queen Street  
Dover, DE 19901  
(P) 302-674-4401  
(F) 302-674-4129

21 West Clarke Avenue  
Milford, DE 19963  
(P) 302-430-5300  
(F) 302-430-5023

**REPORT AND NOTIFICATION TO NRC OF MEDICAL EVENT**

License Number: 07-14850-01

Licensee's Name: Bayhealth Medical Center  
640 South Queen Street  
Dover, DE 19904

Authorized User: John Lahaniatis M.D.

Date / Time of Medical Event: 6/12/2006, 3:00PM

**Description of Event:**

On 6/12/06 at approximately 3:00 PM, an I-125 prostate seed implant was performed in the operating room at Bayhealth Medical Center/Kent General Hospital. At the end of the procedure John Lahaniatis, MD, Radiation Oncologist and authorized user, reviewed the documentation and wrote a note on the Operative Note Sheet indicating the isotope type, number of seeds, number of needles used, individual seed activity, total activity implanted and other patient-related parameters. At this time, it was determined that the implanted activity was different from the planned activity; the total activity calculated was 27 mCi and implanted activity was 34 mCi. This dose was 26% higher than intended.

**Reason Why Event Occurred:**

The event occurred due to incorrect entry of units of dose and activity into the treatment planning computer system.

With prostate seed implants, the total number and placement of radioactive seeds and subsequent computer calculations for dose distribution within the patient are performed in "real time." In order to do this computerized calculation as to how the seeds will be implanted, a physicist using a laptop in the Operating Room, enters specific information related to the type and radioactivity of the seeds, manufacturer, etc.

In this regard there are two options from which to choose: mCi or U. When seed activity is entered in either of the above option "boxes," the alternate unit is calculated and displayed on the following line. It is at this point that the software provides the opportunity to compare the numbers for agreement with the seed activity received from the vendor.

The seed activity ordered was 0.34 mCi per seed. However, the activity entered in software was 0.34 U per seed. The discrepancy was not noted. Therefore, the computer

RECEIVED  
REGION 1  
JUN 12 2006  
10:29 AM

calculated the number of seed required for activity 0.27 mCi per seed, while the activity actually implanted was 0.34 mCi per seed.

It was determined that the total activity implanted was 34 mCi which is 26% more than the planned activity of 27 mCi.

**Effect on Patient:**

Patient may be at risk for long term complications related to higher than planned dose to the rectum and urethra. Therefore the patient will be monitored by Drs. Vallorosi and Lahaniatis on a long-term basis, including physical examinations and appropriate imaging and other medical studies as needed.

**Actions Taken to Prevent Recurrence:**

To confirm radioactive seed information accuracy, a "time out" will be incorporated in the policy #B9810.24 titled "Prostate Implants using I-125 or Pd-103."

The "time out" will take place using a checklist to verify that the correct isotope and source strength have been entered. A second person—a physician, second physicist, dosimetrist or nurse—will verbally call out the numbers entered on the laptop screen. The physicist will then compare those numbers to the vendor's calibration certificate. The Radiation Oncologist will proceed only when the prescribed radioactivity data has been confirmed in the calculations.

**Certificate that Licensee Notified Patient:**

Letter was sent to patient on 6/16/06 from Dr. Lahaniatis.

**Actions Taken:**

All variations in dose delivery above 20% are considered misadministration per NRC 10CFR35.3045(a)(1)(ii) and NRC needs to be notified immediately.

The following actions were taken on 6/12/06:

1. Participating Urologist, Christopher Vallorosi, MD was notified in-person by John Lahaniatis, MD Radiation Oncologist immediately following the implant.
2. Rachael Taylor, MD, Radiation Safety officer was notified in-person by Dr. Lahaniatis.
3. Dr. Lahaniatis notified hospital administration by telephoning Donna Stinson, Administrative Director of Operations, Oncology Service line at 4:00 PM on 6/12/06.
  - a. Ms. Stinson interviewed Sapna Paramale, Medical Physicist at 4:10 PM in the Dover Cancer Center for her description of the event.
4. Drs. Lahaniatis and Vallorosi met with patient's wife and described the event by 4:30 PM.

5. Raji Subramanyam, PhD, Chief Medical Physicist telephoned the NRC King of Prussia office at approximately 5:00 PM and was instructed by voice message to contact the Washington, DC office, where a verbal report was taken.
6. Dr. Lahaniatis telephoned the Seattle Prostate Institute. During the follow-up telephone conversation on 6/13/06, a Seattle-based Radiation Oncologist, requested the post implant dosimetry (with patient identifiers removed) to be sent to him via overnight express service for his review. He agreed to telephone Dr. Lahaniatis with his assessment before 6/16/06.
7. JoAnn Davis, Director of Risk Management was also notified of event by 6:00 PM.

Rachel Taylor M.D.  
Radiation Officer

A handwritten signature in cursive script that reads "Rachel Taylor M.D.".

JUN 26 2006