www.exeloncorp.com

Exelon Generation 4300 Winfield Road Warrenville, IL 60555

RS-06-106

10 CFR 50.90

August 4, 2006

U. S. Nuclear Regulatory Commission ATTN: Document Control Desk Washington, D. C. 20555

> LaSalle County Station, Units 1 and 2 Facility Operating License Nos. NPF-11 and NPF-18 NRC Docket Nos. 50-373 and 50-374

- Subject: Additional Information Supporting the License Amendment Request to Technical Specification 3.7.3, "Ultimate Heat Sink"
- References: 1. Letter from K. R. Jury (Exelon Generation Company, LLC) to U.S. NRC, "Request for a License Amendment to Technical Specification 3.7.3, Ultimate Heat Sink," dated March 13, 2006
  - U.S. NRC to C. M. Crane (Exelon Generation Company, LLC), "LaSalle County Power Station, Units 1 and 2 – Request for Additional Information Related to Ultimate Heat Sink License Amendment Request," dated June 15, 2006
  - 3. Letter from J. A. Bauer (Exelon Generation Company, LLC), "Additional Information Supporting the License Amendment Request to Technical Specification 3.7.3, "Ultimate Heat Sink," dated July 13, 2006

In Reference 1, Exelon Generation Company, LLC, (EGC), requested an amendment to Appendix A, Technical Specifications (TS), of Facility Operating License Nos. NPF-11 and NPF-18 for LaSalle County Station (LSCS) Units 1 and 2 respectively. Specifically, the proposed change increases the temperature limit of the cooling water supplied to the plant from the Core Standby Cooling System (CSCS) pond (i.e., the Ultimate Heat Sink (UHS)) from  $\leq 100^{\circ}$ F to  $\leq 101.5^{\circ}$ F. This increase is achieved by reducing the temperature measurement uncertainty by replacing the existing thermocouples with higher precision temperature measuring equipment.

In Reference 2, the NRC requested additional information to complete the review of the proposed license amendment. In Reference 3, EGC provided the additional information requested.



August 4, 2006 U. S. Nuclear Regulatory Commission Page 2

In an email dated July 19, 2006, from Stephen Sands, NRC Project Manager to Alison Mackellar, EGC Licensing Engineer; additional information was requested to complete the review of the proposed license amendment. A telephone conference was held on July 20, 2006, between the NRC and EGC to discuss the additional questions. At that meeting EGC agreed to provide the following information to the NRC.

- 1. Vendor data sheets for the newly installed equipment used to measure the UHS temperature
- 2. A more detailed calculation of the uncertainty error

Attachment 1 provides the detailed calculation requested including vendor supplied technical data.

Note that the purpose of the detailed calculation is to confirm that the original uncertainty calculation, provided in Reference 3, remains bounding. The detailed calculation was performed in accordance with EGC methodology for instrument uncertainty for safety related indicating loops.

EGC has reviewed the information supporting a finding of no significant hazards consideration that was previously provided to the NRC in Attachment 1 of Reference 1. The supplemental information provided in this submittal does not affect the bases for concluding that the proposed license amendment does not involve a significant hazards consideration.

Should you have any questions concerning this letter, please contact Ms. Alison Mackellar at (630) 657-2817.

I declare under penalty of perjury that the foregoing is true and correct. Executed on the 4th day of August 2006.

Respectfully,

sain MBerry k

Darin M. Benyak Manager – Licensing

Attachment 1: LSCS Calculation L-003230, "CW Inlet Temperature Uncertainty Analysis"

Attachment 1

LaSalle County Station Calculation L-003230

CW Inlet Temperature Uncertainty Analysis

## ATTACHMENT 1 Design Analysis Cover Sheet

				Last P	age No. 14
Analysis No.	L-003230		Revision 000		
EC/ECR No.	361689		Revision 000		
Title:	CW Inlet Temperature L	<b>Uncertainty Anal</b>	ysis		
Station(s)	LaSalle		Compo	onent(s)	- <u> </u>
Unit No.:	1, 2		1TE-CW010	2TE-CW010	)
Discipline	1 & C		1TE-CW011	2TE-CW011	
Description Co Keyword	o <b>de/</b> 104		1TT-CW010	2TT-CW010	)
Safety Class	NSR		1TT-CW011	2TT-CW011	
System Code	CW		U1 Computer Point F285	U2 Compute	er Point F285
Structure	N/A		U1 Computer Point F286	U2 Compute	er Point F286
	CONT	ROLLED DOCI	UMENT REFERENCES		
Document No.		From/To	Document No.		From/To
				×	
Is this Design	Analysis Safeguards?		Yes 🗌 No 🛛	,	
Does this Desi	gn Analysis Contain Un	verified Assun	nptions? Yes 🗌 No 🛛	ATI/AR#	
Is a Suppleme	ntal Review Required?		Yes 🗌 No 🛛	If yes, compl Attachment 3	ete 3
Preparer T.	J. Van Wyk	·····	Mayun		8/2/06
Roviewer V	Print Name		Sign Name		Date
neviewei v.	Print Name		Sign Nome		8/2/06
Method of Revie	ew 🕅 Detailed	Review	Alternate Calculations	Testing	Date
<b>Review Notes:</b>					
Approver	1 1. Laharan		La F		8-2-06
	Print Name		Sign Name		Date
(For External Analyses On Exelon Reviewe	er N/A	$\sim$			
	Print N	lame	Sign Name	9	Date
Approver N/	4				
Departmention of D	Print Name	••••••••••••••••••••••••••••••••••••••	Sign Name		Date
Description of nevision (list affected pages for partials):					

THIS DESIGN ANALYSIS SUPERCEDES:

## CALCULATION TABLE OF CONTENTS

CALCULATION NO. L-003230	Revision 000	PAGE NO. 2
SECTION:	PAGE NO.	SUB-PAGE NO.
TABLE OF CONTENTS	2	
PURPOSE / OBJECTIVE	3	
METHODOLOGY AND ACCEPTANCE CRITERIA	3	
ASSUMPTIONS AND LIMITATIONS	3	
DESIGN INPUT	3	
REFERENCES	6	
CALCULATIONS	7	
SUMMARY AND CONCLUSIONS (Total Error)	13	
ATTACHMENTS: A. Minco® Quotation 160056-2, January 26, 2006	A1	
B. Minco® Drawing S100995, dated 4/27/99	B1	
C. E-mail from Keith Jensen of Minco® to Vikram Shah of LaSalle dated 7/25/06	C1	
D. ifm efector600® TR2432 Operating Instructions, 701724/01, dated 02/04 (Partial)	D1 – D2	
E. Letter from Ameera Shah of ifm efector to Vikram Shah of LaSalle dated 7/26/06	E1	
F. Fluke® 45 Dual Display Multimeter User's Manual, Rev. 4, dated 07/97 (Specification Page only)	F1	
G. SOLA® SDN Power Supplies Specifications for SDN 2.5-24-100P	G1	
H. RTP® RTP2000 Setup and Installation Guide, UG-2000-001, dated 9/12/02 (Partial)	H1	
I. Minco Report of Calibration for Platinum RTD, Model S100995PD, Serial No. P/N366 (Partial)	1 –  2	
J. HP 34401A Multimeter User's Guide, Edition 4, printed February 1996 (Specification Page only)	J1	

CALCULATION NO.	L-003230	<b>Revision 000</b>	PAGE NO. 3 of 14

### 1 PURPOSE / OBJECTIVE

- 1.1 The purpose of this calculation is to evaluate the loop uncertainty for the CW Inlet Temperature Indication Loops. These are revised instrument loops that were implemented by EC359060 for Unit 1 and EC359114 for Unit 2.
- 1.2 These instrument loops provide Ultimate Heat Sink (UHS) temperature indication via the Plant Process Computer (PPC). These new loop configurations replaced the existing thermocouples 1(2)CW010/011 (the sensing elements for computer points F285/F286) with new RTD temperature sensing elements and new temperature compensators (transmitters), and relocated the computer inputs to the appropriate Input/Output (I/O) analog input cards.

## 2 METHODOLOGY AND ACCEPTANCE CRITERIA

- 2.1 The methodology used for this calculation is based on NES-EIC-20.04 "Analysis of Instrument Channel Setpoint Error and Instrument Loop Accuracy", Rev. 4 (Reference 5.1.2). Additionally, for calculating the average uncertainty using up to four indicating loops, the multiple test criterion of ASME PTC 19.1 (Ref. 5.1.4), Section 3.2 was used.
- 2.2 The instrumentation evaluated in this calculation provides indication (via the Plant Process Computer) for Ultimate Heat Sink Temperature. This is a non-safety indication loop, but the indication is used to verify the Technical Specification SR 3.7.3.1 is met. In accordance with Reference 5.1.2, Appendix D, a Level 3 evaluation is appropriate for this analysis. However, in response to questions during the NRC review of the License Amendment Request to increase the UHS temperature surveillance requirement value, this analysis will evaluate all uncertainty terms and determine the total uncertainty value using methodology consistent with safety-related indicating loops (Reference 5.1.2, Appendix D, Level 2).
- 2.3 Temperature, humidity and pressure errors, when available from the manufacturer, are to be evaluated with respect to the conditions specified in the station EQ Zones. If not provided, an evaluation must be made to ensure that the environmental conditions are bounded by the manufacturer's specified operational limits. If the environmental conditions are bounded, these error effects are considered to be included in the manufacturer's reference accuracy.
- 2.4 Published instrument vendor specifications are considered to be based on sufficiently large samples so that the probability and confidence level meets the 2σ criteria, unless stated otherwise by the vendor (Reference 5.1.2, Appendix A, Section 8.0).
- 2.5 For normal error analysis, normal vibrations and seismic effects are considered negligible or capable of being calibrated out in accordance with Appendix I of Reference 5.1.2.
- 2.6 The calibration standard error is considered negligible; the calibration standard error (STD) is more accurate than the M&TE by a ratio of at least 4:1 (Reference 5.1.2, Appendix A, Section 5.1.4).
- 2.7 The insulation resistance error is considered negligible unless the instrumentation is expected to operate in an abnormal or harsh environment (Reference 5.1.2, Appendix A, Section 7.0).
- 2.8 Reference 5.1.2, Appendix I states that the effects of normal radiation are small and accounted for in the periodic calibration process. Outside of containment during normal operation, the uncertainty introduced by radiation effects on components is considered to be negligible.

#### 3 ASSUMPTIONS AND LIMITATIONS

- 3.1 Evaluation of M&TE errors for the digital multimeter is based on the assumption that the test equipment listed in Section 4.5 is used.
- 3.2 It is assumed that the calibration standard of the equipment utilized is more accurate than the M&TE equipment by a ratio of at least 4:1 such that the calibration standard errors can be considered

CALCULATION NO.	L-003230	Revision 000	PAGE NO. 4 of 14

negligible with respect to the M&TE specification per Section2.6. This is considered a reasonable assumption since M&TE equipment is certified to its required accuracy under laboratory conditions.

#### 4 DESIGN INPUTS

- 4.1 The new instrument loops will consist of the following components: high accuracy RTD temperature elements, temperature transmitters, precision input resistors at the field input to the I/O card, and the D/A conversion in the PPC I/O equipment. The loop components evaluated in this document have the following specifications:
- 4.1.1 New Minco RTDs in the existing thermowells (replacing the existing thermocouples). The new RTDs have the following performance specifications (Ref. 5.4.1):

Repeatability: ±0.2°F

[The RTDs are designed to EN60751 Class A specifications with high precision and repeatability requirements. Thus, this specification could be considered to be at a  $3\sigma$  confidence level. However, for conservatism, this specification will be used as a  $2\sigma$  value.]

Drift: ±0.1°F/year (Ref. 5.4.3)

[The study in Reference 5.5.3 shows that RTDs are inherently stable, and after the first few months following installation RTDs attain a stable condition from which it may not drift sufficiently to exceed accuracy limits. RTD cross-calibration is performed to identify if an element has experienced significant drift. Although the RTDs are not separately calibrated, for conservatism the vendor's drift value will be expanded using the loop calibration interval of 4 years (+ 1 year late factor).]

- 4.1.2 The resistance value equivalent to the temperature value of interest (101.5°F) for the RTDs was obtained from the Minco calibration reports for the RTDs installed at LaSalle (Ref. 5.4.10). The highest of the four resistance values was 115.013Ω. This value will be used to determine the M&TE error for the indicating loop (applied to Module 2). The change in resistance per 1°F change in temperature (0.214Ω/°F) was also obtained using the actual resistance values in the calibration reports for 101.5°F and 102.5°F.
- 4.1.3 New ifm® efector600 TR2432 temperature transmitter modules. These new modules have the following performance specification (Ref.5.4.4, 5.4.5):

Accuracy (includes drift):	±0.54°F / 2 years
"Temperature Drift":	±0.1% of measured range/ 10°C

[Note: Ref. 5.4.5 indicates that the accuracy specification includes drift error and is warranted to hold the accuracy and drift within the specified value for 2 years. It further states that testing is performed on 100% of the devices after production to verify conformance with these specifications. Therefore, these values are 3 $\sigma$  confidence level. It also states that the accuracy specification includes the resolution error and electronic component drift, and that there are no other environmental influences that will affect the accuracy specification.]

4.1.4 PPC I/O input card. The I/O input cards have the following performance specification (Ref.5.4.9): [2σ]

Accuracy:

- ±0.025% of full scale (30°F to 120°F)
- 4.2 RTD extension wire has the identical conductor types as the RTD, and therefore there is no emf drop or change in conductor size at the point of connection on the RTD (Ref. 5.4.2).

CAL	CULATION NO. L-00;	3230	Revision 000	<b>PAGE NO.</b> 5 of 14
4.3	The Instrument Loop p performance specification	ower supply is a SOL ions: [2ơ]	A® SDN 2.5-24-100P (Ref. 5	.4.8), which has the following
	Output tolerance: Temperature range: Humidity:	± 2% overall (com -10°C to 60°C < 90% RH, non-co	bined Line, load, time, and ter	mperature related changes)
4.4	The precision signal representation with a tolerance of $\pm 0.0$	sistor at the input tern 02% (Reference 5.3.2	ninals of the I/O card (Module ?) [2σ]	3) is a high-precision resistor
4.5	The loop is calibrated u	sing a variable resist	ance input (to simulate the RT	D input), measured with either

- a Fluke 45 DMM or an HP 34401A, and reading the indicated temperature at the PPC. The calibration procedures (Ref. 5.2.1 and 5.2.2) each specify that one loop will be calibrated using either the Fluke 45 OR the HP 34401A. The other loop must be calibrated using the other DMM.
- 4.5.1 Reference Accuracy for the Fluke 45 (medium speed) on the  $300\Omega$  range is:

 $(\pm 0.05\% \text{ reading} + 2 \text{ LSD} + 0.02\Omega)$  (Ref. 5.4.6) [2**σ**]

4.5.2 Reference Accuracy for the HP 34401A on the  $1k\Omega$  range is:

> ± (0.01% reading + 0.001% range) (Ref. 5.4.7) [2σ]

Temperature coefficient for the HP 34401A on the 1kΩ range is (for 0°C to 18°C and 28°C to 55°C):

± (0.0006% of reading + 0.0001% of range /°C) (Ref. 5.4.7) [2σ]

#### 4.6 LOCAL SERVICE ENVIRONMENTS (Ref. 5.5.2)

Table 4.6			
$\searrow$	RTDs	Ifm efector600 TR2432	Plant Process Computer
EQ Zone		H7	C1A
Location		Turbine Bldg	Control Room (Computer Room)
Temperature		83°F to 102°F	50 to 104°F (Normal: 65 to 85°F)
Pressure		0 "wc	0.125 to +3.0 "wc
Humidity		39 to 47% RH	2.6 to 90% RH [see note below]
Note: Per refere	nce 552 the no	mal expected humidity in this zone	is 20 to 50% PU1

[Note: Per reference 5.5.2, the normal expected humidity in this zone is 20 to 50% RH]

#### 4.7 **Calibration Tolerance**

The calibration tolerance for these indication loops is  $\pm 0.54$ °F. Per Ref. 5.1.2, this is a  $3\sigma$  value.

## CALCULATION NO. L-003230

**Revision 000** 

#### 5 <u>REFERENCES</u>

#### 5.1 METHODOLOGY

- 5.1.1 ANSI/ISA-S67.04-Part 1-1994, "Setpoints for Nuclear Safety Related Instrumentation"
- 5.1.2 NES-EIC-20.04, "Analysis of Instrument Channel Setpoint Error and Instrument Loop Accuracy," Revision 4
- 5.1.3 ANSI/ISA TR67.04.09, "Graded Approaches to Setpoint Determination," dated 10/15/05
- 5.1.4 ASME PTC 19.1, Part 1, "Measurement Uncertainty," 1985

## 5.2 **PROCEDURES**

- 5.2.1 LIP-CW-501 [New loop-specific calibration procedure in development; tracked by CAP process]
- 5.2.2 LIP-CW-601 [New loop-specific calibration procedure in development; tracked by CAP process]

#### 5.3 LASALLE STATION DRAWINGS

- 5.3.1 1 E-1(2)-4022ZC "Schematic Diagram, Circulating Water System CW Pt. 3," as revised by EC359060 and EC359114.
- 5.3.2 1 E-1(2)-4707AA, "Wiring Diagram Analog Input Cabinet 1(2)C91-P607 AITs 1,2,3,4 Left Side," as revised by EC359060 and EC359114.

#### 5.4 VENDOR PRODUCT INFORMATION

- 5.4.1 Minco® Quotation 160056-2, January 26, 2006
- 5.4.2 Minco® Drawing S100995, dated 4/27/99
- 5.4.3 E-mail from Keith Johnson or Minco® to Vikram Shah of LaSalle dated 7/25/06
- 5.4.4 ifm efector600® TR2432 Operating Instructions, 701724/01, dated 02/04
- 5.4.5 Letter from Ameera Shah of ifm efector to Vikram Shah of LaSalle dated 7/26/06
- 5.4.6 Fluke® 45 Dual Display Multimeter Users Manual, Revision 4, dated 07/97
- 5.4.7 HP 34401A Multimeter User's Guide, Edition 4, printed February 1996
- 5.4.8 SOLA® SDN Power Supplies Specifications for SDN 2.5-24-100P
- 5.4.9 RTP® 8436 Series Analog Input Cards Technical Manual, 981-0021-211A, Rev. A, dated 04-96
- 5.4.10 Minco Report of Calibration for Platinum RTD, Model S100995PD, Serial No. P/N366

## 5.5 OTHER REFERENCES

- 5.5.1 LaSalle Technical Specifications, Sections 3.7.3, B 3.7.3, Amendments 178/164
- 5.5.2 LaSalle UFSAR, Rev. 16, Tables 3.11-18 and 3.11-24
- 5.5.3 EPRI TR-103099, "Effects of Resistance Temperature Detector Aging on Cross-Calibration Techniques," Final Report dated June 1994

CALCULATION NO. L-003230

Revision 000

**PAGE NO.** 7 of 14

6 <u>CALCULATIONS</u>

#### 6.1 RTD ERRORS (MODULE 1)

- 6.1.1 Random Errors σ1
- 6.1.1.1 RTD Reference Accuracy RA1

The RTD Reference Accuracy is  $\pm 0.2^{\circ}F$  (Section 4.1.1). This is a  $2\sigma$  value. RA1<sub>2 $\sigma$ </sub> =  $\pm 0.2^{\circ}F/2$ 

 $RA1 = \pm 0.1^{\circ}F$ 

6.1.1.2 RTD Calibration Error CAL1

The RTDs are not separately calibrated. Therefore, there is no calibration tolerance for this module. (The loop calibration tolerance is applied to Module 2, which is the module that is adjusted during loop calibration.)

CAL1 = 0

6.1.1.3 RTD Setting Tolerance ST1

The RTDs are not separately calibrated. Therefore, there is no setting tolerance for this module. (The loop calibration tolerance is applied to Module 2, which is the module that is adjusted during loop calibration.)

ST1 = 0

6.1.1.4 Random Input Errors σ1in

The RTDs are the first modules in the loop. Therefore,

σ1in = 0

6.1.1.5 Drift Error D1

The RTD Drift value (IDE) specified by the vendor is  $\pm 0.1^{\circ}$ F/year. [2 $\sigma$ ] The RTDs are not separately calibrated: RTD cross-calibration is performed to identify if an RTD has experienced significant drift. For conservatism the vendor's drift value will be expanded using the loop calibration interval (Section 4.1.1). The interval for these indicating loops is 4 years. The 25% late factor is 1 year. (VDP is the vendor drift period, or 1 year in this case.)

$D1_{2\sigma}$	=	[IDE] x [(SI + LF)/VDP)] <sup>1/2</sup>
	=	$[0.1^{\circ}F] \times [(4 \text{ years} + 1 \text{ year})/1 \text{ year}]^{1/2}$
	=	0.1°F x 2.236
	=	0.224°F
D1	=	0.112°F

#### 6.1.1.6 RTD Random Error o1

	=	± 0.150 °F
	=	$\pm [(0,12F)^{-} + (0)^{-} + (0)^{-} + (0)^{-} + (0)^{-} + (0,112)^{-}]^{-}$ $\pm 0.150 ^{\circ}F$
σ1		$\pm [(RA1n)^{2} + (CAL1)^{2} + (ST1)^{2} + (\sigma_{1}in)^{2} + (D1)^{2}]^{1/2}$

CALCULATION NO. L-003230		<b>Revision 000</b>	PAGE NO. 8 of 14	
				1
6.1.2	Non-Random Errors <b>2e1</b>			

RTDs are passive devices that produce a resistance signal proportional to temperature. As such, they are not affected by the following non-random effects.

Humidity Effects:	eH1 = 0
Static Pressure Effects:	eSP1 = 0
Ambient Pressure Effects:	eP1 = 0
Power Supply Effects:	eV1 = 0
Seismic Effects:	eS1 = 0
Radiation Effects:	eR1 = 0
Process Effects:	ePr1 = 0

6.1.2.1 Insulation Resistance Errors eIR1

Insulation Resistance error is to be evaluated where actuation functions are expected to operate in an abnormal or harsh environment (Section 2.7). There are no terminal blocks in 100% relative humidity areas, therefore,

elR1 = 0

Since the RTD extension wires are made of the same material as the RTD itself, there is no emf rise or drop across the RTD head terminals (Section 4.2)

eRD1 = 0

6.1.2.3 Temperature Errors eT1

RTDs are designed to exhibit a precise temperature effect that is used to develop the input signal to the loop. Since the RTDs are designed to function at temperatures well above the system design temperature, there is no temperature error other than the reference accuracy error. Therefore,

eT1 = 0

6.1.2.4 Non-Random Input Errors e1in

The RTD is the first module in the loop. Therefore,

e1in = 0

6.1.2.5 Non-Random Error Σe1

 $\Sigma e1 = 0^{\circ}F$ 

#### 6.2 TEMPERATURE TRANSMITTER ERRORS (MODULE 2)

- 6.2.1 Random Error σ2
- 6.2.1.1 Reference Accuracy RA2

Reference Accuracy is  $\pm 0.54^{\circ}$ F (Section 4.1.3). This is a  $3\sigma$  value.

 $RA2 = \pm 0.54^{\circ}F/3 = \pm 0.18^{\circ}F$ 

CALCULATION NO. L-003230	<b>Revision 000</b>	PAGE NO. 9 of 14

Per Reference 5.4.5, this accuracy includes drift and is warranted for 2 years. The calibration interval is 4 years. The 25% late factor is 1 year. (VDP is the vendor drift period, or 2 years in this case.) The formula for applying the surveillance interval to Drift will be applied to the entire RA2 error term.

RA2	=	$\pm$ [IDE] x [(SI + LF)/VDP)] <sup>1/2</sup>
	=	$\pm [0.18^{\circ}F] \times [(4years + 1 year)/2 years)]^{1/2}$
	=	± [0.18°F] x [1.581139]

 $RA2 = \pm 0.285^{\circ}F$ 

#### 6.2.1.2 Calibration Error CAL2

The loop is calibrated using a variable resistance input, measured with a Fluke 189 DMM, and reading the indicated temperature at the PPC.

6.2.1.2.1 Measurement & Test Equipment Error MTE2

#### HP 34401A

Reference Accuracy is the manufacturer's accuracy ( $\pm 0.01\%$  reading + 0.001% of range for the 1k $\Omega$ ) as a 2 $\sigma$  value (Section 5.4.6). The highest reading of interest is 101.5°F. The Minco calibration reports for the RTDs show that the highest resistance value for this temperature is 115.013 $\Omega$ . (Section 4.1.2)

RAMTE2	= ± 0.050°F
	$= \pm 0.0215\Omega \times 1^{\circ}F/0.214\Omega = 0.100^{\circ}F$
	$= \pm 0.0115\Omega + 0.01\Omega = 0.0215\Omega$
$RAMTE_{2\sigma}$	$= \pm 0.01\% \times 115.013\Omega + (0.00001 \times 1000\Omega)$

The manufacturer also specifies a Temperature coefficient for this range (1k $\Omega$ ) for 0°C to 18°C and 28°C to 55°C as 0.0006% of reading + 0.0001% of range per °C. The normal turbine building ambient temperature in the zone where the transmitter is installed varies from 83°F to 102°F(Ref. 5.5.2). For additional conservatism, this range is expanded to 75°F to 102°F (or 23.9°C to 38.9°C). The lower temperature (23.9°C) is within the range where the coefficient is not applicable, so the applicable  $\Delta$ T is: (38.9°C - 28°C) or 10.9°C

TEMTE220	$= \pm (0.0006\% \times 115.013\Omega) + (0.000001 \times 1000\Omega)$
	$= \pm 0.00069\Omega + 0.001\Omega = \pm 0.00169\Omega$
	= ± 0.00169Ω x 1°F/0.214Ω = 0.00789°F
RAMTE2	= ± 0.00395°F

The temperature error is a degradation of the specified accuracy and is not considered an additional random error. Therefore, the total M&TE error for the HP 34401A is:

MTE2 =  $\pm [(0.050^{\circ}F)^{2} + (0.00395^{\circ}F)^{2}]^{1/2}$ 

MTE2 = ± 0.0502°F

### Fluke 45 (medium speed)

Reference Accuracy is the manufacturer's accuracy [ $\pm$  (0.05% reading + 2 LSD + 0.02 $\Omega$ )] as a 2 $\sigma$  value (Section 5.4.6). [The LSD for the Fluke 45 is 0.01 $\Omega$ .] The highest reading of interest is 101.5°F. The Minco calibration reports for the RTDs show that the highest resistance value for this temperature is 115.013 $\Omega$ . (Section 4.1.2)

 $\mathsf{RA}_{2\sigma} = \pm (0.05\% \times 115.013\Omega) + [(2 \times 0.01\Omega) + 0.02\Omega]$ 

## ALCUL ATION DACE

CALCULATION NO. L-003230	Revision 000	PAGE NO. 10 of 14			
= ± 0.0575Ω + 0. = ± 0.0975Ω × 1°	.04Ω = 0.0975Ω °F/0.214Ω = 0.456°F				
MTE2 = ± 0.228°F					
The Fluke 45 (med. speed) M&TE e	error is bounding and will be used to	evaluate total loop uncertainty.			
6.2.1.2.2 Calibration Standard Error STD The calibration standard error is STD2 = 0	<b>12</b> s evaluated as negligible (Section 3.	.2).			
6.2.1.2.3 Loop Calibration Tolerance ST2	2				
The calibration tolerance for this ST2 = $\pm 0.54^{\circ}$ F	s indicating loop is $\pm$ 0.54°F (Section / 3	n 4.7). [3σ]			
ST2 = ±0.18°F					
6.2.1.2.4 Calibration Error CAL2					
The total calibration error for the	e M&TE is:				
$CAL2 = \pm [(MTE2) = \pm [(0.228)$	$(2)^{2} + (STD2)^{2} + (ST2)^{2}]^{1/2}$ $(F)^{2} + (0)^{2} + (018)^{2}]^{1/2}$				
$CAL2 = \pm 0.29^{\circ}F$					
6.2.1.3 Ambient Temperature Error oT2	2				
The vendor states the "tempera measuring range / 10°C (Ref. 4 temperature error. Measuring ra	ture drift" error for the temperature t .1.3) [3 $\sigma$ ]. This is applied in this calc ange: 30 to 120°F = 90°F.	ransmitter as 0.1% of ulation as an ambient			
The normal turbine building aml varies from 83°F to 102°F(Ref. 75°F to 102°F (27°F difference)	bient temperature in the zone where 5.5.2). For additional conservatism,	the transmitter is installed this range is expanded to			
$\sigma T2_{3\sigma} = \pm (0.1\% *$ = $\pm [(0.001)$ = $\pm 0.1519$	∗ Span) ∗ 90°F)/10°C x (27°F x 5°F/8°C) °F / 3				

σT2 = ± 0.051°F

6.2.1.4 Random Input Error o2in

> σ2in = σ1 ± 0.150°F =

#### 6.2.1.5 Power Supply Effects **o2PS**

The transmitter specifications are valid for voltages between 20 and 30 vDC. The 24-volt power supply variability is less than ± 2% all errors combined (4.3). This is equal to 23.5vDC to 24.5vDC. Therefore,

σ2PS = ±0°F

CALCU	LATION NO.	L-003	230	Revision 000	PAGE NO. 11 of 14
	σ2 σ2 <b>σ2</b>	-	$\pm [(RA2)^{2} + (CAL2)^{2} + (\sigma)^{2} + (0.285^{\circ}F)^{2} + (0.290^{\circ}F)^{2} + (0.290^{\circ}F)^{2} + 0.436^{\circ}F$	T2) <sup>2</sup> + (σ2in) <sup>2</sup> + (σ2P <sup>2</sup> + (0.051°F) <sup>2</sup> + (0.15	S) <sup>2</sup> ] <sup>1/2</sup> 0°F) <sup>2</sup> + (0°F) <sup>2</sup> ] <sup>1/2</sup>
6.2.2	Non-Randon	n Error Σ	e2		
6.2.2.1	Humidity Erro	or <b>e2H</b>			
	No humidity at the instrun during norma	effect er nent loca al conditi	rors are provided in the ma ation are within the operatir ons. (Reference 5.1.2, App	nufacturer's specifica ng limits of the module pendix I)	tions, and the humidity conditions a. Humidity errors are negligible
	e2H =	0			

6.2.2.2 Radiation Error e2R

No radiation errors are provided in the manufacturer's specifications. Per Section 2.8, it is reasonable to consider the normal radiation effect as negligible. Therefore,

e2R = 0

6.2.2.3 Seismic Error e2S

No seismic effect errors are provided in the manufacturer's specifications. A seismic event defines a particular type of accident condition. Therefore, there is no seismic error for normal operating conditions

e2S = 0

6.2.2.4 Static Pressure Offset Error e2SP

The transmitter is an electrical device and therefore not affected by static pressure.

e2SP = 0

6.2.2.5 Ambient Pressure Error e2P

The transmitter is an electrical device and therefore not affected by ambient pressure.

e2P = 0

6.2.2.6 Process Error e2Pr

The transmitter receives an analog input from an RTD. Any errors associated with the conversion of temperature to resistance have been accounted for as RTD errors. Therefore,

e2Pr = 0

6.2.2.7 Non-Random Input Error e2in

 $e2in = \Sigma e1 = 0$ 

6.2.2.8 Total Non-Random Error Σe2

 $\Sigma e^2 = e^{2H} + e^{2R} + e^{2S} + e^{2SP} + e^{2P} + e^{2Pr} + e^{2in}$ = 0 + 0 + 0 + 0 + 0 + 0 + 0  $\Sigma e^2 = 0$ 

Σ**e2** = 0

CALCU	ULATION NO. L-003230 Revision 000 PA	AGE NO. 12 of 14
6.3.1	Random Error <b>3</b>	
6.3.1.1	Reference Accuracy RA3	
	Reference Accuracy is $\pm$ 0.025% calibrated range (Ref. 5.4.9). The calibrated 120°F (120°F – 30°F = 90°F).	d range is 30°F to
	$RA3_{2\sigma} = \pm 0.00025 \times 90^{\circ}F = 0.0225^{\circ}F$	
	RA3 = ±0.0113°F	
6.3.1.2	Calibration Error CAL3	
	The I/O module is not separately calibrated; indication is verified during loop	calibration.
	$CAL3 = \pm 0^{\circ}F$	
6.3.1.3	Drift Error <b>D3</b>	
	The vendor does not specify a drift error specification for the I/O module. $D3 = \pm 0^{\circ}F$	
6.3.1.4	Random Input Error <b>o3in</b>	
	$\sigma 3 in = \sigma 2 = \pm 0.437^{\circ} F$	
6.3.1.5	Total Random Error σ3	
	$\sigma 3 = \pm [(RA3)^2 + (CAL3)^2 + (\sigma D3)^2 + (\sigma 3in)^2 + (\sigma 3r)^2]^{1/2}$ $\sigma 3 = \pm [(0.0113^{\circ}F)^2 + (0.0^{\circ}F)^2 + (0^{\circ}F)^2 + (0.436^{\circ}F)^2]^{1/2}$	
	$\sigma 3 = \pm 0.436^{\circ} F$	
6.3.2	Non-Random Error <b>Se3</b>	
6.3.2.1	Humidity Error <b>e3H</b>	
	No humidity effect errors are provided by the manufacturer'; specified RH for 80% RH. The I/O module is located in EQ Zone C1A, (Section 4.6), where ex to 50%. Humidity errors are negligible. (Reference 5.1.2, Appendix I)	PPC equipment is 20 to pected RH levels are 20
	e3H = 0	
6.3.2.2	Radiation Error e3R	
	No radiation errors are provided in the manufacturer's specifications. Per Sec reasonable to consider the normal radiation effect as negligible. Therefore,	tion 2.8, it is
	e3R = 0	
6.3.2.3	Seismic Error e2S	
	No seismic effect errors are provided in the manufacturer's specifications. A s particular accident condition. Therefore, there is no seismic error for normal op	eismic event defines a perating conditions

e3S = 0

6.3.2.4 Static Pressure Offset Error **e3SP** 

CALCULATION NO. L-003230	<b>Revision 000</b>	PAGE NO. 13 of 14

The I/O module is an electrical device and therefore not affected by static pressure.

e3SP = 0

6.3.2.5 Ambient Pressure Error e3P

The I/O module is an electrical device and therefore not affected by ambient pressure.

e3P = 0

6.3.2.6 Process Error e3Pr

The I/O module receives an analog current input from the transmitter. Any errors associated with the conversions of temperature to resistance, and resistance to current have been accounted for as errors associated with modules 1 and 2. Therefore,

(Section 4.4)

e3Pr = 0

6.3.2.7 Input Signal Resistor Error e3SR

e3SR = ± (0.02% \* Span) = ± 0.0002 \* 90°F = ± 0.018 °F

 $e3SR = \pm 0.018^{\circ}F$ 

6.3.2.8 Non-Random Input Error e3in

 $e3in = \Sigma e2 = 0$ 

- 6.3.2.9 Total Non-Random Error Σe3
  - $\Sigma e3 = e3H + e3R + e3S + e3SP + e3P + e3Pr + e3SR + e3in$ = 0 + 0 + 0 + 0 + 0 + 0 + 0.018 + 0

 $\Sigma e3 = 0.018$ 

### 6.4 SUMMARY AND CONCLUSION (TOTAL ERROR)

6.4.1 As discussed in Methodology Section 2.2, Level 2 methodology is applied for determining Total Error for this indication loop:

TE =  $\sigma 3 + \Sigma e 3$ =  $\pm (0.436^{\circ}F) + 0.018^{\circ}F$ =  $\pm 0.454^{\circ}F$ TE =  $\pm 0.454^{\circ}F$ 

#### In conclusion, the total uncertainty for the CW Inlet Temperature Indication loop is $\pm 0.454^{\circ}$ F

6.4.2 To obtain a more accurate value of the UHS temperature using these instruments, the average of the available values can be taken. This assumes that the four readings are sensing the same input temperature and that there is little effect between the input and the measurement point.

$$T_{CWAverage} = \frac{T_{1TE-CW010} + T_{1TE-CW011} + T_{2TE-CW010} + T_{2TE-CW011}}{4}$$

## CALCULATION NO. L-003230 Revision 000 PAGE NO. 14 of 14

The accuracy of this process is considered the same as the accuracy of summing networks addressed in References 5.1.1 and 5.1.2, or by the multiple test criterion of Reference 5.1.4 Section 3.2.

In all of these cases the final random uncertainty ( $\sigma$ ) is the square root sum of the squares of the individual channel random uncertainties considering the multiplier for each of the uncertainties is one divided by the number of channels that are being averaged. The non-random uncertainty (e) will remain the same as for a single loop (Ref. 5.1.4, Section 3.2).

$$\sigma_{Average} = \sqrt{\left(\frac{\sigma_1}{n}\right)^2 + \left(\frac{\sigma_2}{n}\right)^2 + \left(\frac{\sigma_3}{n}\right)^2 + \cdots + \left(\frac{\sigma_n}{n}\right)^2}$$

If all of the instrument loops are identical then this equation will reduce to:

$$\sigma_{Average} = \frac{\sigma_i}{\sqrt{n}} + e$$

Thus for the CW temperatures, the accuracy of the average of the readings for two loops will be:

$$\sigma_{Average} = \frac{0.436}{\sqrt{2}} + e = 0.308 + 0.018 = 0.326 \,^{\circ}\text{F}$$

The accuracy of the average of the readings for three loops will be:

$$\sigma_{Average} = \frac{0.436}{\sqrt{3}} + e = 0.252 + 0.018 = 0.270 \text{ °F}$$

The accuracy of the average of the readings for four loops will be:

$$\sigma_{Average} = \frac{0.436}{\sqrt{4}} + e = 0.218 + 0.018 = 0.236 \,^{\circ}\text{F}$$



A critical component of your success"

L-003230 Rev. 0 Attachment A Page A1 (final) 7300 Commerce Lane Minneapolis, MN 55432 U.S.A. Customer Service Telephone: 763-571-3123 Sales Inquiries Fax: 763-571-0927 Purchase Order Fax: 763-571-0942 E-Mail: custserv@minco.com

#### QUOTATION To: Vikram Shah 160056-2 Quote No: **Exelon** Corporation Page: 1 LaSalle County Nuclear Station Date: January 26, 2006 2601 N 21st Marsailles Road **RTD** Assemblies RFQ: Marseilles IL 61341-9757 Phone: 815-415-3828 CC: Thermo/Cense, Inc. Fax: 942 Turret Court Mundelein, IL 60060 Fax Order to 763-571-0942 or Phone: 847-949-8070,8071 E-Mail Order to custserv@minco.com Fax: 847-949-8074

## Please Reference Above Quote Number When Placing Your Order.

			Unit Price
Item	Description	Quantity	U.S. \$
1	Minco Part # ASSEMBLY	1 - 9	162.60
	Assembly Consisting Of:		
	CGASSY		
	CH359P2T6		
	FG113-1		
	FG750F8M12		
	XS853PD157X4		
	X = Class A sensor.		
	Single Element RTD assembly		
2	Minco Part # XRT07	1 - 9	425.00
	Test charge for a chart of temperature readings at .1F intervals		
	from -272F to 932F	A	
		where the state in the later of the second	totale a contra contra car car contra comme

## Notes:

- 1. These assemblies will replace the existing head that is on the thermowell. This is due to not knowing how long the replacement probe would need to be. The drawing does not provide all of this information to determine the proper length. Lead time for these parts is also relatively short as compared to a special probe.
- 2. 1. Probe length is 15.6". This is the necessary length of the probe to fit in the thermowell and fit into the connection head.

2. The probe diameter is .25", but will fit in the thermowell without any reduction in performance.

3. Drift specifications on the S852 sensor is listed as +/- .2 F per year, repeatability is also +/- .2 F. This specification assumes cycling throughout the full temperature range of the sensor, from - 50C to 260C. A smaller temperature cycle will change the amount of drift.



Print Date: 07/28/2006 10:12

## VanWyk, Thomas J.

From: Sent: To: Subject: Keith Jensen [Keith.Jensen@minco.com] Wednesday, July 26, 2006 9:22 AM Shah, Vikram R. Fwd: Exelon Corporation L-003230 Rev. 0 Attachment C Page C1 (final)



100995.pdf

>>> Keith Jensen 7/25/2006 3:50 PM >>> Vikram Shah 815-415-3828 Exelon Corporation Marsailles IL vikram.shah@exelon.com

XS853PD157X4 RFQ 160056-2

The S100995 probe meets the EN60751 Class A +/- 0.06% @ OC sensor accuracy requirements

Minco estimates the drift per year over the range of 30F to 120F would be expected to be around 0.1F or less (PHP)  $\,$ 

The drawing is attached

Keith Jensen 763-586-2908 Applications Engineer MINCO PRODUCTS INC. Minneapolis MN keith.jensen@minco.com



02/04 Sachnr. 701724/01

## **Technical data**

Operating voltage [V]	30 DC <sup>1)</sup>
Current rating [mA]	250
Short-circuit prot., reverse polarity prot. / overlo	ad prot.
, , , , , , , , , , , , , , , , , , ,	vatchdog
Voltage drop [V].	< 2
Current consumption [mA]	- 552)
Constant current sensor [mA] 0.2 (Pt 1000	$\alpha   \alpha m \alpha n t \rangle$
Constant current sensor [mA]	element
Power on delay time [c]	element)
Power-on delay line [5]	1.5
Response time switching output [ms]	130
Analogue output (measuring range scaleable)	) 10 V
Max. load current output [ $\Omega$ ](U <sub>B</sub> - 10) x 50; 700 at U	B = 24 V
Min. load with voltage output $[\Omega]$	2000
Response time analogue output [ms]	384
Accuracy	
Switching output $[^{\circ}C/^{\circ}F]$ + 0.3	(+0.54)
Analog output $[^{\circ}C/^{\circ}F]$ + 0.3	(+0.54)
Display $[^{\circ}C/^{\circ}F]$ + (0.3/+0.54+	16 Digit)
Resolution	72 Digit/
Switching output [°C/°E]	01/01
Analogue output [OC/9E]	$J_{1} / 0_{1}$
	$J_{1} / J_{1} / J_{1$
Display [ C F]	J. 1 / U. 1
remperature drift [% or value of measuring range/10 K]	$1.\pm 0.1$
Housing material stainless steel (304S15); EPDM/X (Santi	oprene);
PC (Macrolon); Pocan; FPM	1 (Viton)
Operating temperature [°C]	5 +70
Storage temperature [°C]40	) +85
Protection	IP 67. III
Insulation resistance [M $\Omega$ ]	$0 \vee DC$
Shock resistance [g]	11ms)
Vibration resistance [g] 20 (DIN / IEC 68-2-6 10 - 2	000 Hz)
FMC	
EN 61000-4-2 ESD	1/9//
EN 61000-4-3 HE radiated:	
EN 61000 4 A Buret:	
EN 61000-4-4 DUSL	Z KV
	10 V
1) to EN50178 SELV PELV	
referring to ULL' see page 21 (Electrical connection)	

ŧ

referring to UL: see page 21 (Electrical connection).
 <sup>2)</sup> 41 mA when the display is switched off; the values apply to the operating voltage = 24 V and unloaded outputs.

L-003230 Rev. 0 Attachment E Page E1 (final)



# ifm efector inc.

782 Springdale Drive, Exton, PA 19341 • 800-441-8246 • Fax: 800-329-0436 • www.ifmefector.com

July 26, 2006

Mr. Vikram Shah Exelon Corporation 2601 N 21st Rd. Marseilles, Illinois 61341

Dear Vikram:

This letter is in response to your concern about the specifications of the **ifm efector** TR2432 temperature sensor. The following points should clarify the questions that you had:

- After production, 100% of the sensors are verified and tested to the specifications listed on our datasheet.
- The analog accuracy specification of (+/- 0.54°F) already includes the analog resolution value of (0.1°F), and is inclusive of any electronic component drift.
- The temperature drift specification is the electronic drift that occurs for every 10°C change in temperature that occurs in the application. This drift is in addition to the accuracy specification.
- There are no other environmental influences that will affect the accuracy specification.
- These sensors have a warranty period of 2 years.

Please contact me if you have any further questions, or if you require any additional information.

Best regards,

Amer Shah

Ameera Shah Product Support Engineer Fluid Sensors Team

### SPECIFICATIONS - OHMS

## L-003230 Rev. 0 Attachment F Page F1 (final)

## Attachment F: Fluke 45 Accuracy Specifications

OHMS

	Resolution				Typical Full	Max Current
Hange	Slow	Medium	Fast	Accuracy	Full Scale Voltage	Through the Unknown
300Ω	_	10 mΩ	100 MΩ	0.05% + 2 + 0.02Ω	0.25	1 mA
3 kΩ		100 MΩ	1Ω	0.05% + 2	0.24	120 <i>µ</i> A
30 kΩ		1Ω	10Ω	0.05% + 2	0.29	14 µA
300 kΩ		10Ω	100Ω	0.05% + 2	0.29	1.5 µA
3 MΩ		100Ω	1 kΩ	0.06% + 2	0.3	150 μA
30 MΩ		1 kΩ	10 kΩ	0.25% + 3	2.25	320 µA
300 MΩ*	-	100 kΩ	1 MΩ	2%	2.9	320 µA
100Ω	1 mΩ			0.05% + 8 + 0.02Ω	0.09	1 mA
1000Ω	10 mΩ	-		0.05% + 8 + 0.02Ω	0.10	120 <i>µ</i> A
10 kΩ	100 mΩ			0.05% + 8	0.11	14 μA
100 kΩ	1Ω		-	0.05% + 8	0.11	1.5 μA
1000 kΩ	10Ω			0.06% + 8	0.12	150 μA
10 MΩ	100Ω		-	0.25% + 6	1.5	150 µA
100 MΩ*	100 kΩ		-	2 % + 2	2.75	320 µA
*Because of the method used to measure resistance, the 100 M $\Omega$ (slow) and 300 M $\Omega$ (medium and fast)						

ranges cannot measure below 3.2 M $\Omega$  and 20 M $\Omega$ , respectively. "UL" (underload) is shown on the display for resistances below these nominal points, and the computer interface outputs "+1E-9".

#### **Open Circuit Voltage**

3.2 volts maximum on the 100 $\Omega$ , 300 $\Omega$ , 30 M $\Omega$ , 100 M $\Omega$ , and 300 M $\Omega$  ranges, 1.5 volts maximum on all other ranges.

#### **Input Protection**

500V dc or rms ac on all ranges



L-003230 Rev. 0 Attachment G Page G1 (final)



## SDN<sup>™</sup> Specifications (Single Phase)

<b>m 1</b> / 1	Catalog Number						
Description	SDN 2.5-24-100P	SDN 4-24-100LP	SDN 5-24-100P	SDN 10-24-100P	SDN 20-24-100P		
······································	·	I	nput				
Nominal Voltage			115/230 VAC auto select				
-AC Range			85-132/176-264 VAC				
-DC Range <sup>4</sup>	90-375 VDC		210-375 VDC		N/A		
-Frequency			47 - 63 Hz				
Nominal Current <sup>1</sup>	1.3 A. / 0.7 A	2.1 A / 1.0 A	2.2 A/ 1.0 A	5 A / 2 A typ.	9 A/ 3.9 A		
-Inrush current max.	typ. < 25 A	tyr	o. < 20 A	lyp. •	: 40 A		
Efficiency (Losses <sup>2</sup> )	> 87.5% typ (8.6 W)	> 88% typ (13.1 W)	> 88% typ (16.4 W)	> 88% typ (32.7 W)	> 90% typ (48 W)		
Power Factor Correction			Units Fulfill EN61000-3-2				
**************************************	An	0	utput				
Nominal Voltage	24 VDC (22.5 - 28.5 VDC adj.)	24 VDC (22.5 - 25.5 VDC adj.)		24 VDC (22.5 - 28.5 VDC adj.)			
-Tolerance		< ±2% overall (comb	ination Line, load, time and tempera	ature related changes)			
-Rippie <sup>3</sup>			< 50 mVpp				
Nominal Current	2.5 A (60 W)	3.8 A (92 W)	5 A (120 W)	10 A (240 W)	20 A (480 W)		
-Peak Current <sup>4</sup>	1.6x Nominal Current < 2 sec.	4.2 A max at 23.8V	6 A 2x Nominal Current < 2 sec.	12 A 2x Nominal Current < 2 sec.	25 A 2x Nominal Current < 2 sec.		
-Current Limit	Fold I	Forward (Current rises, voltage	drops to maintain constant power di	uring overload up to max peak cur	rent)		
Holdup Time <sup>s</sup>	> 50 ms	> 100 ms	> 100	ms	> 20 ms		
Parallel Operation	Single or Paralle	el use is selectable via Front Pa	nel Switch (SDN4 should not be use	d in parallel as Class 2 rating wou	ld be violated.)		
· · · · ·		Ge	neral				
EMC: -Emissions		EN61000-6-3, -4; Class B I	EN55011, EN55022 Radiated and C	onducted including Annex A.			
-Immunity	EN61000-6-1, -2; EN61000-4-2 Level 4, EN61000-4-3 Level 3; EN61000-4-6 Level 3; EN61000-4-4 Level 4 input and Level 3 output; EN61000-4-5 Isolation Class 4, EN61000-4-11; Transient resistance according to VDE 0160/W2 over entire load range.						
Approval <b>s</b>	EN60950; EN50178; EN60204; UL508 Listed, cULus; UL60950, cRUus, CE (LVD 73/23 & 93/68/EEC). EN61000-3-2, IEC60079-15 (Class 1, Zone 2, Hazardous Location, Groups A, B, C, D w/ T3A temp class up to 60°C Ambient.) SEMI F47 Sag Immunity. SDN2.5 & SDN4 - UL60950 testing to include approval as Class 2 power supply.						
Temperature	Storage: -25°C+85°C Operation10°-60°C full power with operation to 70°C possible with a linear derating to half power from 60°C to 70°C (Convection cooling, no forced air required). Operation up to 50% load permissable with sideways or front side up mounting orientation. The relative humidity is < 90% RH, noncondensing; IEC 68-2-2, 68-2-3. For operation below -10°C, contact Technical Services.						
MTBF:	> 820,000 hours > 640,000 hours > 600,000 hours > 510,000 ho			> 510,000 hours			
- Standard	Beilcore Issue 6 Method 1 Case 3 @ 40C MIL217F @ 30C				MIL217F @ 30C		
Warranty	5 years						
General Protection/Safety	Protected against continuous short-circuit, overload, open-circuit. Protection class 1 (IEC536), degree of protection IP20 (IEC 529) Safe low voltage: SELV (acc.EN60950)			P20 (IEC 529)			
Status Indicators	Green LED and DC OK signal (N.O. Solid State Contact rated 200 mA / 60 VDC)						
		Insta	llation				
Fusing -Input	Internally fused. External 10 A slow acting fusing for the input is recommended to protect input wiring.						
-Output	Outputs are capable of providing high currents for short periods of time for inductive load startup or switching. Fusing may be required for wire/loads if 2x Nominal O/P current rating cannot be tolerated. Continuous current overload allows for reliable fuse tripping.						
Mounting	Simple snap-on system for DIN Rail TS35/7.5 or TS35/15 or chassis-mounted (optional screw mounting set SDN-PMBRK2 required).						
Connections	Input: IP20-rated screw terminals, connector size range: 16-10 AWG (1.5-6 mm2) for solid conductors. 16-12 AWG (0.5-4 mm2) for flexible conductors. Output: Two connectors per output, connector size range: 16-10 AWG (1.5 - 6 mm2) for solid conductors.						
Case		Fully enclosed metal	housing with fine ventilation grid to	keep out small parts.			
-Free Space	25 mm above and below, 10 mm	25 mm left and right, in front	25 mm above and below, 25 mm left and right, 15 mm in front	70 mm above and below 15 mm	, 25 mm left and right, in front		
H x W x D (inches/mm)	4.88 in. x 1.97 in. x 4.55 in. (124 mm x 50 mm x 116 mm)	4.88 in. x 2. (124 mm x 6	56 in. x 4.55 in. 5 mm x 116 mm)	4.88 in. x 3.26 in. x 4.55 in. (124 mm x 83 mm x 116 mm)	4.88 in. x 6.88 in. x 4.55 in. (124 mm x 175 mm x 116 mm)		
Weight (ibs/g)	1 ib (460g)	1.5 lb	s (620g)	2.2 lbs (1100g)	3 lbs (1520g)		
1 Input current rations are conse	ervatively specified with low input, wor	st case efficiency and power facto	All peak current is	s calculated at 24 Volt levels.			

Input current ratings are conservatively specified with low input, worst case efficiency and power factor.

Losses are heat dissipation in watts at full load, nominal input line. Ripple/noise is stated as typical values when measured with a 20 MHz, bandwidth scope and 50 Ohm resistor. Full load, 100 VAC Input @ Tamb = +25"C

Not UL listed for DC input. 6

Visit our website at www.solaheviduty.com or

contact Technical Services at (800) 377-4384 with any questions.

3

## 8436/32 8-Channel Isolated Low-Level Analog Input Card

The RTP8436/32 8-Channel Isolated Analog Input Card provides high accuracy low-level (±160 mV) analog measurements. Sampling transformers provide channel-to-channel isolation. Very high noise immunity is characteristic of the transformer multiplexer, achieving 160 dB of common mode rejection. Immunity to noise is further enhanced with a two-pole low pass filter, set to provide 70 dB of normal mode rejection at 60 Hz.

Analog to digital conversion is performed by a 16-bit switched capacitor successive approximation A/D converter. A precision voltage source provides a self-test function for the card's amplifiers and A/D converter. No field adjustments are necessary after the initial factory setup.

#### Specifications

ų.

Input Signal Range:	± 160 mV
Multiplexer Type:	8-channel solid state multiplexer with individual transformers for complete channel-to-channel isolation
Sample Rate:	50 samples per second per channel
Accuracy:	0.025% of Full Scale
Temperature Ranges:	<ul> <li>25° to +85°C (-13° to +185°F), storage</li> <li>0° to +55°C (+32° to +131°F), standard operating</li> <li>-20° to +60°C (-4° to +140°F), extended operating</li> </ul>
	Note: Input measurements may not meet the accuracy specification at the upper or lower ends of the extended operating range.
Isolation:	600 VAC RMS or 400 VDC 1500 VAC @ 60 Hz for 60 seconds withstand
Common Mode Voltage:	600 VAC RMS or 400 VDC continuous
Common Mode Rejection:	–160 dB at 60 Hz (100 $\Omega$ unbalanced)
Common Mode Crosstalk:	–150 dB at 60 Hz
Normal Mode Rejection:	2-pole low-pass filter, -70 dB at 60 Hz
Input Impedance:	5 $M\Omega$ in parallel with 10 pF at 50 samples/second per channel
Input Bias Current:	8 nA maximum at 50 samples/second per channel
Input Source Impedance:	100 $\Omega$ maximum to meet accuracy specification

Melaton <sup>Ser</sup>ian in c 115200

L-003230 Rev. 0 Attachment I Page I1

\* \* \* Report of Calibration \* \* \* for Platinum Resistance Thermometer Model S100995PD Serial No. P/N366

RTE-CWOIL

<u>(</u>,

÷.,

(

W ......

				L-003230 Rev. 0 Attachment I Page I2 (final)	115	115200	
T(°F)	R(ohms)	T(°F)	R(ohms)	T(°F) R(ohms)	T(°F)	R(ohms)	
100.0 100.1 100.2 100.3 100.4 100.5 100.6 100.7 100.8 100.9	$114.692 \\ 114.713 \\ 114.735 \\ 114.756 \\ 114.777 \\ 114.799 \\ 114.820 \\ 114.842 \\ 114.863 \\ 114.884$	105.0 105.1 105.2 105.3 105.4 105.5 105.6 105.7 105.8 105.9	115.762 115.784 115.805 115.827 115.848 115.869 115.891 115.912 115.934 115.955	110.0116.832110.1116.854110.2116.875110.3116.896110.4116.918110.5116.939110.6116.961110.7116.982110.8117.003110.9117.025	115.0 115.1 115.2 115.3 115.4 115.5 115.6 115.7 115.8 115.9	117.901 117.922 117.944 117.965 117.986 118.008 118.029 118.051 118.072 118.093	
101.0 101.1 101.2 101.3 101.4 101.5 101.6 101.7 101.8 101.9	114.906 114.927 114.949 114.970 114.992 115.013 115.034 115.056 115.077 115.099	106.0 106.1 106.2 106.3 106.4 106.5 106.6 106.7 106.8 106.9	115.976 115.998 116.019 116.041 116.062 116.105 116.126 116.148 116.169	111.0117.046111.1117.067111.2117.089111.3117.110111.4117.132111.5117.153111.6117.174111.7117.196111.8117.217111.9117.238	116.0 116.1 116.2 116.3 116.4 116.5 116.6 116.7 116.8 116.9	118.115 118.136 118.157 118.200 118.221 118.243 118.264 118.286 118.307	
102.0 $102.1$ $102.2$ $102.3$ $102.4$ $102.5$ $102.6$ $102.7$ $102.8$ $102.9$	115.120 115.142 115.163 115.184 115.206 115.227 115.249 115.270 115.291 115.313	107.0 107.1 107.2 107.3 107.4 107.5 107.6 107.7 107.8 107.9	116.190 116.212 116.233 116.255 116.276 116.297 116.319 116.340 116.362 116.383	112.0117.260112.1117.281112.2117.303112.3117.324112.4117.345112.5117.367112.6117.388112.7117.410112.8117.431112.9117.452	117.0 117.1 117.2 117.3 117.4 117.5 117.6 117.7 117.8 117.9	118.328 118.350 118.371 118.392 118.414 118.435 118.456 118.478 118.499 118.520	
103.0 103.1 103.2 103.3 103.4 103.5 103.6 103.7 103.8 103.9	115.334 115.356 115.377 115.399 115.420 115.441 115.463 115.484 115.506 115.527	108.0 108.1 108.2 108.3 108.4 108.5 108.6 108.7 108.8 108.9	116.404 116.426 116.447 116.469 116.511 116.533 116.554 116.576 116.597	113.0117.474113.1117.495113.2117.516113.3117.538113.4117.559113.5117.580113.6117.602113.7117.623113.8117.645113.9117.666	118.0 118.1 118.2 118.3 118.4 118.5 118.6 118.7 118.8 118.9	118.542 118.563 118.585 118.606 118.627 118.649 118.670 118.691 118.713 118.734	
104.0 104.1 104.2 104.3 104.4 104.5 104.6 104.7 104.8 104.9	115.548 115.570 115.591 115.613 115.634 115.655 115.677 115.698 115.720 115.741	109.0 109.1 109.2 109.3 109.4 109.5 109.6 109.7 109.8 109.9	116.618 116.640 116.661 116.683 116.704 116.725 116.747 116.768 116.789 116.811	114.0117.687114.1117.709114.2117.730114.3117.751114.4117.773114.5117.794114.6117.816114.7117.837114.8117.858114.9117.880	119.0 119.1 119.2 119.3 119.4 119.5 119.6 119.7 119.8 119.9	118.755 118.777 118.798 118.819 118.841 118.862 118.883 118.905 118.926 118.947	
105.0	115.762	110.0	116.832	115.0 117.901	120.0	118.969	

ſ

¥

## Attachment J: HP 34401A Accuracy Specifications

Chapter 8 Specifications **DC Characteristics** 

## **DC** Characteristics

ere i

Function	Range [3]	Test Current or Burden Voltage	24 Hour [ 2 ] 23°C ± 1°C	90 Day 23°C ± 5°C	1 Year 23°C ± 5°C	Temperature Coefficient /°C 0°C – 18°C 28°C – 55°C	
DC Voltage	100.0000 mV 1.000000 V 10.00000 V 100.0000 V 1000.000 V		0.0030 + 0.0030 0.0020 + 0.0006 0.0015 + 0.0004 0.0020 + 0.0006 0.0020 + 0.0006	0.0040 + 0.0035 0.0030 + 0.0007 0.0020 + 0.0005 0.0035 + 0.0006 0.0035 + 0.0010	0.0050 + 0.0035 0.0040 + 0.0007 0.0035 + 0.0005 0.0045 + 0.0006 0.0045 + 0.0010	0.0005 + 0.0005 0.0005 + 0.0001 0.0005 + 0.0001 0.0005 + 0.0001 0.0005 + 0.0001	
Resistance [4]	100.0000 Ω 1.000000 kΩ 10.00000 kΩ 100.0000 kΩ 1.000000 MΩ 10.00000 MΩ 100.0000 MΩ	1 mA 1 mA 100 μA 10 μA 5 μA 500 nA 500 nA // 10 MΩ	0.0030 + 0.0030 0.0020 + 0.0005 0.0020 + 0.0005 0.0020 + 0.0005 0.002 + 0.001 0.015 + 0.001 0.300 + 0.010	$\begin{array}{c} 0.008 \pm 0.004 \\ 0.008 \pm 0.001 \\ 0.020 \pm 0.001 \\ 0.800 \pm 0.010 \end{array}$	$\begin{array}{c} 0.010 + 0.004 \\ 0.010 + 0.001 \\ 0.010 + 0.001 \\ 0.010 + 0.001 \\ 0.010 + 0.001 \\ 0.040 + 0.001 \\ 0.800 + 0.010 \end{array}$	0.0006 + 0.0005 0.0006 + 0.0001 0.0006 + 0.0001 0.0006 + 0.0001 0.0010 + 0.0002 0.0030 + 0.0004 0.1500 + 0.0002	
DC Current	10.00000 mA 100.0000 mA 1.000000 A 3.000000 A	< 0.1 V < 0.6 V < 1 V < 2 V	0.005 + 0.010 0.01 + 0.004 0.05 + 0.006 0.10 + 0.020	0.030 + 0.020 0.030 + 0.005 0.080 + 0.010 0.120 + 0.020	0.050 + 0.020 0.050 + 0.005 0.100 + 0.010 0.120 + 0.020	0.002 + 0.0020 0.002 + 0.0005 0.005 + 0.0010 0.005 + 0.0020	
Continuity	1000.0 Ω	1 mA	0.002 + 0.010	0.008 + 0.020	0.010 + 0.020	0.001 + 0.002	
Diode Test	1.0000 V	1 mA	0.002 + 0.010	0.008 + 0.020	0.010 + 0.020	0.001 + 0.002	
DC:DC Ratio	100 mV to 1000 V		(Input Accuracy) + (Reference Accuracy) Input Accuracy = accuracy specification for the HI-LO input signal. Reference Accuracy = accuracy specification for the HI-LO reference input signal.				

## Accuracy Specifications $\pm$ (% of reading + % of range) [1]

Transfer Accuracy (typical)

(24 hour % of range error) 2

#### Conditions:

• Within 10 minutes and  $\pm 0.5^{\circ}C$ .

- Within ±10% of initial value.
- Following a 2-hour warm-up.
- Fixed range between 10% and 100% of full scale.
- Using 6<sup>1</sup>/<sub>2</sub> digit slow resolution (100 PLC).
  Measurements are made using accepted metrology practices.