



South Jersey Healthcare®
Changing Medicine. Changing Lives.

Corporate Office

2950 College Drive
Suite 1E
Vineland, New Jersey 08360
(856) 641-8000

July 28, 2006

Region I USNRC
Division of Materials Licensing
475 Allendale Road
King of Prussia PA 19605

Re: License # NRC 29-19311-01
South Jersey Healthcare - Regional Medical Center

Q-8

MS-12

2006 AUG - 7 PM 12: 17

RECEIVED
REGION I


Dear Sir or Madam:

Please amend our license to allow the use of Gadolinium - 153 for the purpose of attenuation correction source for a Vantage-equipped gamma camera.

| | | |
|---------------------------------|---|--------------------------------|
| Manufacturer | Isotope Products Laboratories (IPL) | North American Scientific(NAS) |
| Model Number | NES-8412 | MED3601 |
| Registration Number | CA0406S204S | CA0510S121S |
| Single Source Strength | 250 mCi, not to exceed 300 mCi | 250 mCi, not to exceed 300 mCi |
| Maximum Possession Limit | 750 mCi | 750 mCi |
| Authorized user | Robert W. Ciphers, M.D. (Current authorized user 35.200 on NRC License No. 29-19311-01) | |

Please contact Jonathan Law (609) 652-3409 if there are any questions regarding this amendment request.

Sincerely,


Wayne Schiffner
Executive Vice President

WCS:pr

139080
NMSS/RGNI MATERIALS-C02