

600 Mountain Avenue Murray Hill, NJ 07974-0636

United States Nuclear Regulatory Commission Nuclear Materials Safety Branch Division of Nuclear Materials Safety Region I 475 Allendale Road King of Prussia, PA 19406 August 10, 2006

9-5

Re: Termination of License SNM-203 07000 222

To Whom It May Concern:

Lucent Technologies' R&D management has determined that there will no longer be a need for the licensed radioactive material that was formerly used under the above license. All materials held under SNM-203 have been disposed of. Enclosed is the record issued by Los Alamos National Laboratory (LANL) wherein it states that it has received the Pu-Be source from the Lucent Technologies' inventory (ATRC #2006:33). This source was the only one held under this license – contrary to the LANL record. The other source received by LANL (ASRO #2006:36), an Am-Be, was held under the Lucent Byproduct license. Also enclosed are copies of the last two leak tests of this source showing that it was intact. Lucent Technologies will retain its Byproduct and Irradiator licenses and continue to use the sources covered by them, 29-00170-03 and 29-00170-08. As soon as resources are available, Lucent will contract for the necessary measurements to terminate the remaining license, SMB-1260.

If you have questions regarding this request, please contact me at (908) 582-5907.

Sincerely,

Richard Quick,

Senior Manager - Radiation Safety Officer

Attachments: LANL Receipts

Source Leak Tests 8/17/05, 12/14/05

Cc:

With Attachments-

NRC Correspondence File

Without Attachments-

**RPC Committee** 

139958 NMSS/RGNI MATERIALS-002



N2 Advanced Nuclear Technology Off-Site Source Recovery (OSR) Project P.O. Box 1663, Mail Stop: J552 Los Alamos, New Mexico 87545 505-667-6701/Fax: 505-665-7913

Date: May 30, 2006 Refer to: N2-06: 147

Samuel J. Collins, Regional Administrator Nuclear Regulatory Commission 475 Allendale Road King of Prussia, PA 19406-1415

Re: ATRC#2006:33 & ASRO#2006:36, License Number SNM-203

Dear Mr. Collins:

Enclosed please find two signed Authorization to Transfer/Relinquishment of Custody form & a Ship/Relinquishment of Ownership form concerning:

Lucent Technologies 600 Mountain Ave., Room 7B-514 Murray Hill, NJ 07974

The radioactive sources described on the forms have been removed from Lucent Technologies in Murray Hill, NJ and are in storage at Los Alamos National Laboratory and at NSSI Sources & Services, Inc. These sources have been transferred to Department of Energy (DOE) ownership and are being stored under DOE license exemption.

This action was completed as part of the Off-Site Source Recovery (OSR) Project managed by this office. If you need any further information on this action, please contact the OSR Project Office at 505/667-6701.

Sincerely,

Shelby Leonard

Team Leader OSR Project

SL/lh

Cy:

Shaun Van Doren, Lucent Technologies Michele Burgess, NRC HQ Robert Campbell, U.S. DOE, NA-211 Joel Grimm, DOE-AL/WMD N2 File OSRP File

# Los Alamos

**NATIONAL LABORATORY** 

### ATRC # 2006:33

### **Off-Site Source Recovery Project** Authorization to Transfer/Relinquishment of Custody

SOURCE CUSTODIAN: Lucent Technologies

TELEPHONE:

908-582-7586

**CONTACT NAME:** 

Shaun Van Doren

FAX:

908-582-3904

ADDRESS:

600 Mountain Ave.

Room 7B-514

Murray Hill, NJ 07974

Lucent Technologies is authorized by the University of California at Los Alamos National Laboratory (UC/LANL), on behalf of the U.S. Department of Energy (DOE), to transfer the sealed source(s) indicated below to Los Alamos National Laboratory.

Lucent Technologies hereby relinquishes and UC/LANL, on behalf of DOE, hereby accepts all ownership of the sealed source(s) indicated below upon receipt or acceptance of the source(s) at Los Alamos National Laboratory.

#### SOURCE INFORMATION

Type:	Model:	Serial No:	Date of Mfg:	Grams Isotopic:
239Pu/Be		80S12		7.949
SOURCE CUS	STODIAN:			
OFFICIAL NA	ME/TITLE: Sha	AME (Please Print or Type)		5 Protect Specialist E (Please Print or Type)
SIGNATURE		Avin (Flease Finit of Type)		E <u>5/10/06</u>
UC/LANL AU	THORIZATION:	,		
OFFICIAL NA	ME/TITLE: She	by J. Leonard, Team Leader, Of	f-Site Source Rec	overy Project
SIGNATURE	Thuly J	Leonard	DAT	E 5-15-06
ACKNOWLE	DGEMENT OF REC	EIPT BY LANL:		
OFFICIAL NA		ARLO 5 RINCON  (AME (Please Print or Type)	TITI	DIATION SPECIALIS  LE (Please Print or Type)
SIGNATURE	Congl		DAT	E 5-16-06

# Los Alamos

NATIONAL LABORATORY

### ASRO # 2006:36

# Off-Site Source Recovery Project Authorization to Ship/Relinquishment of Ownership

SOURCE OWNER:

Lucent Technologies

TELEPHONE:

908-582-7586

LICENSE:

1 NRC - Region I No.: SNM 203 FAX:

908-582-3904

CONTACT NAME:

Shaun Van Doren

ADDRESS:

600 Mountain Ave.

Room 7B-514

Murray Hill, NJ 07974

Lucent Technologies is authorized by the University of California at Los Alamos National Laboratory (UC/LANL), on behalf of the U.S. Department of Energy (DOE), to transfer the sealed source(s) indicated below to the following designated UC/LANL support subcontractor:

NSSI / Sources & Services, Inc.

Texas Department of State Health Services

Radioactive Material License No. L02991

Lucent Technologies hereby relinquishes and UC/LANL, on behalf of DOE, hereby accepts all ownership of the sealed source(s) indicated below upon receipt or acceptance of the source(s) by the designated UC/LANL support subcontractor.

#### SOURCE INFORMATION

Type:	Model:	Serial No:	Activity (Ci):
241Am/Be		MRC-AmBe -3520	0.033
SOURCE OWN			
OFFICIAL NAM		An Doren	TITLE (Please Print or Type)
SIGNATURE _	. NAME	(Please Print or Type)	TITLE (Please Print or Type) DATE 5/10/06
UC/LANL AUT	HORIZATION:		
OFFICIAL NAM	ME/TITLE: Shelby J. 1	Leonard, Team Leader, Off-Site So	urce Recovery Project
SIGNATURE _	Shelly Jak	emand	DATE <u>5-15-06</u>
			•
		BY DESIGNATED UC/LANL S TRACTOR: NSSI / Sources & Ser	
OFFICIAL NAM	ME/TITLE: CARLOS	RINCON	RADIATION SPECLALIST
	NAME	(Please Print or Type)	TITLE (Please Print or Type)
SIGNATURE	Calin		DATE 5-16-06

MH Source Leak Testing

John E. Riley 8/17/05

Wipe tests for removable radioactive contamination were conducted 8/3-9/05 on sealed sources in the 1G Radiation Lab. The leak test materials were counted for radioactivity on the Oxford LB5100 low background alpha/beta system. The system has been calibrated using activity standards whose accuracy is traceable to NIST. The results of the analyses are as shown in the following table.

Count	Nuclide	Nuclide	Result	1
#	Inventory #			
1	14-0001	Co-60	<mda< td=""><td>;<b>*</b></td></mda<>	; <b>*</b>
2	14-0175	Am-241	<mda< td=""><td></td></mda<>	
3	14-0551	Co-60	<mda< td=""><td></td></mda<>	
4	14-0978	Pu-239	<mda< td=""><td>- PUBE - AMBE</td></mda<>	- PUBE - AMBE
5	14-1408	Am-241	<mda< td=""><td>- AMBE</td></mda<>	- AMBE
6	14-2052	Am-241	<mda< td=""><td></td></mda<>	
	Alpha	MDA	<0.0001 μCi	
	· Beta/Gamma	MDA	<0.001 μCi	

There was no detectable radioactivity in the leak tests. The leakage is less than the 0.005  $\mu\text{Ci}$  required under the NRC licenses.

John E. Riley 12/14/05

### MH Source Leak Testing

Wipe tests for removable radioactive contamination were conducted 12/14/2005 on sealed sources on the inventory. The leak test materials were counted for radioactivity on the Oxford LB5100 low background alpha/beta system. Prior to counting, the system was calibrated using activity standards whose accuracy is traceable to NIST – 14-2787, 14-2788, 14-2789. The results of the analyses are as shown in the following table.

	Count	Nuclide	Nuclide	Result	39
	#	Inventory #			
	. 1	20-0001	Ni-63	<mda< td=""><td></td></mda<>	
	2	20-0002	Ni-63	<mda< td=""><td></td></mda<>	
	3	20-0003	Ni-63	<mda< td=""><td></td></mda<>	
	4	14-1809	Co-60	<mda< td=""><td></td></mda<>	
	5	14-2278	Cs-137	<mda< td=""><td></td></mda<>	
	6	14-2120	Ni-63	<mda< td=""><td></td></mda<>	
	7	14-2121	Ni-63	<mda< td=""><td></td></mda<>	
,	8 -	14-2128	Fe-55	<mda< td=""><td></td></mda<>	
	9	14-0001	Co-60	<mda< td=""><td></td></mda<>	
	10	14-0175	Am-241	<mda< td=""><td></td></mda<>	
	11	14-0551	Co-60	<mda< td=""><td>0 0</td></mda<>	0 0
- [	12	14-0978	Pu-239	<mda< td=""><td>— PuBε — AmBε</td></mda<>	— PuBε — AmBε
	13	14-1408	Am-241	<mda< td=""><td>- AMBE</td></mda<>	- AMBE
	14	14-2052	Am-241	<mda< td=""><td></td></mda<>	
		Alpha	MDA	<0.0001 μCi	
		Beta/Gamma	MDA	<0.001 μCi	

There was no detectable radioactivity in the leak tests. The leakage is less than the 0.005  $\mu\text{Ci}$  required under the NRC licenses.