A
030-35639

4/28/06

Nuclear Regulatory Commission
Region III
ATTN: Bill Reichhold
2443 Warrenville, Suite 210
Lisle, IL 60532-4352
630-829-9839

RE: Radioactive materials license number 13-32301-01

Dear Bill Reichhold:

1. We have completed our physical move of the facility to the new site. All radioactive materials have been removed from the old site. A close-out survey (ambient) and (wipe) was performed at the old site. I have enclosed the results of these surveys.
2. Please amend item 10. of the license and delete 4506 South Western Avenue, Marion, Indiana.

I would like to be notified as soon as possible as I am waiting to release the old site for unrestricted use.

Should you have any question regarding this amendment, please contact me at 765-651-9347 or Mark Dater (NC Systems, Inc) at 800-548-4024.

Sincerely,

Gary A. Frick, D.O.
Radiation Safety Officer
Cardiovascular Consultants OF MARION, P.C.
1123 North Western Ave
Marion, IN 46952

315421
REC'D MAY 10 2006

Your Hospital Name
Your Hospital Address
Your Hospital City and State
Your Hospital Phone Number

Wipe Results Report

Thursday, April 27, 2006 2:05 PM

Area: MARION, IN - DECOMMISION

Unrestricted Area, RSO Trigger at 100% of Federal Guidelines, Well

Location: AREA #1

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
Yes	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	154 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-103 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	26 dpm

Location: AREA #2

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-33 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-63 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	-178 dpm

Location: AREA #3

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-8 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-55 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	-93 dpm

Location: AREA #4

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-59 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-87 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	-238 dpm

Location: AREA #5

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-59 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	32 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	-161 dpm

Location: AREA #6

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-33 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-23 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	43 dpm

Location: AREA #7

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-59 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-71 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	-84 dpm

Location: AREA #8

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
[illegible]	[illegible]	[illegible]	[illegible]	57 dpm	657 dpm	60 sec.	94 dpm

Your Hospital Name
Your Hospital Address
Your Hospital City and State
Your Hospital Phone Number

Wipe Results Report

Thursday, April 27, 2006 2:05 PM

Area: MARION, IN - DECOMMISION

Unrestricted Area, RSO Trigger at 100% of Federal Guidelines, Well

Location: AREA #9

>LLD	>Trigger	Isotope	Trigger	LLD	Background	Count Time	Activity
No	No	• Co-57	2000 dpm	57 dpm	657 dpm	60 sec.	-127 dpm
No	No	• Cs-137	200 dpm	42 dpm	351 dpm	60 sec.	-63 dpm
No	No	• Ba-133	200 dpm	71 dpm	1032 dpm	60 sec.	-93 dpm

Technologist: _____

Comments:

Signature: [signature] **Date:** 04-27-2006

Contamination Survey Record

Week

(Survey For Removable Contamination)

Survey Date: 04-25-2006 Time: 1:00 pm By: T. Euler Assay Date: ~~1:30 pm~~ 4-25-2006 Time: 1:30 pm By: T. Euler

Instrument: Bicron, I-537A, c658 H PHA H.V.: Full Open Window

Standard Radionuclide: _____ Assay Activity: _____ Date: _____

Survey Area



ALL READINGS IN MR/HR

Thomas Euler

Sample Analysis

Swipe #	Gross cpm	−	Bkg cpm	=	Net cpm	x	dpm* / cpm	=	dpm	Action Taken**

CARDIOVASCULAR CONSULTANTS

Gary A. Frick, D.O., M.S., F.A.C.C., F.A.C.O.I.

1123 North Western Avenue
Marion, IN 46952
Phone: 765 - 651 - 9347 Fax: 765 - 651- 9346

FACSIMILE TRANSMISSION

TO: Bill Reichhold	**FROM:** Gary Frick
FAX NUMBER: (630) 829-9782	**DATE:** 5/10/06
PHONE NUMBER: (630) 829-9839	**TOTAL NO. OF PAGES INCLUDING COVER:**

SUBJECT:

Radioactive materials license #13-32301-01

☐ **URGENT** ☑ **FOR REVIEW** ☐ **PLEASE REPLY**

NOTES/COMMENTS:

CONFIDENTIALITY NOTICE

The documents accompanying this fax transmission contain confidential information. The information is intended only for the use of the individual (s) or entity named above and may not be released to another individual or entity without express written consent of Cardiovascular Consultants, P.C. If you are not the intended recipient, you are notified that any disclosure, copying, distribution or the taking of any action in reliance on the contents of this information is not permissible. If you have received this fax in error or it was sent to the wrong fax #, please call (765) 651-9347 and ask to speak to the sender. Thank you.

1-650 ... regarding control # 315113

NC Systems, Inc.
Institute For Nuclear Medical Education, Inc.
5660 Airport Blvd., Suite 101, Boulder, Colorado 80301
Phone: (800) 548-4024 • (303) 541-0044 • FAX: (303) 541-0066

Stamp "Received" in The Area

FAX

☒ Standard ☐ Urgent **Total Pages:** 4

From: Mark Dater **To:** Becki **Date:** 4/28/06

FAX Phone: 765-651-9346 **Phone:** 765-651-9347 **Location:** IN

Reference:

Dear Becki:

I am faxing the survey results of the close-out surveys.
Have Dr. Frick sign and fed -ex to NRC ASAP
Please call if you have any questions.

Sincerely,

Mark
NC Systems, Inc.

Sender 1076 OFFICE - Fax 98

This message is intended for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, and exempt from disclosure under applicable law. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. ... by telephone and return the original message to us at the address above via the U.S. Postal services. Thank you.