

U.S. NUCLEAR REGULATORY COMMISSION	Date: 8/24/06
TELEPHONE CONVERSATION RECORD	Time: 9:10 a.m.
Mail Control 2006-001 License No(s). 29-00102-07	Docket No(s). 03020802
or Report No(s).	
Name of Licensee:	Newark Beth Israel Medical Center
Name of Participant(s):	Michael Mink
Telephone No.	(908)788-9440
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Waste return at St. Barnabas
Summary:	Mr. Mink is a consultant to Newark Beth Israel Medical Center. According to him, the Bicron portal monitor used to survey outgoing waste is checked monthly, but is not calibrated. The trigger level is twice background. Workers are told to wait 15 seconds before moving the waste from in front of the detector. If anything is detected they are to hold it for 3 days.
Action Required:	Send to ADAMS.
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date
Prepared & SUNSI Review Completed By:	rwm1 / RA / Date: 8/24/08