

<b>U.S. NUCLEAR REGULATORY COMMISSION</b>		Date: 8/15/2006
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 14:00
Mail Control or Report No(s).	139199	License No(s). 29-31023-01 Docket No(s). 03036880
Name of Licensee:	Cardiac Imaging for Life, P.C.	
Name of Participant(s):	Erin Uciechowsky, Office Manager Steven Courtemanche, Health physicist NRC	
Telephone No.	212-593-5757 610-337-5075	
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Request for termination of the license.	
Summary:	<p>I contacted the license to inform them that the request to terminate the license should include a close-out survey and a description as to whom the sources were transferred. Ms. Uciechowsky stated that the provider of the equipment had the information and he would get back in touch me. I stated that the letter from the consultant should be included as an enclosure but the cover letter should be signed by the licensee.</p> <p>16:00 A representative from Nuclear Diagnostic Products called and asked what sort of information was needed to close out the client's license. I informed him that I needed the results of the removable contamination surveys and ambient radiation level surveys as well as the disposition of the sealed sources that the licensee possessed. The information should come from the licensee with a signed cover letter.</p>	
Action Required:	Place in ADAMS	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SUNSI Review Completed By:	Steven Courtemanche	/ RA /      Date: 8/25/2006