

RECEIVED
REGION 1

2006 AUG 24 PM 2: 30

August 21, 2006

Br. 1

U.S. Nuclear Regulatory Commission
Region I
ATTN: Medical Licensing Assistance Section
475 Allendale Road
King of Prussia, PA 19406

03003162
Amendment to NRC License No. 37-12486-01

Dear Sir/Madam:


We want to amend our license to include the use of Cs-131 sealed sources for brachytherapy.

The vendor is IsoRay, Inc., 350 Hills Street, Suite 106, Richland, WA 99352. The device name is Lawrence CSERION Cs-131 Brachytherapy Seed and the *model is CS-1*.

Since we may be treating several patients simultaneously we need to increase the maximum possession limit for 35.400 radionuclides from *2000mCi to 4000mCi*.

If you require any additional information please contact the Radiation Safety Officer, Dr. Gary Coren.

Sincerely,



Roy Powell
President & Chief Executive Officer

cc. Gary Coren, M.D., RSO

This is to acknowledge the receipt of your letter/application dated

8/21/2006, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment 37-12486-01 There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 139319.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.