



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

September 5, 2006

Docket No. 03032514
Control No. 139270

License No. 37-28696-01

Allen Glotfelty
Executive Director
Associated Cardiologists, P.C.
2808 Old Post Road
Harrisburg, PA 17110

SUBJECT: ASSOCIATED CARDIOLOGISTS, P.C., REQUEST FOR ADDITIONAL
INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO
LICENSE, CONTROL NO. 139270

Dear Mr. Glotfelty:

This is in reference to your letter dated August 7, 2006 requesting to amend Nuclear Regulatory Commission License No. 37-28696-01. In order to continue our review, we need the following additional information:

1. The preceptor, Harry J. Lessig, M.D., used the criteria in Subpart J of 10 CFR 35 for attesting Christopher Spizzieri, D. O. qualification as an authorized user. The definition of an authorized user as written in 10 CFR 35.2 only allows the use of Subpart J to meet these requirements before October 24, 2005. Many of the attributes are the same, however one attribute was not previously included in Subpart J for 35.200 users. Please have the preceptor, Harry J. Lessig, M.D. attest to the following work experience under his supervision: Administering dosages of radioactive drugs to patients or human research subjects (10 CFR 35.290(c)(ii)(F)).
2. The training requirements for use of sealed sources for diagnosis can be found in 10 CFR 35.590. Please submit information that Christopher Spizzieri, D. O. has completed training in the use of the device for patient attenuation correction.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Toolkit Index Page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 139270. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5366.

A. Glotfelty
Associated Cardiologists, P.C.

2

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Dennis R. Lawyer

Dennis R. Lawyer
Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

cc:
Robert A. Skotnicki, D.O., Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML062490028.wpd

SUNSI Review Complete: DLawyer

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	DNMS/RI	<input type="checkbox"/>		
NAME	DLawyer/DRL							
DATE	09/05/2006							

OFFICIAL RECORD COPY