



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

October 3, 2006

Docket No. 03001242
Control No. 139141

License No. 06-00649-03

Harry Evert
Vice President, Administration
Middlesex Hospital
28 Crescent Street
Middletown, CT 06457

SUBJECT: MIDDLESEX HOSPITAL, LICENSE AMENDMENT, CONTROL NO. 139141

Dear Mr. Evert:

This refers to your license amendment request. Enclosed with this letter is the amended license. Please note that, based on my conversations with your Radiation Safety Officer, Joan Merton, the strontium-90 eye applicator has been listed separately on your license for storage only incident to disposal. If you wish to use this device in the future, you must submit an amendment request to NRC for such use.

Please review the enclosed document carefully and be sure that you understand and fully implement all the conditions incorporated into the amended license. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region I Office, Licensing Assistance Team, (610) 337-5239, so that we can provide appropriate corrections and answers.

An environmental assessment for this action is not required, since this action is categorically excluded under 10 CFR 51.22(c)(14).

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Toolkit Index Page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

Thank you for your cooperation.

Sincerely,

Original signed by Donna M. Janda

Donna M. Janda
Health Physicist
Medical Branch
Division of Nuclear Materials Safety

Enclosure:
Amendment No. 50

cc:
Joan Merton, Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML062760342.wpd

SUNSI Review Complete: DJanda

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	DNMS/RI	<input type="checkbox"/>		
NAME	DJanda/DMJ							
DATE	10/3/2006							

OFFICIAL RECORD COPY

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Middlesex Hospital</p> <p>2. 28 Crescent Street Middletown, Connecticut 06457</p>	<p>In accordance with the letter dated July 12, 2006,</p> <p>3. License number 06-00649-03 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date June 30, 2013</p> <hr/> <p>5. Docket No. 030-01242 Reference No.</p>
---	--

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Any byproduct material permitted by 10 CFR 35.100</p> <p>B. Any byproduct material permitted by 10 CFR 35.200</p> <p>C. Any byproduct material permitted by 10 CFR 35.300</p> <p>D. Any byproduct material permitted by 10 CFR 35.400</p> <p>E. Strontium 90</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p> <p>B. Any</p> <p>C. Any</p> <p>D. Sealed Sources (3M Models 6D 6C-CA and 6500 Series; Best Medical International, Inc. Model 2301)</p> <p>E. Sealed Source (ICN Tracer Lab SN 1274)</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. As needed</p> <p>B. As needed</p> <p>C. 1000 millicuries</p> <p>D. 2000 millicuries</p> <p>E. 60 millicuries</p>
---	---	--

<p>9. Authorized use:</p> <p>A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.</p> <p>B. Any imaging and localization study permitted by 10 CFR 35.200.</p> <p>C. Any diagnostic study or therapy procedure permitted by 10 CFR 35.300.</p> <p>D. Any manual brachytherapy procedure permitted by 10 CFR 35.400.</p> <p>E. For storage only incident to disposal.</p>

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
06-00649-03

Docket or Reference Number
030-01242

Amendment No. 50

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 28 Crescent Street, 530, 534 and 536 Saybrook Road, Middletown, Connecticut.
11. The Radiation Safety Officer for this license is Joan Merton.
12. Licensed material is only authorized for use by, or under the supervision of:
- A. Individuals permitted to work as an authorized user or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for medical use as indicated:

Authorized Users

Material and Use

Kenneth I. Bird, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Ravi Jain, M.D.

35.100; 35.200; Oral administration of Iodine-131 in quantities less than or equal to 33 millicuries

Allen B. Silberstein, M.D.

35.300; 35.400

Joseph B. Weissberg, M.D.

35.300; 35.400

Diana Miller Hull, M.D., Ph.D.

35.100; 35.200; Oral administration of Iodine-131 in quantities less than or equal to 33 millicuries

Michael Crain, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Ramon Gonzales, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Eric J. Mannes, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Gregory N. Viscomi, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Craig A. Walden, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Erik Pingoud, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Jeffrey Takahashi, M.D.

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

Robert Wolek, M.D.

35.100; 35.200; 35.300

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
06-00649-03

Docket or Reference Number
030-01242

Amendment No. 50

Authorized Users

Caridad E. Fuertes, M.D.

Todd J. Everett, M.D.

Material and Use

35.100; 35.200; Oral administration of Iodine-131 in quantities less than or equal to 33 millicuries

35.100; 35.200; Iodine-131 sodium iodide for imaging and localization studies

13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.
14. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
15. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated May 22, 2003 (ML031500504)
- B. Letter dated June 16, 2003 (ML031690461)

For the U.S. Nuclear Regulatory Commission

Original signed by Donna M. Janda

Date October 3, 2006

By

Donna M. Janda
Medical Branch
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406