

USNRC
REGION I
Nuclear Materials Section I
475 Allendale Road
King of Prussia, PA 19406

September 26, 2006

Br. 2

Dear Sir:

Enclosed is a survey of our old Nuclear Medicine Hot Lab, we would like the area released for unrestricted use. This is an amendment of our Byproduct Material License No. 37-28414-01. 03031251

- 1) Instrument used for ambient surveys:
Ludlum GM Counter
Meter: Model # 14C
Serial # 67389
Probe: Model EWGM
Serial # 058703 Calibration date 07/14/06
Trigger level for restricted area is 2mR/hr.
Ambient levels for all areas =< .03mR/hr.
- 2) Instrument used for wipe tests:
Sun Nuclear Wipe Counter
SNC Model 1051
Serial # 3165906
Efficiency = 0.6% pass=< 2000DPM

2006 SEP 29 PM 12: 15

RECEIVED
REGION I

Areas surveyed for wipe test:


	Wipe Test	Ambient level results
1. Outside door knob	pass	<.03mR/hr
2. Inside door knob	pass	<.03mR/hr
3. Receipt area	pass	<.03mR/hr
4. Work counter	pass	<.03mR/hr
5. Hot Lab door	pass	<.03mR/hr
6. Waste storage #1	pass	<.03mR/hr
7. Waste storage #2	pass	<.03mR/hr
8. Waste storage #3	pass	<.03mR/hr
9. Cupboard drawer # 1	pass	<.03mR/hr
10. Cupboard drawer # 2	pass	<.03mR/hr
11. Counter drawer #1	pass	<.03mR/hr
12. Counter drawer #2	pass	<.03mR/hr
13. Counter drawer #3	pass	<.03mR/hr
14. Counter drawer #4	pass	<.03mR/hr
15. Counter drawer #5	pass	<.03mR/hr
16. Counter drawer #6	pass	<.03mR/hr
17. Counter drawer #7	pass	<.03mR/hr
18. Counter drawer #8	pass	<.03mR/hr

139482
NMSS/RGNI MATERIALS-002

19.Counter drawer #9	pass	<.03mR/hr
20.Counter drawer #10	pass	<.03mR/hr
21.Counter drawer #11	pass	<.03mR/hr
22.Counter drawer #12	pass	<.03mR/hr
23.Counter drawer #13	pass	<.03mR/hr
24.Counter drawer #14	pass	<.03mR/hr
25.Counter drawer #15	pass	<.03mR/hr
26.Counter drawer #16	pass	<.03mR/hr

Your prompt attention to this matter would be greatly appreciated so we can proceed with construction plans.

Thank you,



Kishor Joshi, MD
Facility President

This is to acknowledge the receipt of your letter/application dated

9/26/2006, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment 37-28414-01 There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 139482.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.