

**From:** Willie Lee  
**To:** murshed.hossain@mail.tju.edu  
**Date:** Wed, Oct 4, 2006 3:02 PM  
**Subject:** Request for Additional Information, Control 139319

REQUEST FOR ADDITIONAL INFORMATION

LICENSEE: Frankford Health System  
LICENSE NUMBER: 37-12486-01  
DOCKET NUMBER: 030-03162  
CONTROL NUMBER: 139319  
DATE: October 4, 2006

ATTN: Murshed Hossain, Medical Physicist

Additional information is needed for completion of the license amendment request (dated 08/21/06) submitted by Roy Powell, President and Chief Executive Officer, as follows:

1. Please describe how you will assure that members of the public will receive less than 500 milli-Rem exposure from a released patient containing Cesium-131 brachytherapy sources.

Please note that you may not reply to this request by return e-mail. Your prompt reply must be in writing by letter, signed by a member of Senior Management, which can be sent by facsimile to (610) 337-5269. Contact me if there are any questions regarding this request.

Willie J. Lee, Health Physicist  
US Nuclear Regulatory Commission  
Region 1  
475 Allendale Road  
King of Prussia, PA 19406  
(610) 337-5090 (Phone)  
(610) 337-5269 (FAX)

**Mail Envelope Properties** (45240562.613 : 7 : 26753)

**Subject:** Request for Additional Information, Control 139319  
**Creation Date** Wed, Oct 4, 2006 3:02 PM  
**From:** Willie Lee

**Created By:** WJL1@nrc.gov

**Recipients**  
mail.tju.edu  
murshed.hossain (murshed.hossain@mail.tju.edu)

**Post Office**

**Route**  
mail.tju.edu

<b>Files</b>	<b>Size</b>	<b>Date &amp; Time</b>
MESSAGE	1841	Wednesday, October 4, 2006 3:02 PM

**Options**

**Expiration Date:** None  
**Priority:** Standard  
**ReplyRequested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard