

Reviewer Notes - Jonathan Rivera
International Isotopes
SSD Case No. 06-44
NR-1235-S-103-S

JR 10/3/06

1. In 4/17/06 application, International Isotopes stated that the INIS-SF-CS-1J and INIS-SF-CS-2J sources meet ANSI/HPS N43.6-1997 classification of ANSI 96C66546. However, according to the testing report provided as attachment to email dated 7/7/06, the classification should be ANSI 96C63546. This report provided, on page 11 of 36, external pressure results for a classification of "3" under Table 1 of ANSI/HPS N43.6-1997. International Isotopes provided additional pressure test information starting on page 29 of 36 of this report. However, this additional information did not include an integrity test as indicated under section 7.3.3 of ANSI/HPS N43.6-1997.
2. SSD case no. 06-44 is a continuation of 06-25 which contains an SSD checklist. Therefore, checklist not included in 06-44.