

mslk  
JI

# Cardiac Testing Center, Inc.

29 South Street  
New Providence, New Jersey 07974  
201-665-9500

October 2, 1991

Jenny M. Johansen, Chief  
Nuclear Materials Safety Section D  
Division of Radiation Safety and Safeguards  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road  
King of Prussia, PA 19406

Re: Cardiac Testing Center  
License No.: 29-20960-01


Dear Ms Johansen:

The response to information requested by Ms Michelle Beardsely is as follows. Replacement pages for the application being reviewed are included. The original page may be discarded.

1. We do not use, nor do we plan to use, Mo-99/Tc-99m generators. We use only unit doses supplied by a radiopharmacy.
2. The range of the survey meters in our possession may be found in the original application under Attachment 9.6. The lowest range is, of course, on the order of 0.01 mR/hr.
3. A list of companies that we will use for calibrating our survey meters is enclosed. This page shall replace the page in our original license application.
4. Trigger levels have been revised and the page corresponding to the application is enclosed to replace the original page.

Please contact me at (908) 665-9500 if you have additional questions.

Sincerely,

  
Mark J. Krell, M.D.  
Radiation Safety Officer  
and Partner

OFFICIAL RECORD COPY ML 10

114342  
OCT 07 1991