

<b>U.S. NUCLEAR REGULATORY COMMISSION</b>		Date: 10/11/2006
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 1:00 PM
Mail Control or Report No(s).	139482	License No(s). 37-28414-01 Docket No(s). 03031251
Name of Licensee:	Cherry Tree Medical Associates	
Name of Participant(s):	Mary Chamberlin, Imaging Supervisor	
Telephone No.	724-439-4150	
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	RAI, Purpose of Request	
Summary:	Based on our letter dated January 13, 2006, we stated that the old hot lab could be released for unrestricted use. Why have you submitted a similar request. The hot lab was modified again and has been changed again. She said that there was information submitted earlier, however I did not locate any in Adams. I discussed the notification requirement in 10 CFR 35.14(b)(4) in which they just need to notify of us the changes within 30 days. I discussed information important to decommissioning, however the licensee only uses sealed sources and technicum 99m. I stated that we needed a diagram of what the areas of material use would be.	
Action Required:	She will send in the diagram of material usage.	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SUNSI Review Completed By:	Dennis Lawyer	/ RA /      Date: 10/11/2006