



DONOHUE CARDIOLOGY ASSOCIATES

EXCELLENCE IN VASCULAR HEALTH

BOARD CERTIFIED PHYSICIANS IN CARDIOLOGY

September 20, 2006

CARDIOLOGISTS

BRYAN C. DONOHUE, MD
CHRISTOPHER C. ALLEN, MD
KIRK F. MUSSELMAN, MD
SANJAYA N. SAHETA, MD
JOHN F. PENSOCK, MD, FACC
LINDA GORDON, CRNP
MONCEL LOGAN-DEITZ, PA-C
TAMMY FENDYA, CRNP

OFFICES

WASHINGTON
50 BERRY ROAD
WASHINGTON, PA 15301
PHONE: (724) 222-1125
FAX: (724) 222-1373

SHADYSIDE
SHADYSIDE PLACE
580 S. AIKEN AVE., SUITE 430
PITTSBURGH, PA 15232
PHONE: (412) 235-1140
FAX: (412) 235-1141

MONROEVILLE
SUITE 208
2550 MOSSIDE BLVD
MONROEVILLE, PA 15146
PHONE: (412) 235-1140

BETHEL
1300 OXFORD DRIVE
BETHEL PARK, PA 15102
PHONE: (724) 222-1125
(412) 235-1140

CENTERVILLE
1070 OLD NATIONAL PIKE
FREDERICKTOWN, PA 15333
PHONE: (724) 632-5801

U.S. Nuclear Regulatory Commission
Materials Licensing Branch
Region I
475 Allendale Road
King of Prussia, PA 19406

J-6
MS-16

RE: Amendment Request for Radioactive Material license
No. 37-30735-01 03036027
Donohue Cardiology Associates

Gentlemen:

Please delete or Bethel Park office from our license effectively immediately. We have removed all radioactive materials and have performed a close out survey (enclosed).

If you have any questions or require additional information, please do not hesitate to contact the undersigned.

Sincerely,

Patricia R. Edwards
Administrator

Christopher Allen, MD
Vice-President, RSO

Perna Health Physics, Inc.

Leak Test Certificate

Facility: Donohue Cardiology Associates UPMC
 Dates: 30-Aug-06
 By: MTP

Source ID	SN	Isotope	QTY (mCi)	Cal Date	Test Results
E-vial	934-17-8	Ba-133	0.2818	1-Jul-02	<.005 uCi
E-vial	BM0637002-06	Cs-137	0.210	7-Apr-03	<.005 uCi
Flood	1185-118	Co-57 flood	10.000	1-Jun-06	<.005 uCi

Calibration Source Analysis

Counter: Caprac

Background (cpm)	
Co-57	Cs-137
428	428

Efficiency:	Isotope	DPM	Gross Ct	Efficiency	MDA
	Co-57	1057.64	78470	93.27%	103
	Cs-137	164522.95	46400	27.94%	115

Precision of measurements: 95% confidence level
 All sources NIST traceable with +/- 5% accuracy


Sample Analysis

Samples:	SN	Isotope	CPM	DPM	uCi
	934-17-8	Ba-133	420	-9	-3.86E-06
	BM0637002-06	Cs-137	420	-9	-3.86E-06
	1185-118	Co-57 flood	420	-9	-3.86E-06

Precision of Measurements: 95% Confidence Level

Mark T. Perna
 M.S., DABR

Digitally signed by Mark T. Perna, M.S., DABR
 DN: cn=Mark T. Perna, o=Perna Health Physics, Inc., ou=PHPI, email=mark@pernahealth.com, c=US



Perna Health Physics, Inc.

Radiation Safety Officer