



J-9
MS-16

**Respond to: Paul J. Early, Corporate RSO
P.O. Box 340
Bemus Point, NY 14712**

**PH: 716.386.3860
FX: 716.386.5118
VM: 858.537.2137**

October 6, 2006

VIA FAX (610-337-5769)
2

Tara Weidner
U.S. Nuclear Regulatory Commission, Region I
475 Allendale Road
King of Prussia PA 19406-1415

Re: July 19, 2006 Request for HGNC as a base location

License No. 31-30666-01
Docket No. 030-35802
Control No. 139297

Dear Tara,

Please rescind our above referenced letter to add HGNC Cardiology back as one of our base locations. We will no longer be performing imaging at that location. Although we have been operating there as a temporary job site for the past 6 months, I have included a copy of the last area survey performed during our last day at that location.

Sincerely,

Paul J. Early, DABSNM, DABPM

*Saved
10/9/06. SPJ.
10/10/06*



SYSTEM # CARDIUS II DAILY QC FORM

BASE LOCATION: HGNC Cardiology

10-01-06 10:48 FROM DIGIRAD

INSTRUMENTS USED:

Model	Serial Number
1. Ludlum 14C	213183
2. Ludlum 14C	210230
3. NaI probe 44-3	223526

DATE: 10/6/06

Response & Battery Acceptable? Yes No
 Response & Battery Acceptable? Yes No
 Response & Battery Acceptable? Yes No


INSTRUMENTS USED FOR:

Packages	Areas	Personnel	Initials
	<input checked="" type="checkbox"/>		<i>[Signature]</i>

PACKAGE RECEIPT:

Supplier	Nuclide	Package Units	Activity mBq	Dose Rate (msv/hr) 3 Feet Contact	Wipes (dpm)	Contents Verified	Facility of use	Initials
Cardinal Health	<u>99mTc</u>	<u>Radio Active white T</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>HGNC</u>	
Cardinal Health	<u>99mTc</u>	<u>Radio Active white T</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>HGNC</u>	
Cardinal Health	<u>99mTc</u>	<u>Radio Active white T</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>HGNC</u>	
Cardinal Health	<u>99mTc</u>	<u>Radio Active white T</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>HGNC</u>	
Cardinal Health	<u>99mTc</u>	<u>Radio Active white T</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>[Handwritten]</u>	<u>HGNC</u>	

AREA SURVEY RESULTS



Area	msv/hr	Net dpm
A. Prep Area	<u>0.03</u>	<u>[Handwritten]</u>
B. Injection Area	<u>0.03</u>	<u>[Handwritten]</u>
C. Radioactive Waste	<u>0.03</u>	<u>[Handwritten]</u>
D. Computer/Camera	<u>0.03</u>	<u>[Handwritten]</u>
E. Treadmill	<u>0.03</u>	<u>[Handwritten]</u>
F. Floor	<u>0.03</u>	<u>[Handwritten]</u>
G. Cold Trash	<u>0.03</u>	<u>[Handwritten]</u>
H. Other Hot Lab	<u>0.03</u>	<u>[Handwritten]</u>
I. Unrestricted Area	<u>0.03</u>	<u>[Handwritten]</u>
J. Background	<u>0.0</u>	<u>[Handwritten]</u>

Notify RSO if any survey is: 2 msv/hr
 Surveyor: Anderson Meter SN: 213183 2000 dpm

Individual	Hands msv/hr	Feet msv/hr	Clothing msv/hr
1.			
2.			
3.			
4.			

Package: Returned to Pharmacy	msv/hr Surface	msv/hr Background	dpm Surface
Package: Returned to Pharmacy	msv/hr Surface	msv/hr Background	dpm Surface
Package: Returned to Pharmacy	msv/hr Surface	msv/hr Background	dpm Surface
Package: Returned to Pharmacy	msv/hr Surface	msv/hr Background	dpm Surface

Threshold Limits:

	Unrestricted	Restricted
Surveys	.5 msv/hr	.5 msv/hr
Wipes	2000 dpm/100cm ²	20,000 dpm/100cm ²

close out survey

610-366-1483

T-562 P02/02 U-728