



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

October 12, 2006

Docket No. 03002631
Control No. 138313

License No. 32-04054-04

Colonel Colin M. Greene
U.S. Army, Deputy Commander for Clinical Services
Womack Army Medical Center
ATTN: MCXC-PMS-RP
(Health Physics/Radiation Protection Office)
Building #4-2817 Reilly Road
Fort Bragg, NC 28310

SUBJECT: WOMACK ARMY MEDICAL CENTER, CORRECTED COPY OF LICENSE,
CONTROL NO. 138313

Dear Colonel Greene:

Enclosed is the Corrected Copy of Amendment No. 36 for License No. 32-04054-04. In accordance with the telephone call on October 11, 2006, the name of Dr. Santiago-Maldonado has been corrected.

We apologize for any inconvenience this error may have caused.

Sincerely,

Original signed by Pamela J. Henderson

Pamela J. Henderson, Chief
Medical Branch
Division of Nuclear Materials Safety

Enclosure:
Corrected Copy of Amendment No. 36

cc:
Captain Margaret Myers, Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML062850546.wpd

SUNSI Review Complete: S Xu

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI	N		
NAME	SXu/Sx		PLanzisera/PL		PHenderson/PJH			
DATE	10/12/2006		10/12/2006		10/12/2006			

OFFICIAL RECORD COPY

CORRECTED COPY

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Womack Army Medical Center ATTN: MCXC-PMS-RP (Health Physics/Radiation Protection Office)</p> <p>2. Building #4-2817 Reilly Road Fort Bragg, North Carolina 28310</p>	<p>In accordance with the letter dated January 19, 2006,</p> <p>3. License No. 32-04054-04 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration Date: February 28, 2012</p> <hr/> <p>5. Docket No. 030-02631</p>
---	---

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Any byproduct material permitted by 10 CFR 35.100</p> <p>B. Any byproduct material permitted by 10 CFR 35.200</p> <p>C. Any byproduct material permitted by 10 CFR 35.300</p> <p>D. Any byproduct material permitted by 10 CFR 31.11</p> <p>E. Hydrogen 3</p> <p>F. Carbon 14</p> <p>G. Phosphorus 32</p> <p>H. Sulfur 35</p> <p>I. Iodine 125</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p> <p>B. Any</p> <p>C. Any</p> <p>D. Prepackaged Kits</p> <p>E. Any</p> <p>F. Any</p> <p>G. Any</p> <p>H. Any</p> <p>I. Any</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. As needed</p> <p>B. As needed</p> <p>C. As needed [not to exceed 1.5 curies (Ci) of I-131]</p> <p>D. As needed</p> <p>E. 100 millicuries (mCi)</p> <p>F. 90 mCi</p> <p>G. 10 mCi</p> <p>H. 10 mCi</p> <p>I. 50 mCi</p>
---	--	--

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

CORRECTED COPY

License No.
32-04054-04

Docket No.
030-02631

Amendment No.
36

- | | | |
|---|--|--|
| 6. Byproduct, source, and/or special nuclear material | 7. Chemical and/or physical form | 8. Maximum amount that licensee may possess at any one time under this license |
| J. Gadolinium 153 | J. Sealed source (North American Scientific Model MED3601; DuPont Merck Model NES-8412; Isotope Products Labs Model A3410) | J. 1 Ci total, no single source to exceed 300 mCi |
| K. Gadolinium 153 | K. Sealed source (Isotope Products Labs Model 301B; Du Pont Merck Model NES-8424) | K. 4 Ci total, no single source to exceed 323 mCi |
| L. Barium 133 | L. Sealed source (Isotope Products Labs Model PHI-133 GFS Series) | L. 46 mCi |
-
9. Authorized use:
- A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.
 - B. Any imaging and localization study permitted by 10 CFR 35.200.
 - C. Any diagnostic study or therapy procedure permitted by 10 CFR 35.300.
 - D. In vitro studies.
 - E. - I. Research and development as defined in 10 CFR 30.4; animal studies.
 - J. For possession and use in ADAC Laboratories "Vantage" Nonuniform Attenuation Correction System in/on gamma cameras for medical use. For storage in shipping container pursuant to source exchange.
 - K. For possession and use in SMV International Transmission Attenuation Correction Source Holder Model No. PS 96 on gamma cameras for medical use. For storage in shipping container pursuant to source exchange.
 - L. For possession and use in Marconi (formerly Picker) Beacon Model PHI-0094 and/or N211XXX non-linear attenuation correction device for medical use.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

CORRECTED COPY

License No.
32-04054-04

Docket No.
030-02631

Amendment No.
36

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at Womack Army Medical Center, Building #4-2817 Reilly Road, Fort Bragg, North Carolina.
11. The Radiation Safety Officer (RSO) for this license is Captain Margaret C. Myers.
12. Licensed material is only authorized for use by, or under the supervision of:
- A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for medical use as indicated:

<u>Authorized User</u>	<u>Material and Use</u>
Nikolaos Tsolomitis Lomis, M.D.	35.100; 35.200
Hillarie Ann Saul, M.D.	35.100; 35.200
Elisa C. K. Morgan, M.D.	35.100; 35.200
Kenneth Alan Griggs, M.D.	35.100; 35.200
Johan K. Ahn, M.D.	35.100; 35.200
Fred Anthony Caruso, M.D.	35.100; 35.200
Charles Bernard Gantt, Jr., M.D.	35.100; 35.200
Faheem H. Hussain, M.D.	35.100; 35.200
Vimal K. Sodhi, M.D.	35.100; 35.200; 35.300
Antonie Martinus Romyn, M.D.	35.100; 35.200; 35.300
Joseph Carvalho, Jr., M.D.	35.100; 35.200
Patricia M. Kulas, M.D.	35.100; 35;200; 35.300
Sun Yong Kim, M.D.	35.100; 35;200; 35.300
Ida M. Santiago-Maldonado, M.D.	35.100; 35.200; Oral administration of sodium iodide iodine-131; parenteral administration of any beta emitter or photon-emitting radionuclide with a photon energy less than 150 keV

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

CORRECTED COPY

License No.
32-04054-04

Docket No.
030-02631

Amendment No.
36

C. The following individuals are authorized users for non-medical uses as indicated:

User

Material and Use

Captain Margaret C. Myers

All, as part of performing radiation safety duties

D. Licensed material specified in Items 6.D.-I., shall be used by, or under the supervision of, individuals designated by the licensee's RSC and RSO. Designated personnel shall meet the training and experience criteria specified in 10 CFR 33.15. The licensee shall maintain records of persons designated as users.

13. Experimental animals administered licensed material or their products shall not be used for human consumption.
14. Licensed material shall not be used in products distributed to the public.
15. Installation, initial radiation survey, relocation, or removal from service of devices containing sealed sources shall be performed by persons specifically licensed by the Commission or an Agreement State to perform such services. Maintenance and repair of devices and installation, replacement, and disposal of sealed sources shall be performed only by persons specifically licensed by the Commission or an Agreement State to perform such services.
16. The licensee is authorized to hold radioactive material with a physical half-life of less than 120 days for decay-in-storage before disposal without regard to its radioactivity if it:
- A. Monitors byproduct material at the surface before disposal and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding; and
 - B. Removes and obliterates all radiation labels, except for radiation labels on materials that are within containers and that will be managed as biomedical waste after they have been released from the licensee; and
 - C. Maintains records of the disposal of licensed material for three years. The record must include the date of the disposal, the survey instrument used, the background radiation level, the radiation level measured at the surface of each waste container, and the name of the individual who performed the disposal.
17. The licensee shall conduct a physical inventory every six months, or at other intervals approved by the U. S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET****CORRECTED COPY**License No.
32-04054-04Docket No.
030-02631Amendment No.
36

18. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
19. For sealed sources not associated with 10 CFR Part 35 use, the following conditions apply:
- A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U. S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
 - B. Notwithstanding Paragraph A of this Condition, sealed sources designed to primarily emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed three months.
 - C. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U. S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
 - D. Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.
 - E. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
 - F. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U. S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
 - G. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U. S. Nuclear Regulatory Commission or an Agreement State to perform such services.
 - H. Records of leak test results shall be kept in units of microcuries and shall be maintained for five years.
20. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

CORRECTED COPY

License No.
32-04054-04

Docket No.
030-02631

Amendment No.
36

21. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.
22. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U. S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- | | | |
|----|------------------------------------|---------------|
| A. | Application dated February 6, 1991 | [ML020420007] |
| B. | Application dated January 16, 2002 | [ML020280398] |
| C. | Letter dated December 5, 1991 | [ML020420005] |
| D. | Letter dated January 8, 1992 | [ML020390523] |
| E. | Letter dated March 12, 1993 | [ML020390522] |
| F. | Letter dated March 25, 1993 | [ML020390518] |
| G. | Letter dated December 17, 1993 | [ML020390515] |
| H. | Letter dated January 24, 1994 | [ML020390514] |
| I. | Letter dated October 21, 1994 | [ML020390512] |
| J. | Letter dated September 13, 1995 | [ML020390509] |
| K. | Letter dated March 24, 1997 | [ML020390497] |
| L. | Letter dated October 17, 1997 | [ML020390490] |
| M. | Letter dated November 21, 1997 | [ML020390487] |
| N. | Letter dated July 24, 1998 | [ML020390481] |
| O. | Letter dated September 2, 1998 | [ML020390477] |
| P. | Letter dated July 1, 1999 | [ML020390473] |
| Q. | Letter dated July 26, 1999 | [ML020390469] |
| R. | Letter dated February 14, 2000 | [ML003693701] |
| S. | Letter dated June 13, 2000 | [ML003725076] |
| T. | Letter dated February 13, 2001 | [ML010610394] |
| U. | Letter dated August 7, 2001 | [ML012640384] |
| V. | Letter dated December 10, 2001 | [ML013460075] |
| W. | Letter dated January 14, 2002 | [ML020280398] |
| X. | Letter dated July 15, 2003 | [ML033000535] |
| Y. | Letter dated August 5, 2004 | [ML042320155] |

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

CORRECTED COPY

License No.
32-04054-04

Docket No.
030-02631

Amendment No.
36



For the U. S. Nuclear Regulatory Commission

Date October 12, 2006

By *Original signed by Penny Lanzisera*
 Penny Lanzisera
 Medical Branch
 Division of Nuclear Materials Safety
 Region I
 King of Prussia, Pennsylvania 19406