

**From:** Penny Lanzisera  
**To:** karenwagner@catholichealth.net  
**Date:** Fri, Oct 13, 2006 2:16 PM  
**Subject:** Request for Additional Information for Amendment Request, MCN 139339

St. Joseph Medical Center  
L.N. 37-02763-01  
D.N. 030-03025  
Mail Control No. 139339

Dear Karen,

Please provide the following additional information:

- 1) Clarify whether the mailing address should be changed to P.O. Box 316 with this amendment.
- 2) Provide procedures for assuring rolling lead shields stay in place when used for in-patient radiopharmaceutical therapies and manual brachytherapy.
- 3) Confirm that the HDR unit is the only radiation producing unit used in the HDR treatment room.
- 4) Describe "each room where mHDR treatments will be given" or confirm that only the room described in the letter dated July 28, 2006, and the currently licensed HDR treatment room (until move of the HDR unit) will be used for treatments.
- 5) Confirm that the HDR unit storage box will be securely affixed to a permanent structure in the treatment room.

In addition, please note that determination of dosages in nuclear medicine must be performed in accordance with 10 CFR 35.63.

You may fax your reply to my attention at 610-337-5269. Please refer to Mail Control No. 139339 in your response.

Penny Lanzisera  
US NRC Region 1

**Mail Envelope Properties** (452FD7E5.757 : 17 : 55940)

**Subject:** Request for Additional Information for Amendment Request, MCN  
139339

**Creation Date:** Fri, Oct 13, 2006 2:16 PM

**From:** Penny Lanzisera

**Created By:** PAN@nrc.gov

**Recipients**

catholichealth.net

karenwagner (karenwagner@catholichealth.net)

**Post Office**

**Route**

catholichealth.net

**Files**

MESSAGE

**Size**

1862

**Date & Time**

Friday, October 13, 2006 2:16 PM

**Options**

**Expiration Date:** None

**Priority:** Standard

**ReplyRequested:** No

**Return Notification:** None

**Concealed Subject:** No

**Security:** Standard