



3105 LIMESTONE ROAD • SUITE 202 • WILMINGTON, DE 19806
(302) 994-6500 • Fax (302) 994-6922

5521 SILVERSIDE ROAD • SUITE 1-A • WILMINGTON, DE 19810
(302) 477-6500 • Fax (302) 477-6509

1100 FORREST AVE. • DOVER, DE 19904
(302) 672-1990 • Fax (302) 672-1899

550 SOUTH DUPONT PLAZA • MILFORD, DE 19968
(302) 422-6050 • Fax (302) 422-6620

1606 SAVANNAH RD. • SUITE 3 • LEWES, DE 19958
(302) 644-3522 • Fax (302) 645-1228

October 16, 2006

P-7

Licensing Branch
Nuclear Material Safety Section
Division of Safety and Safeguards
U.S.N.R.C. – Region 1
475 Allendale Road
King of Prussia, PA 19406

NRC License No. 07-27897-01 03029521

Dear Mr. Thompson:

Per our telephone conversation, you will find attached a copy of the sealed source leak test that was performed at our old office, 1606 Savannah Road, prior to transport to our new office at 16704 King's Highway.

The sealed sources were packaged tightly in a box and sealed with packing tape. The box was then placed inside the lead module used for isotope storage. The outside of the lead module surveyed at less than .05 mR/hr and was transported to the new office. The sealed sources consist of a 137-Cs vial source, 57-Co vial source, and 137-Cs rod source.

Please contact me at 302-994-6500 if you have any further questions. Thank you for your consideration in this matter.

Sincerely,

AnnMarie Messick, BS, NCT, CNMT
Director of Cardiac Diagnostic Center

139521
NMSS/RGNI MATERIALS-002

Cardiac Diagnostic Center
Nuclear Medicine Department
Lewes, Delaware

NRC # 07-27897-01
October 16, 2006
Page 1 of 1

Sealed Sources Leak Test : # 16

Test Date & Time : 09-30-2006 07:10
Instrument Model : 2200
Serial # : 164586
Background : 763.00 CPM
Instrum. Uniq# : 1
Comment :

Technologist : KJR
Manufacturer : Ludlum
Instrum Eff. : 9.81 %
Trigger Level : 0.005 uCi

#	Chemical Name	Serial #	Cal. Date	Predicted Act.	Wipe Result	Chemical Eff.	Corr. Factor	Removed Act	In Use
1	Co-57	BM0657-042-18	08-25-2006	4,9043 mCi	785 CPM	0.00	1.00	0.000101 uCi	Y
2	Cs-137	S356042-050	01-21-2000	162,9051 uCi	791 CPM	0.00	1.00	0.000129 uCi	Y
3	Cs-137	S139S037	01-03-2000	0.0942 uCi	160 CPM	0.00	1.00	0.000000 uCi	Y

This Report indicates that all above sealed sources have PASSED a Leak Test.
Removed Activity for sources was below the 0.005 uCi NRC/STATE Trigger Limit

Technologist Signature : Kim Roberts