

U.S. NUCLEAR REGULATORY COMMISSION		Date: 10/17/2006
TELEPHONE CONVERSATION RECORD		Time: 1:20 pm
Mail Control 139248 or Report No(s).	License No(s). 37-15400-01	Docket No(s). 03009047
<p>Name of Licensee: St. Mary Hospital</p> <p>Name of Participant(s): Susan Monte, Rad. Oncology/Radiology Administrator Shawn McNeeley, Medical Physicist</p> <p>Telephone No. 215-710-5359</p> <p>Subject: Additional information for HDR authorization (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary: Ms. Monte and Mr. McNeeley confirmed the following items:</p> <ol style="list-style-type: none"> 1. The HDR treatment room door has a "Caution, High Radiation Area" sign posted in addition to a "Caution, Radioactive Materials" posting. 2. The emergency procedures for HDR use will include the telephone numbers for the RSO, authorized users, and authorized medical physicists. <p>Mr. McNeeley also stated that St. Mary Hospital no longer performs intravascular brachytherapy and has returned the Novoste A1000 device to the manufacturer. He will provide documentation of the final disposition of the licensed material so this authorization can be removed from St. Mary's license. In addition, Mr. McNeeley stated that St. Mary Hospital had transferred their Cs-137 brachytherapy sealed sources to another licensee and will provide the documentation to remove these sources from St. Mary's license.</p>		
<p>Action Required: Wait for additional information from licensee.</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SUNSI Review Completed By: Donna M. Janda / RA / Date: 10/17/2006</p>		