

U.S. NUCLEAR REGULATORY COMMISSION		Date: 10/17/2006
TELEPHONE CONVERSATION RECORD		Time: 1:20 pm
Mail Control or Report No(s).	139248	License No(s). 37-15400-01 Docket No(s). 03009047
Name of Licensee:	St. Mary Hospital	
Name of Participant(s):	Susan Monte, Rad. Oncology/Radiology Administrator Shawn McNeeley, Medical Physicist	
Telephone No.	215-710-5359	
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Additional information for HDR authorization	
Summary:	<p>Ms. Monte and Mr. McNeeley confirmed the following items:</p> <ol style="list-style-type: none"> 1. The HDR treatment room door has a "Caution, High Radiation Area" sign posted in addition to a "Caution, Radioactive Materials" posting. 2. The emergency procedures for HDR use will include the telephone numbers for the RSO, authorized users, and authorized medical physicists. <p>Mr. McNeeley also stated that St. Mary Hospital no longer performs intravascular brachytherapy and has returned the Novoste A1000 device to the manufacturer. He will provide documentation of the final disposition of the licensed material so this authorization can be removed from St. Mary's license. In addition, Mr. McNeeley stated that St. Mary Hospital had transferred their Cs-137 brachytherapy sealed sources to another licensee and will provide the documentation to remove these sources from St. Mary's license.</p>	
Action Required:	Wait for additional information from licensee.	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SUNSI Review Completed By:	Donna M. Janda	/ RA / Date: 10/17/2006