



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

October 20, 2006

Docket No. 03013617
Control No. 139416

License No. 45-17898-01

Vincent Mascatello, M.D.
Radiation Safety Officer
Reston Hospital Center, LLC
1850 Town Center Parkway
Reston, VA 20190

**SUBJECT: RESTON HOSPITAL CENTER, LLC, REQUEST FOR ADDITIONAL
INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO
LICENSE, CONTROL NO. 139416**

Dear Dr. Mascatello:

This is in reference to your letter received September 18, 2006 requesting to amend Nuclear Regulatory Commission License No. 45-17898-01. In order to continue our review, we need the following additional information:

1. In support of your request to authorize Michael Rethy, M.D. for materials permitted under 10 CFR 35.100, 35.200, and 31.11, please provide one of the following:
 - a. Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the physician was specifically named as an authorized user for the uses requested; or
 - b. Description of the training and experience identified in 10 CFR Part 35.290(c)(1) demonstrating that the proposed authorized user is qualified by training and experience for the use requested; and
 - c. Written certification, signed by a preceptor physician authorized user, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an authorized user for the medical uses authorized has been achieved.

NRC Form 313A (enclosed) may be used to document the training and work experience described above.

2. In support of your request to authorize David E. Dubois, M.D. for the use of sodium iodide I-131, please provide the following:
 - a. Description of the training identified in 10 CFR Part 35.394(c)(1) and (2); and

V. Mascatello
Reston Hospital Center, LLC

2

- b. Documentation of 3 cases each of oral administration of less than or equal to 33 millicuries of sodium iodide I-131 and oral administration of greater than 33 millicuries of sodium iodide I-131, and
- c. Written certification, signed by a preceptor physician authorized user, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an authorized user for the medical uses authorized has been achieved.

NRC Form 313A (enclosed) may be used to document the training and work experience described above.

3. Your application should have been signed by a management representative rather than yourself. Please submit a letter signed by a management representative indicating that management has reviewed the application and concurs in the statements and representations contained therein. Note also that a management representative should sign all future correspondence that requests a change in your license.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Toolkit Index Page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 139416. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5272.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Tara L. Weidner

Tara L. Weidner
Health Physicist
Medical Branch
Division of Nuclear Materials Safety

Enclosure:
NRC Form 313A

DOCUMENT NAME: C:\FileNet\ML062930112.wpd

SUNSI Review Complete: TWeidner

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI			
NAME	TWeidner/TLW							
DATE	10-20-2006							

OFFICIAL RECORD COPY