

PORTABLE/GAMMA FRISK SOURCE CHECK FORM

Meter Type
E-600

Detector Type
SPA-3

Detector Number
51837

Source ID
277

Acceptance Criteria
- 20%

Acceptance Criteria
+ 20%

PRE USE CHECKS										POST USE CHECKS									
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int				
6-30-05	1200	SAT	N/A	6310	298000	291690	DN	6-30-05	1535	SAT	N/A	6730	307000	300070	DN				
7-7-05	0732	SAT	N/A	6970	306000	299030	DN	7-7-05	1545	SAT	N/A	7050	306000	298950	DN				
7-11-05	0645	SAT	N/A	6000	297000	291000	DN	7-12-05	0630	SAT	N/A	6730	287000	280270	DN				
7-12-05	0630	SAT	N/A	6730	287000	280270	DN	7-12-05	1145	SAT	N/A	6860	276000	269140	DN				
7-14-05	0605	SAT	N/A	7650	307000	299350	DN	7-15-05	0600	SAT	N/A	7200	287000	279800	DN				
7-15-05	0600	SAT	N/A	7200	287000	279800	DN	7-15-05	1315	SAT	N/A	7050	287000	279950	DN				
7-18-05	0615	SAT	N/A	6860	302000	295140	DN	7-18-05	1545	SAT	N/A	7550	274000	266450	DN				
7-19-05	0610	SAT	N/A	8110	299000	290890	DN	7-20-05	0635	SAT	N/A	5900	296000	290100	DN				
7-20-05	0635	SAT	N/A	5900	296000	290100	DN	7-20-05	1605	SAT	N/A	6830	292000	285170	DN				
7-21-05	0625	SAT	N/A	6790	304000	297210	DN	7-21-05	1610	SAT	N/A	7120	306000	298880	DN				
7-22-05	0620	SAT	N/A	6650	292000	285350	DN	7-22-05	1520	SAT	N/A	7460	282000	274540	DN				
7-23-05	0610	SAT	N/A	6800	310000	303200	DN	7-23-05	0610	SAT	N/A	7330	308000	300670	DN				
7-25-05	0610	SAT	N/A	7330	308000	300670	DN	7-24-05	0605	SAT	N/A	7420	270000	282580	DN				
7-26-05	0605	SAT	N/A	7420	290000	282580	DN	7-26-05	1515	SAT	N/A	7140	290000	282860	DN				
								N/A											

RF Supervisor Review:

[Signature]

PORTABLE/GAMMA FEISK

SOURCE CHECK FORM

E-600
Meter Type

SPA-3
Detector Type

51837
Detector Number

277
Source ID

230683
Net Acceptance Criteria - 20%

346025
Net Acceptance Criteria + 20%

PRE USE CHECKS										POST USE CHECKS									
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int				
6-30-05	1200	SAT	N/A	6310	298000	291690	DN	6-30-05	1535	SAT	N/A	6930	307000	300070	DN				
7-7-05	0732	SAT	N/A	6970	306000	299030	DN	7-7-05	1545	SAT	N/A	7050	306000	298950	DN				
7-11-05	0645	SAT	N/A	6000	297000	291000	DN	7-12-05	0630	SAT	N/A	6730	287000	280270	DN				
7-12-05	0630	SAT	N/A	6730	287000	280270	DN	7-12-05	1145	SAT	N/A	6860	276000	269140	DN				
7-14-05	0605	SAT	N/A	7650	307000	299350	DN	7-15-05	0600	SAT	N/A	7200	287000	279000	DN				
7-15-05	0600	SAT	N/A	7200	287000	279800	DN	7-15-05	1315	SAT	N/A	7050	287000	279950	DN				
7-18-05	0615	SAT	N/A	6860	302000	295140	DN	7-18-05	1545	SAT	N/A	7550	274000	266450	DN				
7-19-05	0610	SAT	N/A	8110	299000	290890	DN	7-20-05	0635	SAT	N/A	5900	296000	290100	DN				
7-20-05	0635	SAT	N/A	5900	296000	290100	DN	7-20-05	1605	SAT	N/A	6830	292000	285170	DN				
7-21-05	0625	SAT	N/A	6790	304000	297210	DN	7-21-05	1610	SAT	N/A	7120	306000	298880	DN				
7-22-05	0620	SAT	N/A	6650	292000	285350	DN	7-22-05	1520	SAT	N/A	7460	282000	274540	DN				
7-23-05	0610	SAT	N/A	6800	310000	303200	DN	7-23-05	0610	SAT	N/A	7330	308000	300670	DN				
7-25-05	0610	SAT	N/A	7330	308000	300670	DN	7-26-05	0605	SAT	N/A	7420	290000	282580	DN				
7-26-05	0605	SAT	N/A	7420	290000	282580	DN	7-26-05	1545	SAT	N/A	7140	290000	282860	DN				
								N/A											

RF Supervisor Review: *[Signature]*

PORTABLE/GAMMA FRISK

SOURCE CHECK FORM

E-600
Meter -
Type

SPA-3
Detector
Type

61034
Detector
Number

277
Source
ID

233810
Net
Acceptance
Criteria
- 20%

350716
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
7-20-05	1120	SAT	N/A	5690	276000	270310	DN	7-20-05	1610	SAT	N/A	6240	296000	289760	DN
7-21-05	0840	SAT	N/A	7310	278000	270690	DN	7-21-05	0820	SAT	N/A	6810	272000	265190	DN
7-21-05	0820	SAT	N/A	6810	277000	270190	DN	7-22-05	0615	SAT	N/A	6980	304000	297020	DN
7-22-05	0615	SAT	N/A	6980	304000	297020	DN	7-22-05	1510	SAT	N/A	7740	290000	282260	DN
7-23-05	0615	SAT	N/A	7000	303000	296000	DN	7-23-05	1420	SAT	N/A	7140	290000	282860	DN
7-25-05	0615	SAT	N/A	7230	300000	292770	DN	7-26-05	0615	SAT	N/A	7570	294000	286430	DN
7-27-05	0605	SAT	N/A	7750	292000	284250	DN	7-27-05	1540	SAT	N/A	7700	298000	290300	DN
7-28-05	0450	SAT	N/A	6970	304000	297030	DN	7-28-05	0600	SAT	N/A	7670	288000	280330	DN
7-29-05	0600	SAT	N/A	7670	288000	280330	DN	7-29-05	1532	SAT	N/A	7300	287000	279700	DN
7-30-05	0625	SAT	N/A	6880	303000	296120	DN	7-30-05	1520	SAT	N/A	7890	279000	271110	DN
8-1-05	0555	SAT	N/A	7360	288000	280640	DN	8-2-05	0620	SAT	N/A	6750	289000	282250	DN
8-2-05	0620	SAT	N/A	6750	289000	282250	DN	8-3-05	0610	SAT	N/A	7050	295000	287950	DN
8-3-05	0610	SAT	N/A	7050	295000	287950	DN	8-4-05	0615	SAT	N/A	7030	289000	281970	DN
8-4-05	0615	SAT	N/A	7030	289000	281970	DN	8-5-05	1011	SAT	N/A	7270	291000	283730	DN
						N	A								

RP Supervisor Review:

[Signature]

COPY

PORTABLE/GAMMA F-15 SOURCE CHECK FORM

Meter Type: SPA-3 Detector Type: SPA-3 Source ID: 277 Net Acceptance Criteria: 195710.4

Detector Number: 61035 Net Acceptance Criteria: 293565.6

Alarm Check: N/A BKG Counts: 5260 SRC Counts: 243000 Net Acceptance Criteria: 293565.6

Audible Check: SAT Time: 1623 Date: 6-1-6 Int: CSC

Alarm Check: N/A BKG Counts: 5050 SRC Counts: 244000 Net Acceptance Criteria: 293565.6

Audible Check: SAT Time: 1725 Date: 6-5-6 Int: CSC

Alarm Check: N/A BKG Counts: 5160 SRC Counts: 244000 Net Acceptance Criteria: 293565.6

PRE USE CHECKS							POST USE CHECKS								
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
6-1-6	0550	SAT	N/A	5360	245000	239640	US	6-1-6	1623	SAT	N/A	5260	243000	237740	CSC
6-2-6	0540	SAT	N/A	5050	244000	238950	US	6-2-6	1405	SAT	N/A	5250	245000	239750	CSC
6-5-6	0550	SAT	N/A	5200	246000	240800	US	6-5-6	1725	SAT	N/A	5780	244000	235220	CSC
6-6-6	0540	SAT	N/A	4580	243000	235420	US	6-6-6	1530	SAT	N/A	5160	244000	238840	CSC
6-7-6	0540	SAT	N/A	4710	244000	239280	US								*
6-8-6	0550	SAT	N/A	5000	244000	239000	US	6-8-6	1623	SAT	N/A	5340	245000	239640	CSC
6-9-6	0600	SAT	N/A	4880	245000	240120	US	6-9-6	1655	SAT	N/A	5130	259000	253870	CSC
6-10-6	0500	SAT	N/A	5050	260000	254950	CSC	6-10-6	1700	SAT	N/A	5000	259000	254000	US
6-12-6	0540	SAT	N/A	4900	260000	255100	US	6-12-6	1700	SAT	N/A	5270	259000	253730	CSC
6-13-6	0540	SAT	N/A	5430	258000	252570	US								*
6-14-6	0600	SAT	N/A	5120	279000	253880	US	6-14-6	1645	SAT	N/A	4380	243000	238620	CSC
6-15-6	0530	SAT	N/A	5210	242000	236740	US	6-15-6	1525	SAT	NA	5510	243000	237490	CSC
6-15-6	1525	SAT	NA	5510	243000	237490	CSC	6-15-6	1650	SAT	NA	4840	244000	239160	CSC
6-15-6	1650	SAT	N/A	4640	244000	239160	CSC	6-16-6	0540	SAT	N/A	5090	243000	237910	US
6-16-6	0540	SAT	N/A	5090	243000	237910	US	6-16-6	1300	SAT	NA	4730	243000	237770	CSC

RP Supervisor Review: Brohna Euckson * NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

PORTABLE/GAMMA FRIS 4 SOURCE CHECK FORM

6-600 Meter Type
 SPA-3 Detector Type
 726557 Detector Number
 277 Source ID
 202240 Net Acceptance Criteria - 20%
 303360 Net Acceptance Criteria + 20%

PRE USE CHECKS										POST USE CHECKS									
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int				
6-2-06	0705	SAT	NA	5790	237000	231610	CP	6-2-06	1410	SAT	NA	4530	237000	232470	CP				
6-5-06	0530	SAT	N/A	5440	238000	232560	W	6-5-06	1645	SAT	N/A	56740	238000	232460	CP				
6-6-06	0540	SAT	N/A	5510	237000	231490	D	6-6-06	1410	SAT	NA	5780	237000	231720	CP				
6-6-06	1400	SAT	NP	5280	237000	231720	CP	6-6-06	1715	SAT	NA	5490	238000	232510	CP				
6-8-06	1715	SAT	N/A	5000	236000	232510	CP	6-7-06	0540	SAT	N/A	5200	238000	233000	WS				
6-7-06	0540	SAT	N/A	5000	238200	233000	W	6-8-06	0550	SAT	N/A	5180	237000	231820	WS				
6-8-06	0550	SAT	N/A	5780	237000	231820	W	6-8-06	1630	SAT	NA	5160	239000	233840	CP				
6-9-06	0600	SAT	N/A	5620	234000	232230	W	6-9-06	1705	SAT	NA	5520	251000	245480	CP				
6-10-06	0530	SAT	NP	3790	252000	248010	CP	6-10-06	1700	SAT	N/A	5020	249000	243980	WS				
6-12-06	0540	SAT	N/A	5680	249000	248920	W												
6-13-06	0540	SAT	N/A	5190	251000	245810	W	6-13-06	1630	SAT	NA	5250	257000	245750	CP				
6-14-06	0610	SAT	N/A	5290	252000	246710	W	6-14-06	1630	SAT	NP	4730	235000	230270	CP				
6-15-06	0530	SAT	N/A	5660	236000	230340	W	6-16-06	0540	SAT	N/A	4240	234000	231760	WS				
6-16-06	0540	SAT	N/A	4240	236000	231760	W	6-16-06	1530	SAT	NA	5200	235000	228800	CP				
7-6-06	0830	SAT	NP	5040	234000	228960	W	7-6-06	1615	SAT	NA	5060	233000	227950	W				

RP Supervisor Review: W. [Signature] (1) * NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA FRIS. SOURCE CHECK FORM

E-600 Meter Type SPA-3 Detector Type 725329 Detector Number 277 Source ID 18720 Net Acceptance Criteria - 20% 282180 Net Acceptance Criteria + 20%

PRE USE CHECKS										POST USE CHECKS									
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int				
6-10-06	0530	SAT	N/A	4870	234000	229180	WS	6-10-06	0700	SAT	N/A	4500	239000	234500	WS				
6-17-06	0540	SAT	N/A	4590	224000	219410	WS								①				
6-18-06	0540	SAT	N/A	5400	224000	218600	WS	6-18-06	0705	SAT	N/A	4480	224000	219520	WS				
6-20-06	0550	SAT	N/A	4520	225000	220480	WS								①				
6-21-06	0550	SAT	N/A	4670	221000	216360	WS	6-21-6	0500	SAT	NA	5090	223000	217910	WS				
6-22-06	0510	SAT	N/A	4610	224000	219290	WS	6-22-6	0645	SAT	NA	4800	225000	220200	WS				
6-23-06	0545	SAT	N/A	5030	224000	218970	WS								②				
6-26-06	0545	SAT	N/A	3890	221000	217110	WS	6-26-6	0645	SAT	NA	4840	223000	218160	WS				
6-27-06	0530	SAT	N/A	4480	224000	219520	WS								①				
6-28-06	0545	SAT	N/A	3790	223000	219010	WS	6-28-6	0630	SAT	NA	510	221000	215890	WS				
6-28-06	0550	SAT	N/A	4280	222000	217650	WS	6-29-6	0635	SAT	NA	4680	221000	216320	WS				
6-30-06	0625	SAT	NA	3930	229000	225070	WS	6-30-6	0610	SAT	NA	4460	230000	225540	WS				
7-10-06	0742	SAT	N/A	4860	223000	218140	WS	7-10-6	0700	SAT	NA	4800	222000	217200	WS				
7-7-06	0540	SAT	N/A	4620	222000	218380	WS								①				
7-10-06	0545	SAT	N/A	5250	223000	217750	WS	7-13-06	0730	SAT	N/A	5430	222000	216570	WS				

RP Supervisor Review: W. W. [Signature] (1) ① NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA FR-10 A SOURCE CHECK FORM

E-600
Meter
Type

SPA-3
Detector
Type

726557
Detector
Number

277
Source
ID

202240
Net
Acceptance
Criteria
- 20%

303360
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
6-2-6	0705	SAT	NA	5990	237000	231610	CP	6-2-6	1410	SAT	NA	4530	237000	232470	CP
6-5-6	0550	SAT	N/A	5410	238000	232560	CP	6-5-6	1645	SAT	NA	5540	238000	232460	CP
6-6-6	0540	SAT	N/A	5510	237000	231490	CP	6-6-6	14:00	SAT	NA	5780	237000	231720	CP
6-6-6	14:00	SAT	NA	5250	237000	231720	CP	6-6-6	1715	SAT	NA	5490	238000	232510	CP
6-7-6	1715	SAT	N/A	⁵⁴²⁰ 5200	238000	232510	CP	6-7-6	0540	SAT	N/A	5000	238000	233000	US
6-7-6	0540	SAT	N/A	5000	238200	233008	US	6-8-6	0550	SAT	N/A	5180	237000	231820	US
6-8-6	0550	SAT	N/A	5180	237000	231820	US	6-8-6	1620	SAT	NA	5160	239000	233840	CP
6-9-6	0600	SAT	N/A	5670	239000	232330	US	6-9-6	1705	SAT	NA	5520	251000	245480	CP
6-10-6	0520	SAT	NA	3290	252000	248010	CP	6-10-6	1700	SAT	N/A	5020	249000	243980	US
6-12-6	0540	SAT	N/A	5080	249000	243920	US							*	
6-13-6	0540	SAT	N/A	5190	251000	245810	US	6-13-6	1630	SAT	NA	5250	257000	245750	CP
6-14-6	0610	SAT	N/A	5290	252000	246710	US	6-14-6	1630	SAT	NA	4730	235000	230270	CP
6-15-6	0530	SAT	N/A	5660	236000	230340	US	6-16-6	0540	SAT	N/A	4240	236000	231760	US
6-16-6	0540	SAT	N/A	4240	236000	231760	US	6-16-6	15:30	SAT	NA	5200	237000	229800	CP
7-10-6	0830	SAT	NA	5040	234000	228960	CP	7-10-6	16:15	SAT	NA	5060	233000	227950	CP

RP Supervisor Review: W. W. [Signature] (1)

* NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA FRIS & SOURCE CHECK FORM

E-600
Meter
Type

SPA-3
Detector
Type

2056
Detector
Number

277
Source
ID

181748
Net
Acceptance
Criteria
- 20%

272620
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
6-9-06	0600	SAT	N/A	4930	224000	219070	WS	6-9-06	1650	SAT	NA	5320	237000	231680	CF
6-10-06	0520	SAT	N/A	4520	238000	233480	CF	6-10-06	1700	SAT	N/A	4650	238000	233320	WS
6-12-06	0540	SAT	N/A	5140	236000	230860	WS	6-12-06	1705	SAT	N/A	5130	239000	232870	CF
6-13-06	0540	SAT	N/A	4890	238000	233110	WS								*
6-14-06	0600	SAT	N/A	5370	238000	232630	WS	6-14-06	1635	SAT	NA	4950	224000	219050	CF
6-15-06	0530	SAT	N/A	5360	224000	218640	CF	6-15-06	1645	SAT	NA	4680	223000	218320	CF
7-1 ^A -06	0550	SAT	N/A	4760	221000	216240	WS								*
7-20-06	0545	SAT	N/A	5160	223000	217840	WS	7-20-06	0550	SAT	N/A	5550	220000	214450	WS
7-21-06	0550	SAT	N/A	5550	220000	214450	WS								*
7-24-06	0610	SAT	N/A	5520	222000	216480	WS	7-24-06	15:35	SAT	NA	5130	216000	210870	CF
7-25-06	0545	SAT	N/A	4910	221000	211090	WS	7-25-06	0600	SAT	N/A	5760	220000	214240	WS
7-26-06	0600	SAT	N/A	5760	220000	214240	WS								*
7-27-06	0600	SAT	N/A	4940	220000	215060	WS								*
7-28-06	0600	SAT	N/A	5880	226000	214120	WS	7-28-06	1605	SAT	NA	5420	226000	214580	CF
7-31-06	0550	SAT	N/A	4580	226000	215420	WS								*

RP Supervisor Review: Whiskey (1)

Not used

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA FR15 & SOURCE CHECK FORM

E-600
Meter
Type

SMA-3
Detector
Type

70052
Detector
Number

277
Source
ID

191000
Net
Acceptance
Criteria
- 20%

287000
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
6/5/06	0550	SAT	N/A	4480	240000	235520	NS								*
6-6-06	0540	SAT	N/A	6120	240000	233680	NS	6-6-6	17:00	SAT	NA	5760	239000	233240	NS
6-7-06	0540	SAT	N/A	4960	239000	234040	NS	6-7-06	1510	SAT	N/A	5680	238000	232320	NS
6-9-06	0640	SAT	N/A	5490	237000	231510	NS	6-9-6	1705	SAT	NA	5050	252000	246950	NS
6-10-6	0505	SAT	ND	4790	252000	247210	NS	6-10-6	1700	SAT	N/A	4390	251000	240610	NS
6-12-06	0540	SAT	N/A	4860	251000	246140	NS								*
6-13-06	0540	SAT	N/A	5550	247000	241450	NS	6-13-6	1635	SAT	NA	5480	252000	246520	NS
6-14-06	0600	SAT	N/A	5700	252000	246300	NS	6-14-6	1640	SAT	NA	6230	234000	22770	NS
6-15-06	0530	SAT	N/A	4670	235000	230330	NS	6-15-6	1650	SAT	NA	5090	237000	23910	NS
6-16-06	0545	SAT	N/A	5700	236000	230300	NS								*

RP Supervisor Review: W. Bailey (1) * NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA KRIS SOURCE CHECK FORM

COPY

E-600
Meter
Type

SPA-3
Detector
Type

61035
Detector
Number

277
Source
ID

195710.4
Net
Acceptance
Criteria
- 20%

293565.6
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS								
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	
6-1-06	0550	SAT	N/A	5360	245000	239640	US	6-1-6	1623	SAT	NA	5260	243000	237740	CSC	
6-2-06	0540	SAT	N/A	5050	244000	238950	US	6-2-6	1405	SAT	NA	5250	245000	239750	CSC	
6-5-06	0550	SAT	N/A	5200	246000	240800	US	6-5-6	1725	SAT	NA	5780	244000	238220	CSC	
6-6-06	0540	SAT	N/A	4580	243000	238420	US	6-6-6	1530	SAT	NA	5160	244000	238840	CSC	
6-7-06	0540	SAT	N/A	4710	244000	239290	US								*	
6-8-06	0550	SAT	N/A	5000	244000	239000	US	6-8-6	1623	SAT	NA	5310	245000	239660	CSC	
6-9-06	0600	SAT	N/A	4880	245000	240120	US	6-9-6	1655	SAT	NA	5130	259000	253870	CSC	
6-10-06	0500	SAT	N/A	5050	260000	254950	CSC	6-10-6	1700	SAT	N/A	5000	259000		US	
6-12-06	0540	SAT	N/A	4900	260000	255100	US	6-12-6	1700	SAT	NA	5270	259000		CSC	
6-13-06	0540	SAT	N/A	5430	258000	252570	US	—————								*
6-14-06	0100	SAT	N/A	5120	259000	253880	US	6-14-6	1645	SAT	NA	4380	243000		US	
6-15-06	0530	SAT	N/A	5210	242000	236790	US	6-15-6	1525	SAT	NA	5510	243000	237490	CSC	
6-15-6	1525	SAT	NA	5510	24300	237490	CSC	6-15-6	1650	SAT	NA	4840	244000	239160	CSC	
6-15-6	1650	SAT	NA	4840	244000	239160	CSC	6-16-6	0540	SAT	N/A	5090	24300	237910	US	
6-16-06	0540	SAT	N/A	5090	24300	237910	US	6-16-6	1300	SAT	NA	5230	243000	23770	CSC	

RP Supervisor Review: Breana Erickson ⁽¹⁾

* NOT USED

⁽¹⁾ If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.