

U.S. NUCLEAR REGULATORY COMMISSION	Date: 10/27/2006
TELEPHONE CONVERSATION RECORD	Time: 09:10 AM
Mail Control 139551 License No(s). 45-24867-01	Docket No(s). 03029501
or Report No(s).	
Name of Licensee:	Cardiac Diagnostic Services of Virginia, Inc.
Name of Participant(s):	Neil C. Smarte, RSO
Telephone No.	703-641-0500
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	RAI, Use of gadolinium
Summary:	I asked Mr. Smarte if he had gadolinium 153 sources. He stated that do not plan on using the gadolinium 153 sources any more and they had two sources in storage. Currently those sources were approximately 1 millicurie. They could be maintained at their facility under 10 CFR 25.65. I also stated that we needed to have the training and experience detailed for Dr. Pulerwitz. I recommended NRC form 313A and walked him though getting it from the NRC website. I stated that I would send a letter highlighting the requested information.
Action Required:	Send request for additional information letter.
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date
Prepared & SUNSI Review Completed By:	Dennis Lawyer / RA / Date: 10/27/2006