



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



## Certificate of Calibration

Certificate #

0972-01

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer Model: KEITHLEY 614

Description: ELECTROMETER

Serial No: 0555368

Asset No: 002792

Procedure: KEITHLEY 614 - 10 MAY 2006

Work Order: 303070650

Date Issued: Aug 21, 2006

Calibration Date: Aug 21, 2006

\*Calibration Due: Feb 21, 2007

Calibration Location: Bldg. 64

Environment: Temp. 72.0°F Hum. 43 %RH

\*\*Data Type: FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
000108	7993	GENERAL RADIO	1433G	DECADE RESISTOR	Apr 24, 07
000184	27002	GENERAL RADIO	1433-T	DECADE RESISTOR	Apr 24, 07
012066	MY45040419	HEWLETT-PACKARD	3458A/OPT 002	MULTIMETER	Apr 11, 07
000101	86052	BIDDLE	72-5346-1	DECADE RESISTOR	Jan 05, 07
000182	5200003	FLUKE	5700A/EP	CALIBRATOR	Sep 29, 06
000201	5195014	FLUKE	5725A	AMPLIFIER	Sep 14, 07

Reviewed by: blt ( ) jrg ( ) pwc ( ) rgh ( )

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Leo Salazar

Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303070650	Mfr.	KEITHLEY	Technician	lgs
Asset No.	002792	Model	614		
Serial No.	0555368	Type.	Electrometer	Cal Date.	21-Aug-06
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Test Limits	+/-Uncertainty	Found/Left
Zero	mVolts	mVolts	mVolts	mVolts	mVolts	Results
	0.00000	0.00001	0.00001	0.00001	0.0000012	Pass
DCV	Volts	Volts	Volts	Volts	Volts	
0.2 V	0.19000	0.19001	0.00001	0.00017	0.000022	Pass
	-0.19000	-0.18999	0.00001	0.00017	0.000022	Pass
2 Volt	1.9000	1.9001	0.0001	0.0016	0.00012	Pass
	-1.9000	-1.8999	0.0001	0.0016	0.00012	Pass
20 Volt	19.000	19.000	0.000	0.016	0.0012	Pass
	-19.000	-18.998	0.002	0.016	0.0012	Pass
DC Amps	uAmps	uAmps	uAmps	uAmps	uAmps	
200 uAmp	190.0	190.0	0.0	0.7	0.120	Pass
	nAmps	nAmps	nAmps	nAmps	nAmps	
200 nAmp	190.0	190.0	0.0	1.0	0.12	Pass
	pAmps	pAmps	pAmps	pAmps	pAmps	
2000 pAmp	1900	1894	-6	29	1.2	Pass
Resistance	kOhm	kOhm	kOhm	kOhm	kOhm	
20 kOhm	19.00	18.98	-0.02	0.11	0.012	Pass
200 kOhm	190.0	189.7	-0.3	1.1	0.12	Pass
	MOhm	MOhm	MOhm	MOhm	MOhm	
20 MOhm	10.00	9.99	-0.01	0.10	0.012	Pass
	GOhm	GOhm	GOhm	GOhm	GOhm	
20 GOhm	10.00	10.02	0.02	0.02	0.014	Pass
Charge Calibration	nC	nC	nC	nC	nC	
2 n Coulomb	1.0000	1.0464	0.0464	0.0500	0.00012	Pass

END OF REPORT