

PORTABLE/GAMMA FR... & SOURCE CHECK FORM

Meter Type: E-600
 Detector Type: SMA-3
 Detector Number: 726557
 Source ID: 277
 Net Acceptance Criteria: 202240
 Net Acceptance Criteria + 20%: 303360

PRE USE CHECKS

POST USE CHECKS

PRE USE CHECKS					POST USE CHECKS										
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
6-2-06	0705	SAT	NA	5790	237000	231610	CF	6-2-06	1410	SAT	NA	4590	237000	232470	CF
6-5-06	0550	SAT	N/A	5410	238000	232560	NA	6-5-06	1645	SAT	NA	5640	238000	232460	CF
6-6-06	0540	SAT	N/A	5510	237000	231490	W	6-6-06	1400	SAT	NA	5780	237000	231720	CF
6-6-06	1400	SAT	N/A	5780	237000	231720	CF	6-6-06	1715	SAT	NA	5490	238000	232510	CF
6-8-06	1715	SAT	N/A	5490	238000	232510	ESC	6-7-06	0540	SAT	N/A	5000	238000	233000	WS
6-7-06	0540	SAT	N/A	5000	238000	233000	W	6-8-06	0550	SAT	N/A	5180	237000	231820	WS
6-8-06	0550	SAT	N/A	5180	237000	231820	WS	6-8-06	1430	SAT	NA	5160	239000	233840	ESC
6-9-06	0600	SAT	N/A	5270	236000	231820	WS	6-9-06	1705	SAT	NA	5520	251000	245480	CF
6-10-06	0870	SAT	NA	3720	252000	247010	CF	6-10-06	1700	SAT	N/A	5020	249000	243980	WS
6-12-06	0540	SAT	N/A	5080	249000	248920	WS								*
6-13-06	0540	SAT	N/A	5140	251000	245810	WS	6-13-06	1630	SAT	NA	5250	251000	245750	CF
6-14-06	0600	SAT	N/A	5290	252000	246710	WS	6-14-06	1630	SAT	NP	4730	235000	230270	CF
6-15-06	0530	SAT	N/A	5060	236000	230340	WS	6-16-06	0540	SAT	N/A	4240	234000	231760	WS
6-16-06	0540	SAT	N/A	4240	236000	231760	WS	6-16-06	1530	SAT	NA	5200	231000	229800	CF
7-10-06	0830	SAT	NP	5040	234000	228160	WS	7-10-06	1615	SAT	NA	5050	233000	227950	CF

* NOT USED

RP Supervisor Review: W. W. W. (1)

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA P.R.I.S. SOURCE CHECK FORM

E-600
Meter
Type

SPD-3
Detector
Type

61035
Detector
Number

277
Source
ID

19570.4
Net
Acceptance
Criteria
- 20%

293565.6
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
7-10-6	0835	SAT	NP	4850	241000	236150	CF	7-10-6	1605	SAT	NA	5310	241000	235090	CF
9-19-6	1300	SAT	NA	5240	237000	231200	CF	9-19-6	1720	SAT	N/A	5000	237000	231400	WS
N/A															

RP Supervisor Review: *[Signature]* (1)

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY