CONVERSATION RECORD

TIME | DATE 11-02-2006

O VISIT O CONFERENCE X TELEPHONE

O INCOMING
X OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT	ORGANIZATION (OFFICE, DEPT.ETC.)	TELEPHONE NO.
Harold Legate	Essroc Cement, Inc.	574-739-6131
Plant Chemist	Logansport, Indiana	FAX -722-2168

SUBJECT

Control No. 315676 – Re - Amendment request by NRC FORM 314, dated August 23, 2006, regarding request to terminate their NRC license, License No. 13-09312-02.

SUMMARY

In a telephone conversation with Mr. Legate on this date, I related that we will need the following additional information before proceeding with their request:

1. Please provide a copy of a leak test record, current (within 3 years) as of the date of transfer (9-26-2005) of the gauge/sealed source to Ohmart, which confirms the non-leaking status of the sealed source.

Licensee was requested to provide a copy of the leak test record to me by facsimile at 630-515-1259. Any questions may be directed to me at 630-829-9829.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at http://www.nrc.gov/reading-rm/adams.html (the Public Electronic Reading Room).

ACTION REQUIRED

Licensee to respond as requested.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Loren J. Hueter	Loren J. Hueter	11-02-2006

ACTION TAKEN

SIGNATURE	TITLE	DATE