



**UNITED STATES  
NUCLEAR REGULATORY COMMISSION**

REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, ILLINOIS 60532-4352

NOV 02 2006

Steven L. Hossler, M.D.  
Radiation Safety Officer  
Warsaw Health System, LLC  
d/b/a Kosciusko Community Hospital  
2101 East Dubois Drive  
Warsaw, IN 46580-3288

Dear Dr. Hossler:

This refers to your amendment request dated October 12, 2006, to add gadolinium-153 for use in a Siemens attenuation correction system. Based on several telephone conversations I had with Daniel Trigg, Chief Technologist of your staff, you provided a letter dated October 27, 2006, to withdraw your request. The basis for your request was that 10 CFR 35.65 grants authorization for any medical licensee to possess and use, in part, transmission sealed sources, not exceeding 30 millicuries each, which have been manufactured and distributed by a person licensed under 10 CFR 32.74 of this chapter or equivalent Agreement State regulations. Since you determined that your request meets the requirements in 10 CFR 35.65 and you no longer wish to amend your NRC license, your request is voided without prejudice to resubmission.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>. The enclosed license document is exempt from public disclosure in accordance with 10 CFR 2.390, because its disclosure to unauthorized individuals could present a security vulnerability.

If you have any questions or require clarification on any of the information stated above, please contact me at (630) 829-9873.

Sincerely,

  
James R. Mullauer, M.H.S.  
Health Physicist  
Materials Licensing Branch

License No. 13-32129-01  
Docket No. 030-34848