

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 11/15/06

**TELEPHONE CONVERSATION RECORD**

Time:

Mail Control or Report No(s).	139607	License No(s).	37-31198-01 (NEW)	Docket No(s).	030-37360
----------------------------------	--------	----------------	-------------------	---------------	-----------

Name of Licensee: College Fields MRI &amp; Medical Imaging

Name of Participant(s): Charles Giomuso, Proposed RSO

Telephone No. 216-272-4669

Subject: RAI (139607) dated 11/15/06

(NOTE: This will be used as the  
Documents Title in ADAMS)

**Summary:**

College Fields MRI and Medical Imaging requested an NRC license on 10/17/06. The application contained the Radiation Safety Program (RSP) commitments, as opposed to the statements contained in NUREG 1556, Vol. 9, Rev. 1. I informed Mr. Giomuso that the licensee was not required to provide the NRC with their RSP during the licensing process and that the RSP would be reviewed during the inspection process. Mr. Giomuso will revise the application to contain the information in the checklist contained in NUREG 1556, Vol. 9, Rev. 1. In addition, the following information will be provided:

1. Possession limit for I-131.
2. Information on the location of the camera and Hot Lab on the floor plan of the building.
3. Confirm that the Hot Lab door is one that will lock.
4. Because Mr. Giomuso is an outside consultant, he will address the following:
  - a. Describe the control over the radiation safety program that will be delegated so that the consultant-RSO will be able to exercise authority over authorized users when confronted with radiation safety problems that require implementation of corrective actions.
  - b. Describe the relationship that will exist between the consultant-RSO and your institutional management regarding expenditure of funds to facilitate the objectives of your radiation safety program and related regulatory requirements.
  - c. Identify other commitments of the consultant-RSO for other NRC or Agreement State licensed facilities, along with a description of how the consultant-RSO will allocate time to permit the performance of the duties of the RSO as described in the regulations. State the consultant-RSO's minimum amount of on-site time (hours per week).
  - d. Appoint an in-house representative who will serve as the point of contact during the RSO's absence. This person may be allowed to assist the consultant RSO with limited authority.
  - e. Describe the overall availability of the consultant-RSO to respond to questions or operational issues that arise during the conduct of your radiation safety program and related regulatory requirements. Specify the maximum amount of time it will take the RSO to arrive at the facility in the event of an emergency that requires his presence.

Action Required: Wait for additional information

<b>Document Availability:</b>		<u>  X  </u> <b>Publicly Available</b>	<u>      </u> Non-Publicly Available
<u>  X  </u> Non-Sensitive	<u>      </u> Non-Sensitive Copyright	<u>      </u> Sensitive	<u>      </u> Sensitive Copyright
<u>      </u> Immediate Release	<u>  X  </u> Normal Release	<u>      </u> Delay Release Date	
<b>Prepared &amp; SUNSI Review Completed</b> TLWeidner		<i>/ RA /</i>	Date: 11/15/06
<b>By:</b>			