

PRELIMINARY NOTIFICATION OF EVENT OR UNUSUAL OCCURRENCE PNO-I-06-013

This preliminary notification constitutes EARLY notice of events of POSSIBLE safety or public interest significance. The information is as initially received without verification or evaluation, and is basically all that is known by the Region I staff on this date.

**Facility**

Princeton University  
Office of Environmental Health and Safety  
262 Alexander Street  
Princeton, NJ 08544

**Licensee Emergency Classification**

Notification of Unusual Event  
 Alert  
 Site Area Emergency  
 General Emergency  
 Not Applicable

Docket No.: 030-00882  
License No.: 29-05185-24

**SUBJECT: LEAKING STRONTIUM-90 PLATED SOURCE IDENTIFIED AT UNIVERSITY**

On November 15, 2006, the Princeton University Radiation Safety Officer (RSO) notified the NRC Operations Center that a leaking strontium-90 (Sr-90) source had been identified. The Sr-90 is electroplated on a piece of foil located within a brass cylinder. The source has a current activity of 180 microcuries. The Sr-90 source was identified as leaking on November 13, 2006. The previous leak test performed in May 2006 did not show any contamination.

The source was stored in a drawer located in a room in Jadwin Hall. Surveys show that the drawer is contaminated and that some locations on the floor in this room have minor contamination. The source was used by one researcher in a room across the hall from the storage location between May 2006 and September 2006. The source was not used in October or November. Plates that came into contact with the source during the research activity in this room are contaminated. Complete surveys of the room where the source was used are underway. Surveys of the hallway between the two rooms show no contamination. Only three individuals have been in contact with the source: the researcher; the Radiation Safety technician who performed the leak test; and the RSO who has been examining the source. Preliminary surveys of the researcher's office did not identify contamination. A urine bioassay will be performed when the researcher returns to the University next week.

The licensee notified NRC concerning this matter in accordance with 10 CFR 30.50(b)(1)(i) and will follow with a 30-day written report. The University notified personnel in Jadwin Hall about the leaking source. An article was published in the University newspaper and in the local newspaper. The Region I Office of Public Affairs is prepared to respond to media inquiries.

The State of New Jersey has been notified. This information is current as of 4:00 p.m., November 15, 2006.

Contact: Kathy Modes                      James Dwyer  
(610) 337-5251                      (610) 337-5309

DISTRIBUTIONS:

BY E-MAIL:

OWFN

Chairman Klein  
Comm. McGaffigan Jr.  
Comm. Merrifield  
Comm. Jaczko  
Comm. Lyons

TWFN

NMSS  
OIG  
RES

OIP

OCA

OGC

EDO

OE

NRR

SECY

Regional Offices

BY FAX:

DOT:Trans

INPO

NSAC

RI Resident Office

Licensee: \_\_\_\_\_  
(Reactor Licensees)

SUNSI REVIEW COMPLETED BY: JPD 11/15/06

DOCUMENT NAME: S:\PN\PN106013.wpd

OFFICE	RI/DNMS	RI/DNMS	RI/DNMS	RI/SLO	RI/PAO
NAME	KModes kad	JDwyer jpd1	BHolian beh	NMcNamara mmm3	DScrenci dps
DATE	11/15/2006	11/15/2006	11/15/2006	11/15/2006	11/15/2006