U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 11/22/06

Time:

Mail Control or Report No(s). 139352 License No(s). 29-04333-01 Docket No(s). 030-02476

Name of Licensee:	Trinitas Hospital
Name of Participant(s):	Linda Veldkamp
Telephone No.	908-994-5226
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	RAI (139352) dated 11/22/06
Summary:	I spoke to Ms. Veldkamp to confirm that at no time would they have the two HDR units loaded with sources and stored in the same room. According to Ms. Veldkamp, the licensee will be removing the source (~ 2 curies) from the old HDR unit and placing a new 12 Ci source in the new unit at the same time. At no time will they have greater than 22 Ci co-located.
Action Required:	Complete the amendment

Document Availability: X **Publicly Available** ____ Non-Publicly Available

X Non-Sensitive ____ Non-Sensitive Copyright ____ Sensitive ____ Sensitive Copyright

____ Immediate Release X Normal Release ____ Delay Release Date

Prepared & SUNSI Review Completed By: TLWeidner ***/ RA /*** Date: 11/22/06