U.S. NUCLEAR REGULATORY COMMISSION TELEPHONE CONVERSATION RECORD	Date: 11/27/06 Time: 9:30 a.m.

Mail Control or Report No(s). 2006-001 License No(s). 29-09701-01 Docket No(s). 03002522

Name of Licensee:	Riverview Medical Center
Name of Participant(s):	Michael Cammarano
Telephone No.	(732)530-2529
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	State inspection comments regarding dose calibrator linearity
Summary:	I told Mr. Cammarano that the hospital was in violation of its procedures regarding the testing of dose calibrator linearity to the highest activity actually used. Mr. Cammarano stated that he would send confirmation of corrective actions regarding this violation.
Action Required:	Issue NOV.

Document Availability: X **Publicly Available** ____ Non-Publicly Available

X Non-Sensitive ____ Non-Sensitive Copyright ____ Sensitive ____ Sensitive Copyright

____ Immediate Release X Normal Release ____ Delay Release Date

Prepared & SUNSI Review Completed By: RWM1 ***/ RA /*** Date: 11/27/06